



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: P4722

Contract: WALS01

Lab Code: CHEM

Case No.: P4722

SAS No.: P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB83	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	14:22	LB133407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: P4722

Contract: WALS01

Lab Code: CHEM

Case No.: P4722

SAS No.: P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB75	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	14:29	LB133407
CCB76	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	15:07	LB133407
CCB77	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	15:37	LB133407
CCB78	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	16:01	LB133407
CCB79	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	16:27	LB133407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/14/2024	14:36	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	14:36	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	14:36	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	14:36	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	14:36	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	14:36	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	14:36	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	14:36	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	14:36	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	14:36	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	14:36	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	14:36	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	14:36	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	14:36	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	14:36	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	14:36	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	14:36	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	14:36	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	14:36	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	14:36	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	14:36	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	14:36	LB133457

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>P4722</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4722</u>	SAS No.: <u>P4722</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/14/2024	15:27	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	15:27	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	15:27	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	15:27	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	15:27	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	15:27	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	15:27	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	15:27	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	15:27	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	15:27	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	15:27	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	15:27	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	15:27	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	15:27	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	15:27	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	15:27	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	15:27	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	15:27	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	15:27	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	15:27	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	15:27	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	15:27	LB133457
CCB02	Aluminum	100	+/-100	U	100	P	11/14/2024	15:40	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	15:40	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	15:40	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	15:40	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	15:40	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	15:40	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	15:40	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	15:40	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	15:40	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	15:40	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	15:40	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	15:40	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	15:40	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	15:40	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	15:40	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	15:40	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	15:40	LB133457

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	15:40	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	15:40	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	15:40	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	15:40	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	15:40	LB133457
CCB03	Aluminum	100	+/-100	U	100	P	11/14/2024	16:31	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	16:31	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	16:31	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	16:31	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	16:31	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	16:31	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	16:31	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	16:31	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	16:31	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	16:31	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	16:31	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	16:31	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	16:31	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	16:31	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	16:31	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	16:31	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	16:31	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	16:31	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	16:31	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	16:31	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	16:31	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	16:31	LB133457
CCB04	Aluminum	100	+/-100	U	100	P	11/14/2024	17:27	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	17:27	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	17:27	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	17:27	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	17:27	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	17:27	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	17:27	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	17:27	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	17:27	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	17:27	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	17:27	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	17:27	LB133457

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>P4722</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4722</u>	SAS No.: <u>P4722</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	17:27	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	17:27	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	17:27	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	17:27	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	17:27	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	17:27	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	17:27	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	17:27	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	17:27	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	17:27	LB133457
CCB05	Aluminum	100	+/-100	U	100	P	11/14/2024	18:17	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	18:17	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	18:17	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	18:17	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	18:17	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	18:17	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	18:17	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	18:17	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	18:17	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	18:17	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	18:17	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	18:17	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	18:17	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	18:17	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	18:17	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	18:17	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	18:17	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	18:17	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	18:17	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	18:17	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	18:17	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	18:17	LB133457
CCB06	Aluminum	100	+/-100	U	100	P	11/14/2024	19:07	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	19:07	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	19:07	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	19:07	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	19:07	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	19:07	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	19:07	LB133457

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>P4722</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4722</u>	SAS No.: <u>P4722</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	19:07	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	19:07	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	19:07	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	19:07	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	19:07	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	19:07	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	19:07	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	19:07	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	19:07	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	19:07	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	19:07	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	19:07	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	19:07	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	19:07	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	19:07	LB133457
CCB07	Aluminum	100	+/-100	U	100	P	11/14/2024	19:59	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	19:59	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	19:59	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	19:59	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	19:59	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	19:59	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	19:59	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	19:59	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	19:59	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	19:59	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	19:59	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	19:59	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	19:59	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	19:59	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	19:59	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	19:59	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	19:59	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	19:59	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	19:59	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	19:59	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	19:59	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	19:59	LB133457
CCB08	Aluminum	100	+/-100	U	100	P	11/14/2024	21:42	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	21:42	LB133457

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>P4722</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4722</u>	SAS No.: <u>P4722</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	21:42	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	21:42	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	21:42	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	21:42	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	21:42	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	21:42	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	21:42	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	21:42	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	21:42	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	21:42	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	21:42	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	21:42	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	21:42	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	21:42	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	21:42	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	21:42	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	21:42	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	21:42	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	21:42	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	21:42	LB133457
CCB09	Aluminum	100	+/-100	U	100	P	11/14/2024	22:08	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	22:08	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	22:08	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	22:08	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	22:08	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	22:08	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	22:08	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	22:08	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	22:08	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	22:08	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	22:08	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	22:08	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	22:08	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	22:08	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	22:08	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	22:08	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	22:08	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	22:08	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	22:08	LB133457

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	22:08	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	22:08	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	22:08	LB133457
CCB10	Aluminum	100	+/-100	U	100	P	11/14/2024	23:03	LB133457
	Antimony	50.0	+/-50.0	U	50.0	P	11/14/2024	23:03	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	23:03	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	23:03	LB133457
	Beryllium	6.00	+/-6.00	U	6.00	P	11/14/2024	23:03	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	23:03	LB133457
	Calcium	2000	+/-2000	U	2000	P	11/14/2024	23:03	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	23:03	LB133457
	Cobalt	30.0	+/-30.0	U	30.0	P	11/14/2024	23:03	LB133457
	Copper	20.0	+/-20.0	U	20.0	P	11/14/2024	23:03	LB133457
	Iron	100	+/-100	U	100	P	11/14/2024	23:03	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	23:03	LB133457
	Magnesium	2000	+/-2000	U	2000	P	11/14/2024	23:03	LB133457
	Manganese	20.0	+/-20.0	U	20.0	P	11/14/2024	23:03	LB133457
	Nickel	40.0	+/-40.0	U	40.0	P	11/14/2024	23:03	LB133457
	Potassium	2000	+/-2000	U	2000	P	11/14/2024	23:03	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	23:03	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	23:03	LB133457
	Sodium	2000	+/-2000	U	2000	P	11/14/2024	23:03	LB133457
	Thallium	40.0	+/-40.0	U	40.0	P	11/14/2024	23:03	LB133457
	Vanadium	40.0	+/-40.0	U	40.0	P	11/14/2024	23:03	LB133457
	Zinc	40.0	+/-40.0	U	40.0	P	11/14/2024	23:03	LB133457



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: P4722

Contract: WALS01

Lab Code: CHEM

Case No.: P4722

SAS No.: P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB75	Mercury	0.20	+/-0.20	U	0.20	CV	11/08/2024	10:21	LB133352

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>P4722</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4722</u>		SAS No.: <u>P4722</u>					
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB51	Mercury	0.20	+/-0.20	U	0.20	CV	11/08/2024	10:25	LB133352
CCB52	Mercury	0.20	+/-0.20	U	0.20	CV	11/08/2024	10:56	LB133352
CCB53	Mercury	0.20	+/-0.20	U	0.20	CV	11/08/2024	11:23	LB133352
CCB54	Mercury	0.20	+/-0.20	U	0.20	CV	11/08/2024	11:34	LB133352

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: P4722

Contract: WALS01

Lab Code: CHEM

Case No.: P4722

SAS No.: P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:16	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:16	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>P4722</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4722</u>	SAS No.: <u>P4722</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:50	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:50	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	15:33	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	15:33	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	16:28	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	16:28	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	17:22	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	17:22	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:08	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:08	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:59	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:59	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	19:51	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	19:51	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	20:57	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	20:57	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
CCB09	Barium	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	21:55	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	21:55	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
CCB10	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	23:35	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	23:35	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: P4722

Contract: WALS01

Lab Code: CHEM

Case No.: P4722

SAS No.: P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	20:14	LB133525
	Barium	100	+/-100	U	100	P	11/19/2024	20:14	LB133525
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	20:14	LB133525
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	20:14	LB133525
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	20:14	LB133525
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	20:14	LB133525
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	20:14	LB133525

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	20:47	LB133525
	Barium	100	+/-100	U	100	P	11/19/2024	20:47	LB133525
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	20:47	LB133525
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	20:47	LB133525
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	20:47	LB133525
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	20:47	LB133525
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	20:47	LB133525
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	21:42	LB133525
	Barium	100	+/-100	U	100	P	11/19/2024	21:42	LB133525
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	21:42	LB133525
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	21:42	LB133525
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	21:42	LB133525
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	21:42	LB133525
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	21:42	LB133525