

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	257000	255000	101	216000	294000	11/14/2024	14:54	LB133457
	Antimony	-0.36			-50	50	11/14/2024	14:54	LB133457
	Arsenic	-1.10			-20	20	11/14/2024	14:54	LB133457
	Barium	3.24	6.0	54	-94	106	11/14/2024	14:54	LB133457
	Beryllium	1.25			-6	6	11/14/2024	14:54	LB133457
	Cadmium	6.64	1.0	664	-5	7	11/14/2024	14:54	LB133457
	Calcium	244000	245000	100	208000	282000	11/14/2024	14:54	LB133457
	Chromium	55.5	52.0	107	42	62	11/14/2024	14:54	LB133457
	Cobalt	2.07			-30	30	11/14/2024	14:54	LB133457
	Copper	-4.72	2.0	236	-18	22	11/14/2024	14:54	LB133457
	Iron	98700	101000	98	85600	116500	11/14/2024	14:54	LB133457
	Lead	8.61			-12	12	11/14/2024	14:54	LB133457
	Magnesium	271000	255000	106	216000	294000	11/14/2024	14:54	LB133457
	Manganese	5.69	7.0	81	-13	27	11/14/2024	14:54	LB133457
	Nickel	2.90	2.0	145	-38	42	11/14/2024	14:54	LB133457
	Potassium	24.4			0	0	11/14/2024	14:54	LB133457
	Selenium	-16.4			-20	20	11/14/2024	14:54	LB133457
	Silver	2.63			-10	10	11/14/2024	14:54	LB133457
	Sodium	34.7			0	0	11/14/2024	14:54	LB133457
	Thallium	12.7			-40	40	11/14/2024	14:54	LB133457
	Vanadium	6.61			-40	40	11/14/2024	14:54	LB133457
	Zinc	6.48			-40	40	11/14/2024	14:54	LB133457
ICSAB01	Aluminum	271000	247000	110	209000	285000	11/14/2024	15:18	LB133457
	Antimony	629	618	102	525	711	11/14/2024	15:18	LB133457
	Arsenic	117	104	112	88.4	120	11/14/2024	15:18	LB133457
	Barium	535	537	100	437	637	11/14/2024	15:18	LB133457
	Beryllium	521	495	105	420	570	11/14/2024	15:18	LB133457
	Cadmium	1020	972	105	826	1120	11/14/2024	15:18	LB133457
	Calcium	254000	235000	108	199000	271000	11/14/2024	15:18	LB133457
	Chromium	568	542	105	460	624	11/14/2024	15:18	LB133457
	Cobalt	516	476	108	404	548	11/14/2024	15:18	LB133457
	Copper	493	511	96	434	588	11/14/2024	15:18	LB133457
	Iron	103000	99300	104	84400	114500	11/14/2024	15:18	LB133457
	Lead	56.9	49.0	116	37	61	11/14/2024	15:18	LB133457
	Magnesium	279000	248000	112	210000	286000	11/14/2024	15:18	LB133457
	Manganese	515	507	102	430	584	11/14/2024	15:18	LB133457
	Nickel	1020	954	107	810	1100	11/14/2024	15:18	LB133457
	Potassium	-0.81			0	0	11/14/2024	15:18	LB133457
	Selenium	35.8	46.0	78	26	66	11/14/2024	15:18	LB133457
	Silver	206	201	102	170	232	11/14/2024	15:18	LB133457
	Sodium	-5.35			0	0	11/14/2024	15:18	LB133457
	Thallium	103	108	95	68	148	11/14/2024	15:18	LB133457

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	511	491	104	417	565	11/14/2024	15:18	LB133457
	Zinc	852	952	90	809	1095	11/14/2024	15:18	LB133457

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	5.52			-20	20	11/08/2024	14:24	LB133365
	Barium	1.84	6.0	31	-94	106	11/08/2024	14:24	LB133365
	Cadmium	6.11	1.0	611	-5	7	11/08/2024	14:24	LB133365
	Chromium	54.3	52.0	104	42	62	11/08/2024	14:24	LB133365
	Lead	9.97			-12	12	11/08/2024	14:24	LB133365
	Selenium	-14.4			-20	20	11/08/2024	14:24	LB133365
	Silver	-0.087			-10	10	11/08/2024	14:24	LB133365
ICSAB01	Arsenic	115	104	111	88.4	120	11/08/2024	14:30	LB133365
	Barium	496	537	92	437	637	11/08/2024	14:30	LB133365
	Cadmium	979	972	101	826	1120	11/08/2024	14:30	LB133365
	Chromium	521	542	96	460	624	11/08/2024	14:30	LB133365
	Lead	57.3	49.0	117	37	61	11/08/2024	14:30	LB133365
	Selenium	31.2	46.0	68	26	66	11/08/2024	14:30	LB133365
	Silver	198	201	98	170	232	11/08/2024	14:30	LB133365
ICSA01	Arsenic	7.44			-20	20	11/19/2024	20:23	LB133525
	Barium	4.72	6.0	79	-94	106	11/19/2024	20:23	LB133525
	Cadmium	5.53	1.0	553	-5	7	11/19/2024	20:23	LB133525
	Chromium	59.7	52.0	115	42	62	11/19/2024	20:23	LB133525
	Lead	9.28			-12	12	11/19/2024	20:23	LB133525
	Selenium	-19.2			-20	20	11/19/2024	20:23	LB133525
	Silver	-0.77			-10	10	11/19/2024	20:23	LB133525
ICSAB01	Arsenic	115	104	111	88.4	120	11/19/2024	20:27	LB133525
	Barium	493	537	92	437	637	11/19/2024	20:27	LB133525
	Cadmium	1030	972	106	826	1120	11/19/2024	20:27	LB133525
	Chromium	583	542	108	460	624	11/19/2024	20:27	LB133525
	Lead	57.7	49.0	118	37	61	11/19/2024	20:27	LB133525
	Selenium	30.0	46.0	65	26	66	11/19/2024	20:27	LB133525
	Silver	202	201	100	170	232	11/19/2024	20:27	LB133525