

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
Matrix: Water **Level:** LOW **Client ID:** WB-307-SB02A
Sample ID: P4718-03 **Spiked ID:** P4718-03A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Barium	ug/L	75 - 125	2690		1850		1000	84		P