

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **Level:** LOW **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
Matrix: Water **Sample ID:** P4722-15 **Client ID:** WC-3(0-6)DUP
Percent Solids for Sample: NA **Duplicate ID** P4722-15DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	29.5	J	29.1	J	1		P
Antimony	ug/L	20	2.70	J	3.27	J	19		P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	76.6		76.6		0		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	0.32	J	0.36	J	14		P
Calcium	ug/L	20	167000		167000		0		P
Chromium	ug/L	20	1.28	J	1.41	J	10		P
Cobalt	ug/L	20	15.0	U	15.0	U			P
Copper	ug/L	20	10.4		10.6		2		P
Iron	ug/L	20	96.2		95.2		1		P
Lead	ug/L	20	12.8		13.3		4		P
Magnesium	ug/L	20	6910		6840		1		P
Manganese	ug/L	20	21.7		21.7		0		P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Nickel	ug/L	20	3.85	J	3.81	J	1		P
Potassium	ug/L	20	3330		3310		1		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	19600		19500		1		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	55.1		54.2		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **Level:** LOW **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
Matrix: Water **Sample ID:** P4722-15MS **Client ID:** WC-3(0-6)MSD
Percent Solids for Sample: NA **Duplicate ID** P4722-15MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	983		1010		3		P
Antimony	ug/L	20	390		405		4		P
Arsenic	ug/L	20	396		414		4		P
Barium	ug/L	20	172		176		2		P
Beryllium	ug/L	20	93.8		96.0		2		P
Cadmium	ug/L	20	89.8		93.5		4		P
Calcium	ug/L	20	165000		167000		1		P
Chromium	ug/L	20	188		195		4		P
Cobalt	ug/L	20	91.5		94.9		4		P
Copper	ug/L	20	149		155		4		P
Iron	ug/L	20	1510		1580		5		P
Lead	ug/L	20	451		468		4		P
Magnesium	ug/L	20	7610		7740		2		P
Manganese	ug/L	20	116		119		3		P
Mercury	ug/L	20	4.03		4.41		9		CV
Nickel	ug/L	20	231		240		4		P
Potassium	ug/L	20	8110		8410		4		P
Selenium	ug/L	20	927		982		6		P
Silver	ug/L	20	35.7		37.1		4		P
Sodium	ug/L	20	20300		20900		3		P
Thallium	ug/L	20	869		925		6		P
Vanadium	ug/L	20	145		150		3		P
Zinc	ug/L	20	152		157		3		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **Level:** LOW **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
Matrix: Water **Sample ID:** P4718-03 **Client ID:** WB-307-SB02DUP
Percent Solids for Sample: NA **Duplicate ID** P4718-03DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1850		1790		3		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	16.1	J	14.3	J	12		P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **Level:** LOW **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
Matrix: Water **Sample ID:** P4718-03MS **Client ID:** WB-307-SB02MSD
Percent Solids for Sample: NA **Duplicate ID** P4718-03MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4010		3850		4		P
Barium	ug/L	20	2780		2560		8		P
Cadmium	ug/L	20	1120		1080		4		P
Chromium	ug/L	20	2340		2270		3		P
Lead	ug/L	20	5130		4900		5		P
Mercury	ug/L	20	33.6		40.7		19		CV
Selenium	ug/L	20	9590		9210		4		P
Silver	ug/L	20	362		343		5		P