

Continuing Calibration Report for SequenceID : FC110724AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC067656.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 8272580.000  | 60.000  | 3.171  | 6.452  | 137876.333 | 135469.641 | -1.777 |
| Aliphatic C12-C16 | 5401387.000  | 40.000  | 6.453  | 9.842  | 135034.675 | 131681.730 | -2.546 |
| Aliphatic C16-C21 | 7780791.000  | 60.000  | 9.843  | 13.198 | 129679.850 | 127235.092 | -1.921 |
| Aliphatic C21-C28 | 9757366.000  | 80.000  | 13.199 | 16.853 | 121967.075 | 120090.398 | -1.563 |
| Aliphatic C28-C40 | 12371203.000 | 120.000 | 16.854 | 21.696 | 103093.358 | 101093.195 | -1.979 |
| Aliphatic EPH     | 43583327.000 | 360.000 | 3.171  | 21.696 | 121064.797 | 118799.912 | -1.906 |