

Continuing Calibration Report for SequenceID : FC110724AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067666.D

Aliphatic C9-C12	8377570.000	60.000	3.171	6.452	139626.167	135469.641	-3.068
Aliphatic C12-C16	5480862.000	40.000	6.453	9.842	137021.550	131681.730	-4.055
Aliphatic C16-C21	7888638.000	60.000	9.843	13.198	131477.300	127235.092	-3.334
Aliphatic C21-C28	9900793.000	80.000	13.199	16.853	123759.913	120090.398	-3.056
Aliphatic C28-C40	12533852.000	120.000	16.854	21.696	104448.767	101093.195	-3.319
Aliphatic EPH	44181715.000	360.000	3.171	21.696	122726.986	118799.912	-3.306