

Continuing Calibration Report for SequenceID : FE110924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE051106.D

Aliphatic C9-C12	9286782.000	60.000	3.140	6.758	154779.700	139914.732	-10.624
Aliphatic C12-C16	6146623.000	40.000	6.759	10.191	153665.575	140635.568	-9.265
Aliphatic C16-C21	8900683.000	60.000	10.192	13.550	148344.717	137768.027	-7.677
Aliphatic C21-C28	11665637.000	80.000	13.551	17.206	145820.463	133951.574	-8.861
Aliphatic C28-C40	15458424.000	120.000	17.207	22.053	128820.200	128748.418	-0.056
Aliphatic EPH	51458149.000	360.000	3.140	22.053	142939.303	134589.790	-6.204