

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/09/2024

Lab Sample No.: HSTDCCC750 Data File : PS028379.D Time Analyzed: 00:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.550	9.452	9.652	703.660	712.500	-1.2
2,4,5-TP(Silvex)	9.257	9.160	9.360	705.350	712.500	-1.0
2,4-D	8.376	8.279	8.479	683.840	705.000	-3.0
2,4-DB	10.126	10.029	10.229	689.080	712.500	-3.3
2,4-DCAA	7.247	7.150	7.350	729.100	750.000	-2.8
DICAMBA	7.436	7.338	7.538	700.700	705.000	-0.6
DICHLORPROP	8.145	8.048	8.248	689.680	705.000	-2.2
Dinoseb	11.341	11.245	11.445	684.820	705.000	-2.9

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/09/2024

Lab Sample No.: HSTDCCC750 Data File : PS028379.D Time Analyzed: 00:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.329	10.233	10.433	699.380	712.500	-1.8
2,4,5-TP(Silvex)	9.909	9.811	10.011	696.990	712.500	-2.2
2,4-D	9.002	8.905	9.105	674.630	705.000	-4.3
2,4-DB	10.897	10.800	11.000	688.400	712.500	-3.4
2,4-DCAA	7.753	7.655	7.855	729.560	750.000	-2.7
DICAMBA	7.954	7.856	8.056	698.730	705.000	-0.9
DICHLORPROP	8.671	8.573	8.773	682.450	705.000	-3.2
Dinoseb	11.277	11.180	11.380	675.370	705.000	-4.2