

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8151A

DataFile : PS028373.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: P4739-01MS	TP-14MS											
	DICAMBA	187.7	0	68.5	ug/Kg	36				10	112	
	DICHLORPROP	187.7	0	40.4	ug/Kg	22				10	113	
	2,4-D	187.7	0	72.7	ug/Kg	39				10	144	
	2,4,5-TP(Silvex)	187.7	0	27.4	ug/Kg	15				10	114	
	2,4,5-T	187.7	0	37.0	ug/Kg	20				10	115	
	2,4-DB	187.7	0	27.6	ug/Kg	15				10	140	
	Dinoseb	187.7	0	14.2	ug/Kg	8	*			10	118	