

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8151A

DataFile : PS028374.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	TP-14MSD											
P4739-01MSD	DICAMBA	187.8	0	67.7	ug/Kg	36		0		10	112	20
	DICHLORPROP	187.8	0	39.6	ug/Kg	21		5		10	113	20
	2,4-D	187.8	0	71.0	ug/Kg	38		3		10	144	20
	2,4,5-TP(Silvex)	187.8	0	26.6	ug/Kg	14		7		10	114	20
	2,4,5-T	187.8	0	36.3	ug/Kg	19		5		10	115	20
	2,4-DB	187.8	0	26.9	ug/Kg	14		7		10	140	20
	Dinoseb	187.8	0	15.0	ug/Kg	8	*	0		10	118	20