

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107784.D Time Analyzed: 16:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.524	5.423	5.623	465.890	500.000	-6.8
Aroclor-1016-2	5.546	5.445	5.645	452.740	500.000	-9.5
Aroclor-1016-3	5.610	5.507	5.707	444.240	500.000	-11.2
Aroclor-1016-4	5.706	5.604	5.804	472.470	500.000	-5.5
Aroclor-1016-5	6.002	5.899	6.099	494.720	500.000	-1.1
Aroclor-1260-1	7.132	7.026	7.226	469.850	500.000	-6.0
Aroclor-1260-2	7.389	7.282	7.482	493.320	500.000	-1.3
Aroclor-1260-3	7.752	7.643	7.843	493.890	500.000	-1.2
Aroclor-1260-4	7.977	7.868	8.068	522.330	500.000	4.5
Aroclor-1260-5	8.291	8.181	8.381	525.080	500.000	5.0
Decachlorobiphenyl	10.079	9.958	10.158	46.400	50.000	-7.2
Tetrachloro-m-xylene	4.372	4.274	4.474	56.540	50.000	13.1

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107784.D Time Analyzed: 16:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	572.960	500.000	14.6
Aroclor-1016-2	4.742	4.645	4.845	588.760	500.000	17.8
Aroclor-1016-3	4.918	4.821	5.021	566.130	500.000	13.2
Aroclor-1016-4	4.960	4.862	5.062	508.790	500.000	1.8
Aroclor-1016-5	5.173	5.076	5.276	562.090	500.000	12.4
Aroclor-1260-1	6.205	6.107	6.307	525.520	500.000	5.1
Aroclor-1260-2	6.392	6.294	6.494	543.960	500.000	8.8
Aroclor-1260-3	6.546	6.448	6.648	513.070	500.000	2.6
Aroclor-1260-4	7.017	6.918	7.118	479.570	500.000	-4.1
Aroclor-1260-5	7.258	7.159	7.359	492.600	500.000	-1.5
Decachlorobiphenyl	8.640	8.539	8.739	43.780	50.000	-12.4
Tetrachloro-m-xylene	3.642	3.545	3.745	57.900	50.000	15.8