

LAB CHRONICLE

OrderID:	P4722	OrderDate:	11/5/2024 3:33:08 PM
Client:	Walsh Construction Company II, LLC	Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2
Contact:	Kayla Timony	Location:	L23,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4722-03	WC-1(0-6)	SOIL	PCB	8082A	11/05/24	11/07/24	11/07/24	11/05/24
P4722-08	WC-2(0-6)	SOIL	PCB	8082A	11/05/24	11/07/24	11/07/24	11/05/24
P4722-13	WC-3(0-6)	SOIL	PCB	8082A	11/05/24	11/07/24	11/07/24	11/05/24

Hit Summary Sheet
SW-846

SDG No.: P4722

Order ID: P4722

Client: Walsh Construction Company II, LLC

Project ID: NYCDEP C547A - Shafts 17B-1 & 18B

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : WC-1(0-6)								
P4722-03	WC-1(0-6)	SOIL	Aroclor-1254	60.7	P	2.90	18.1	ug/kg
P4722-03	WC-1(0-6)	SOIL	Aroclor-1260	54.7		3.10	18.1	ug/kg
Total Concentration:				115.400				
Client ID : WC-2(0-6)								
P4722-08	WC-2(0-6)	SOIL	Aroclor-1254	41.7		3.00	18.8	ug/kg
P4722-08	WC-2(0-6)	SOIL	Aroclor-1260	43.2		3.20	18.8	ug/kg
Total Concentration:				84.900				
Client ID : WC-3(0-6)								
P4722-13	WC-3(0-6)	SOIL	Aroclor-1254	116		3.20	19.6	ug/kg
P4722-13	WC-3(0-6)	SOIL	Aroclor-1260	78.8		3.40	19.6	ug/kg
Total Concentration:				194.800				



SAMPLE DATA

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	11/05/24	
Client Sample ID:	WC-1(0-6)		SDG No.:	P4722	
Lab Sample ID:	P4722-03		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	93.8	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107781.D	1	11/07/24 08:40	11/07/24 14:31	PB164748

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.60	U	3.60	18.1	ug/kg
11104-28-2	Aroclor-1221	6.80	U	6.80	18.1	ug/kg
11141-16-5	Aroclor-1232	3.60	U	3.60	18.1	ug/kg
53469-21-9	Aroclor-1242	3.60	U	3.60	18.1	ug/kg
12672-29-6	Aroclor-1248	8.40	U	8.40	18.1	ug/kg
11097-69-1	Aroclor-1254	60.7	P	2.90	18.1	ug/kg
37324-23-5	Aroclor-1262	4.90	U	4.90	18.1	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	18.1	ug/kg
11096-82-5	Aroclor-1260	54.7		3.10	18.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.1		32 - 144	106%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.3		32 - 175	121%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/05/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/05/24
Client Sample ID:	WC-2(0-6)	SDG No.:	P4722
Lab Sample ID:	P4722-08	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.1
Sample Wt/Vol:	30.1	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107782.D	1	11/07/24 08:40	11/07/24 14:47	PB164748

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.70	U	3.70	18.8	ug/kg
11104-28-2	Aroclor-1221	7.10	U	7.10	18.8	ug/kg
11141-16-5	Aroclor-1232	3.80	U	3.80	18.8	ug/kg
53469-21-9	Aroclor-1242	3.70	U	3.70	18.8	ug/kg
12672-29-6	Aroclor-1248	8.70	U	8.70	18.8	ug/kg
11097-69-1	Aroclor-1254	41.7		3.00	18.8	ug/kg
37324-23-5	Aroclor-1262	5.10	U	5.10	18.8	ug/kg
11100-14-4	Aroclor-1268	3.80	U	3.80	18.8	ug/kg
11096-82-5	Aroclor-1260	43.2		3.20	18.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.4		32 - 144	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.6		32 - 175	103%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	11/05/24	
Client Sample ID:	WC-3(0-6)		SDG No.:	P4722	
Lab Sample ID:	P4722-13		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.6	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107783.D	1	11/07/24 08:40	11/07/24 15:04	PB164748

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	19.6	ug/kg
11104-28-2	Aroclor-1221	7.40	U	7.40	19.6	ug/kg
11141-16-5	Aroclor-1232	3.90	U	3.90	19.6	ug/kg
53469-21-9	Aroclor-1242	3.90	U	3.90	19.6	ug/kg
12672-29-6	Aroclor-1248	9.10	U	9.10	19.6	ug/kg
11097-69-1	Aroclor-1254	116		3.20	19.6	ug/kg
37324-23-5	Aroclor-1262	5.30	U	5.30	19.6	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	19.6	ug/kg
11096-82-5	Aroclor-1260	78.8		3.40	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.0		32 - 144	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.7		32 - 175	89%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: **P4722**

Client: **Walsh Construction Company II, LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO107183.D	PIBLK-PO107183.D	Tetrachloro-m-xylene	1	20	22.5	112		60	140
		Decachlorobiphenyl	1	20	23.4	117		60	140
		Tetrachloro-m-xylene	2	20	22.1	110		60	140
		Decachlorobiphenyl	2	20	22.8	114		60	140
I.BLK-PO107773.D	PIBLK-PO107773.D	Tetrachloro-m-xylene	1	20	20.8	104		60	140
		Decachlorobiphenyl	1	20	21.9	109		60	140
		Tetrachloro-m-xylene	2	20	20.8	104		60	140
		Decachlorobiphenyl	2	20	22.9	114		60	140
PB164748BL	PB164748BL	Tetrachloro-m-xylene	1	20	20.0	100		32	144
		Decachlorobiphenyl	1	20	21.5	108		32	175
		Tetrachloro-m-xylene	2	20	19.9	100		32	144
		Decachlorobiphenyl	2	20	21.9	110		32	175
PB164748BS	PB164748BS	Tetrachloro-m-xylene	1	20	21.3	107		32	144
		Decachlorobiphenyl	1	20	22.4	112		32	175
		Tetrachloro-m-xylene	2	20	20.2	101		32	144
		Decachlorobiphenyl	2	20	23.2	116		32	175
P4720-01MS	JC-701-COMP-01MS	Tetrachloro-m-xylene	1	20	23.1	116		32	144
		Decachlorobiphenyl	1	20	26.4	132		32	175
		Tetrachloro-m-xylene	2	20	21.9	110		32	144
		Decachlorobiphenyl	2	20	29.5	148		32	175
P4720-01MSD	JC-701-COMP-01MSD	Tetrachloro-m-xylene	1	20	22.8	114		32	144
		Decachlorobiphenyl	1	20	27.3	136		32	175
		Tetrachloro-m-xylene	2	20	21.9	109		32	144
		Decachlorobiphenyl	2	20	29.8	149		32	175
P4722-03	WC-1(0-6)	Tetrachloro-m-xylene	1	20	18.8	94		32	144
		Decachlorobiphenyl	1	20	24.3	121		32	175
		Tetrachloro-m-xylene	2	20	21.1	106		32	144
		Decachlorobiphenyl	2	20	20.1	101		32	175
P4722-08	WC-2(0-6)	Tetrachloro-m-xylene	1	20	15.5	78		32	144
		Decachlorobiphenyl	1	20	20.6	103		32	175
		Tetrachloro-m-xylene	2	20	19.4	97		32	144
		Decachlorobiphenyl	2	20	16.6	83		32	175
P4722-13	WC-3(0-6)	Tetrachloro-m-xylene	1	20	17.6	88		32	144
		Decachlorobiphenyl	1	20	17.7	89		32	175
		Tetrachloro-m-xylene	2	20	20.0	100		32	144
		Decachlorobiphenyl	2	20	14.5	73		32	175
I.BLK-PO107788.D	PIBLK-PO107788.D	Tetrachloro-m-xylene	1	20	20.4	102		60	140
		Decachlorobiphenyl	1	20	19.8	99		60	140
		Tetrachloro-m-xylene	2	20	20.8	104		60	140
		Decachlorobiphenyl	2	20	18.6	93		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8082A

DataFile : PO107779.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: P4720-01MS	JC-701-COMP-01MS											
	AR1016	183.5	0	203	ug/kg	111				55	146	
	AR1260	183.5	25.7	232	ug/kg	112				45	144	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8082A

DataFile : PO107780.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JC-701-COMP-01MSD											
P4720-01MSD	AR1016	183.6	0	204	ug/kg	111		0		55	146	20
	AR1260	183.6	25.7	233	ug/kg	113		1		45	144	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8082A Datafile : PO107776.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB164748BS	AR1016	166.6	163	ug/kg	98				71	120		
	AR1260	166.6	167	ug/kg	100				65	130		

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB164748BL

Lab Name: CHEMTECH

Contract: WALS01

Lab Code: CHEM Case No.: P4722

SAS No.: P4722 SDG NO.: P4722

Lab Sample ID: PB164748BL

Lab File ID: PO107775.D

Matrix: (soil/water) Solid

Extraction: (Type) _____

Sulfur Cleanup: (Y/N) N

Date Extracted: 11/07/2024

Date Analyzed (1): 11/07/2024

Date Analyzed (2): 11/07/2024

Time Analyzed (1): 12:54

Time Analyzed (2): 12:54

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB164748BS	PB164748BS	PO107776.D	11/07/2024	11/07/2024
JC-701-COMP-01MS	P4720-01MS	PO107779.D	11/07/2024	11/07/2024
JC-701-COMP-01MSD	P4720-01MSD	PO107780.D	11/07/2024	11/07/2024
WC-1 (0-6)	P4722-03	PO107781.D	11/07/2024	11/07/2024
WC-2 (0-6)	P4722-08	PO107782.D	11/07/2024	11/07/2024
WC-3 (0-6)	P4722-13	PO107783.D	11/07/2024	11/07/2024

COMMENTS: _____



CALIBRATION SUMMARY

A
B
C
D
E
F
G

GC Column: ZB-MR1 ID: 0.32 (mm)

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.06	10.06	10.06	10.06	10.06	10.06	9.96	10.16
Tetrachloro-m-xylene	4.37	4.37	4.37	4.37	4.37	4.37	4.27	4.47

A
B
C
D
E
F
G

A
B
C
D
E
F
G

GC Column: ZB-MR2 ID: 0.32 (mm)

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.64	8.64	8.64	8.64	8.64	8.64	8.54	8.74
Tetrachloro-m-xylene	3.65	3.64	3.65	3.64	3.65	3.65	3.55	3.75

A
B
C
D
E
F
G

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Instrument ID: ECD_O

Calibration Date(s): 10/15/2024 10/16/2024

Calibration Times: 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO107184.D</u>		CF 750 = <u>PO107185.D</u>			
CF 500 = <u>PO107186.D</u>		CF 250 = <u>PO107187.D</u>		CF 050 = <u>PO107188.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	252081604	261702867	271939076	291118488	272229420	269814291 5
Aroclor-1016-2	(2)	373760062	383028583	397887352	419075308	410673320	396884925 5
Aroclor-1016-3	(3)	234678633	244145249	256575754	276405664	244042480	251169556 6
Aroclor-1016-4	(4)	185622493	193848624	202963902	215081552	165543020	192611918 10
Aroclor-1016-5	(5)	176326470	182141932	191487930	203514892	159796500	182653545 9
Aroclor-1260-1	(1)	238480871	246497651	259155974	277220676	271587560	258588546 6
Aroclor-1260-2	(2)	243402558	251747185	264083512	284216460	272650860	263220115 6
Aroclor-1260-3	(3)	167353086	171560129	181491762	195138068	183182660	179745141 6
Aroclor-1260-4	(4)	162125557	166999933	175289716	187582032	185522120	175503872 6
Aroclor-1260-5	(5)	268686175	273623828	283035894	299271412	297351500	284393762 5
Decachlorobiphenyl		2362750210	2427138920	2496479000	2603827440	2376686800	2453376474 4
Tetrachloro-m-xylene		8902656430	9087809293	9285762860	9548006280	8745262400	9113899453 3
Aroclor-1242-1	(1)	207947063	205572916	222208808	238956456	243522880	223641625 8
Aroclor-1242-2	(2)	302439626	299949616	323550412	342804408	348858880	323520588 7
Aroclor-1242-3	(3)	191912697	193178427	209638152	224306728	225847640	208976729 8
Aroclor-1242-4	(4)	150447174	145338196	162797508	173135964	168308740	160005516 7
Aroclor-1242-5	(5)	141009343	143408436	153689098	164623016	179679780	156481935 10
Decachlorobiphenyl		2339420350	2427742867	2492700840	2551827400	2426410200	2447620331 3
Tetrachloro-m-xylene		8860043620	8651743680	9146784160	9504075440	9174828800	9067495140 4
Aroclor-1248-1	(1)	155913014	164800668	175191332	184609772	188867720	173876501 8
Aroclor-1248-2	(2)	222845576	238202488	253637968	270074056	280366620	253025342 9
Aroclor-1248-3	(3)	231904923	246233752	259706092	274388212	266197800	255686156 7
Aroclor-1248-4	(4)	237210289	246633865	260533272	272872732	271860120	257822056 6
Aroclor-1248-5	(5)	236273499	246591072	260733536	276314120	280304600	260043365 7
Decachlorobiphenyl		2339780580	2371533013	2504089680	2566158680	2396701200	2435652631 4
Tetrachloro-m-xylene		8719403610	9119992693	9358904180	9492784200	8995773400	9137371617 3
Aroclor-1254-1	(1)	246822697	256579683	264992242	286189960	289157060	268748328 7
Aroclor-1254-2	(2)	348120894	361641555	373031238	401237980	402995760	377405485 6
Aroclor-1254-3	(3)	336293322	345653409	355909306	378498460	372630440	357796987 5
Aroclor-1254-4	(4)	213138598	220171503	227187688	242530220	244046960	229414994 6
Aroclor-1254-5	(5)	194609638	201069208	205869028	220139672	216546640	207646837 5
Decachlorobiphenyl		2357746210	2401897693	2508594000	2601578200	2359367400	2445836701 4
Tetrachloro-m-xylene		8883763230	9164627320	9184369060	9508828480	8791509400	9106619498 3
Aroclor-1268-1	(1)	351862425	352413819	364622844	381760528	369380480	364008019 3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	313096403	314491445	324862806	339637696	320299560	322477582	3
Aroclor-1268-3	(3)	272236453	272760276	282620036	293538700	275661940	279363481	3
Aroclor-1268-4	(4)	112236308	111261111	114714594	116428084	101458340	111219687	5
Aroclor-1268-5	(5)	836674479	833098204	847063314	864357444	781067560	832452200	4
Decachlorobiphenyl		4074190590	4136234160	4241958160	4399826360	4004084600	4171258774	4
Tetrachloro-m-xylene		9139601460	8935957933	9311606620	9495189880	8951901000	9166851379	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Instrument ID: ECD_O

Calibration Date(s): 10/15/2024 10/16/2024

Calibration Times: 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO107184.D</u>		CF 750 = <u>PO107185.D</u>				
CF 500 = <u>PO107186.D</u>		CF 250 = <u>PO107187.D</u>		CF 050 = <u>PO107188.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	98519363	99934743	102532802	106601384	108740020	103265662	4
Aroclor-1016-2	(2)	140533795	142751421	144315942	146877552	130360060	140967754	5
Aroclor-1016-3	(3)	75819146	77111448	78799028	81783348	80215260	78745646	3
Aroclor-1016-4	(4)	61532960	63413097	65607722	69564368	70423600	66108349	6
Aroclor-1016-5	(5)	78878086	80970861	83522668	86940988	80356660	82133853	4
Aroclor-1260-1	(1)	150450531	152030172	155131756	161942364	156218200	155154605	3
Aroclor-1260-2	(2)	173752920	181526657	184104290	189736236	151457920	176115605	8
Aroclor-1260-3	(3)	168454604	170292340	172108108	175991544	151894120	167748143	6
Aroclor-1260-4	(4)	143637024	144895129	147115166	150514212	137242140	144680734	3
Aroclor-1260-5	(5)	340738400	339782780	336796716	340391952	289493480	329440666	7
Decachlorobiphenyl		2730622670	2745732653	2784792900	2852828640	2594005000	2741596373	3
Tetrachloro-m-xylene		3312014480	3338472613	3349065240	3230716080	2862214000	3218496483	6
Aroclor-1242-1	(1)	80390662	80447493	83937762	87266896	86674240	83743411	4
Aroclor-1242-2	(2)	114487494	111266404	117741916	120667112	114031640	115638913	3
Aroclor-1242-3	(3)	61817452	60461905	64396992	66865868	63624080	63433259	4
Aroclor-1242-4	(4)	60582529	60636219	64384594	68045260	66598220	64049364	5
Aroclor-1242-5	(5)	73763103	75675432	77215916	80736528	80226620	77523520	4
Decachlorobiphenyl		2686985320	2764418280	2764436780	2798541920	2665265400	2735929540	2
Tetrachloro-m-xylene		3315724680	3241948653	3341198040	3335574400	2920075400	3230904235	6
Aroclor-1248-1	(1)	60218805	63038185	64879120	66667236	62402400	63441149	4
Aroclor-1248-2	(2)	85173308	89108361	93061950	96541480	93540340	91485088	5
Aroclor-1248-3	(3)	89257070	93226611	97377878	100891904	96212540	95393201	5
Aroclor-1248-4	(4)	105934439	110533117	114760412	117772688	111691080	112138347	4
Aroclor-1248-5	(5)	102828085	105359652	109801380	114217964	119140640	110269544	6
Decachlorobiphenyl		2669365810	2682497173	2791873700	2836714480	2679938000	2732077833	3
Tetrachloro-m-xylene		3278417600	3389239440	3425836480	3362184040	2965722400	3284279992	6
Aroclor-1254-1	(1)	159855276	163854537	165548332	171717704	162964860	164788142	3
Aroclor-1254-2	(2)	138468840	142269335	144546922	151325972	148203140	144962842	3
Aroclor-1254-3	(3)	227061110	231385951	232213406	238575248	218867080	229620559	3
Aroclor-1254-4	(4)	128242906	130249965	131325350	135207044	122091880	129423429	4
Aroclor-1254-5	(5)	191904554	195153152	195378104	201188864	171636600	191052255	6
Decachlorobiphenyl		2726927150	2738190720	2773447120	2836815760	2598739800	2734824110	3
Tetrachloro-m-xylene		3347799560	3397335627	3354246660	3362236280	2930639600	3278451545	6
Aroclor-1268-1	(1)	419191400	409991492	413863584	410849472	368955300	404570250	5

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	389672996	380096764	383518208	378699872	334345600	373266688	6
Aroclor-1268-3	(3)	345284332	337828173	339535756	338755760	307004380	333681680	5
Aroclor-1268-4	(4)	129107993	127445005	127285362	129260796	113749480	125369727	5
Aroclor-1268-5	(5)	1049237500	1019311260	1017080130	996315476	856840280	987756929	8
Decachlorobiphenyl		4873889790	4739451973	4823511780	4890043960	4430545400	4751488581	4
Tetrachloro-m-xylene		3440286630	3329771373	3416104800	3331345720	2963806400	3296262985	6

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Instrument ID: ECD_O Date(s) Analyzed: 10/15/2024 10/16/2024

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.57	4.47	4.67	110010000
		2	4.66	4.56	4.76	78836600
		3	4.73	4.63	4.83	234100000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.73	4.63	4.83	195844000
		2	5.26	5.16	5.36	103025000
		3	5.54	5.44	5.64	179001000
		4	5.70	5.60	5.80	90165400
		5	5.80	5.70	5.90	65903200
Aroclor-1262	500	1	7.74	7.64	7.84	245548000
		2	8.28	8.18	8.38	309420000
		3	8.59	8.49	8.69	209196000
		4	8.68	8.58	8.78	163066000
		5	9.32	9.22	9.42	102764000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Instrument ID: ECD_O Date(s) Analyzed: 10/15/2024 10/16/2024

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.86	3.76	3.96	34895600
		2	3.94	3.84	4.04	26624800
		3	4.02	3.92	4.12	81218600
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.02	3.92	4.12	67327000
		2	4.75	4.65	4.85	63205200
		3	4.92	4.82	5.02	34475000
		4	5.00	4.90	5.10	31466000
		5	5.18	5.08	5.28	33083400
Aroclor-1262	500	1	6.76	6.66	6.86	212876000
		2	7.26	7.16	7.36	363332000
		3	7.54	7.44	7.64	137384000
		4	7.61	7.51	7.71	264574000
		5	8.10	8.00	8.20	112833000

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Continuing Calib Date: 11/07/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 09:00 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.53	5.52	5.42	5.62	-0.01
Aroclor-1016-2	(2)	5.55	5.55	5.45	5.65	0.00
Aroclor-1016-3	(3)	5.61	5.61	5.51	5.71	0.00
Aroclor-1016-4	(4)	5.71	5.70	5.60	5.80	-0.01
Aroclor-1016-5	(5)	6.00	6.00	5.90	6.10	0.00
Aroclor-1260-1	(1)	7.13	7.13	7.03	7.23	0.00
Aroclor-1260-2	(2)	7.39	7.38	7.28	7.48	-0.01
Aroclor-1260-3	(3)	7.75	7.74	7.64	7.84	-0.01
Aroclor-1260-4	(4)	7.98	7.97	7.87	8.07	-0.01
Aroclor-1260-5	(5)	8.29	8.28	8.18	8.38	-0.01
Tetrachloro-m-xylene		4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl		10.08	10.06	9.96	10.16	-0.02

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Continuing Calib Date: 11/07/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 09:00 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.72	4.73	4.63	4.83	0.01
Aroclor-1016-2	(2)	4.74	4.75	4.65	4.85	0.01
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.18	5.08	5.28	0.01
Aroclor-1260-1	(1)	6.20	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.00
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.64	3.65	3.55	3.75	0.01
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107769.D Time Analyzed: 09:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.423	5.623	533.240	500.000	6.6
Aroclor-1016-2	5.547	5.445	5.645	524.920	500.000	5.0
Aroclor-1016-3	5.610	5.507	5.707	512.270	500.000	2.5
Aroclor-1016-4	5.707	5.604	5.804	536.660	500.000	7.3
Aroclor-1016-5	6.003	5.899	6.099	549.900	500.000	10.0
Aroclor-1260-1	7.133	7.026	7.226	569.760	500.000	14.0
Aroclor-1260-2	7.389	7.282	7.482	591.260	500.000	18.3
Aroclor-1260-3	7.752	7.643	7.843	572.350	500.000	14.5
Aroclor-1260-4	7.978	7.868	8.068	583.920	500.000	16.8
Aroclor-1260-5	8.293	8.181	8.381	596.880	500.000	19.4
Decachlorobiphenyl	10.079	9.958	10.158	51.570	50.000	3.1
Tetrachloro-m-xylene	4.373	4.274	4.474	54.530	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107769.D Time Analyzed: 09:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.723	4.626	4.826	561.030	500.000	12.2
Aroclor-1016-2	4.742	4.645	4.845	577.570	500.000	15.5
Aroclor-1016-3	4.919	4.821	5.021	552.200	500.000	10.4
Aroclor-1016-4	4.960	4.862	5.062	485.110	500.000	-3.0
Aroclor-1016-5	5.173	5.076	5.276	550.490	500.000	10.1
Aroclor-1260-1	6.203	6.107	6.307	559.730	500.000	11.9
Aroclor-1260-2	6.391	6.294	6.494	577.550	500.000	15.5
Aroclor-1260-3	6.545	6.448	6.648	543.970	500.000	8.8
Aroclor-1260-4	7.016	6.918	7.118	540.980	500.000	8.2
Aroclor-1260-5	7.257	7.159	7.359	574.450	500.000	14.9
Decachlorobiphenyl	8.638	8.539	8.739	56.920	50.000	13.8
Tetrachloro-m-xylene	3.643	3.545	3.745	57.040	50.000	14.1

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Continuing Calib Date: 11/07/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 16:52 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.52	5.52	5.42	5.62	0.00
Aroclor-1016-2	(2)	5.55	5.55	5.45	5.65	0.00
Aroclor-1016-3	(3)	5.61	5.61	5.51	5.71	0.00
Aroclor-1016-4	(4)	5.71	5.70	5.60	5.80	-0.01
Aroclor-1016-5	(5)	6.00	6.00	5.90	6.10	0.00
Aroclor-1260-1	(1)	7.13	7.13	7.03	7.23	0.00
Aroclor-1260-2	(2)	7.39	7.38	7.28	7.48	-0.01
Aroclor-1260-3	(3)	7.75	7.74	7.64	7.84	-0.01
Aroclor-1260-4	(4)	7.98	7.97	7.87	8.07	-0.01
Aroclor-1260-5	(5)	8.29	8.28	8.18	8.38	-0.01
Tetrachloro-m-xylene		4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl		10.08	10.06	9.96	10.16	-0.02

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Continuing Calib Date: 11/07/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 16:52 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.72	4.73	4.63	4.83	0.01
Aroclor-1016-2	(2)	4.74	4.75	4.65	4.85	0.01
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.18	5.08	5.28	0.01
Aroclor-1260-1	(1)	6.21	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.00
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.64	3.65	3.55	3.75	0.01
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107784.D Time Analyzed: 16:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.524	5.423	5.623	465.890	500.000	-6.8
Aroclor-1016-2	5.546	5.445	5.645	452.740	500.000	-9.5
Aroclor-1016-3	5.610	5.507	5.707	444.240	500.000	-11.2
Aroclor-1016-4	5.706	5.604	5.804	472.470	500.000	-5.5
Aroclor-1016-5	6.002	5.899	6.099	494.720	500.000	-1.1
Aroclor-1260-1	7.132	7.026	7.226	469.850	500.000	-6.0
Aroclor-1260-2	7.389	7.282	7.482	493.320	500.000	-1.3
Aroclor-1260-3	7.752	7.643	7.843	493.890	500.000	-1.2
Aroclor-1260-4	7.977	7.868	8.068	522.330	500.000	4.5
Aroclor-1260-5	8.291	8.181	8.381	525.080	500.000	5.0
Decachlorobiphenyl	10.079	9.958	10.158	46.400	50.000	-7.2
Tetrachloro-m-xylene	4.372	4.274	4.474	56.540	50.000	13.1

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107784.D Time Analyzed: 16:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.723	4.626	4.826	572.960	500.000	14.6
Aroclor-1016-2	4.742	4.645	4.845	588.760	500.000	17.8
Aroclor-1016-3	4.918	4.821	5.021	566.130	500.000	13.2
Aroclor-1016-4	4.960	4.862	5.062	508.790	500.000	1.8
Aroclor-1016-5	5.173	5.076	5.276	562.090	500.000	12.4
Aroclor-1260-1	6.205	6.107	6.307	525.520	500.000	5.1
Aroclor-1260-2	6.392	6.294	6.494	543.960	500.000	8.8
Aroclor-1260-3	6.546	6.448	6.648	513.070	500.000	2.6
Aroclor-1260-4	7.017	6.918	7.118	479.570	500.000	-4.1
Aroclor-1260-5	7.258	7.159	7.359	492.600	500.000	-1.5
Decachlorobiphenyl	8.640	8.539	8.739	43.780	50.000	-12.4
Tetrachloro-m-xylene	3.642	3.545	3.745	57.900	50.000	15.8

Analytical Sequence

Client: Walsh Construction Company II, LLC	SDG No.: P4722
Project: NYCDEP C547A - Shafts 17B-1 & 18B-1 Sta;	Instrument ID: ECD_O
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 10/15/2024 10/15/2024

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/15/2024	18:08	PO107183.D	10.06	4.37
AR1660ICC1000	AR1660ICC1000	10/15/2024	18:27	PO107184.D	10.06	4.37
AR1660ICC750	AR1660ICC750	10/15/2024	18:45	PO107185.D	10.06	4.37
AR1660ICC500	AR1660ICC500	10/15/2024	19:03	PO107186.D	10.06	4.37
AR1660ICC250	AR1660ICC250	10/15/2024	19:21	PO107187.D	10.06	4.37
AR1660ICC050	AR1660ICC050	10/15/2024	19:39	PO107188.D	10.06	4.37
AR1221ICC500	AR1221ICC500	10/15/2024	19:57	PO107189.D	10.06	4.37
AR1232ICC500	AR1232ICC500	10/15/2024	20:15	PO107190.D	10.06	4.37
AR1242ICC1000	AR1242ICC1000	10/15/2024	20:34	PO107191.D	10.06	4.37
AR1242ICC750	AR1242ICC750	10/15/2024	20:52	PO107192.D	10.06	4.37
AR1242ICC500	AR1242ICC500	10/15/2024	21:10	PO107193.D	10.06	4.37
AR1242ICC250	AR1242ICC250	10/15/2024	21:28	PO107194.D	10.06	4.37
AR1242ICC050	AR1242ICC050	10/15/2024	21:46	PO107195.D	10.06	4.37
AR1248ICC1000	AR1248ICC1000	10/15/2024	22:04	PO107196.D	10.06	4.37
AR1248ICC750	AR1248ICC750	10/15/2024	22:22	PO107197.D	10.06	4.37
AR1248ICC500	AR1248ICC500	10/15/2024	22:41	PO107198.D	10.06	4.37
AR1248ICC250	AR1248ICC250	10/15/2024	22:59	PO107199.D	10.06	4.37
AR1248ICC050	AR1248ICC050	10/15/2024	23:17	PO107200.D	10.06	4.37
AR1254ICC1000	AR1254ICC1000	10/15/2024	23:35	PO107201.D	10.06	4.37
AR1254ICC750	AR1254ICC750	10/15/2024	23:53	PO107202.D	10.06	4.37
AR1254ICC500	AR1254ICC500	10/16/2024	00:11	PO107203.D	10.06	4.37
AR1254ICC250	AR1254ICC250	10/16/2024	00:29	PO107204.D	10.06	4.37
AR1254ICC050	AR1254ICC050	10/16/2024	00:47	PO107205.D	10.06	4.37
AR1262ICC500	AR1262ICC500	10/16/2024	01:05	PO107206.D	10.06	4.37
AR1268ICC1000	AR1268ICC1000	10/16/2024	01:23	PO107207.D	10.06	4.37
AR1268ICC750	AR1268ICC750	10/16/2024	01:41	PO107208.D	10.06	4.37
AR1268ICC500	AR1268ICC500	10/16/2024	01:59	PO107209.D	10.06	4.37
AR1268ICC250	AR1268ICC250	10/16/2024	02:18	PO107210.D	10.06	4.37
AR1268ICC050	AR1268ICC050	10/16/2024	02:36	PO107211.D	10.06	4.37
AR1660CCC500	AR1660CCC500	11/07/2024	09:00	PO107769.D	10.08	4.37
IBLK	IBLK	11/07/2024	10:06	PO107773.D	10.08	4.37
PB164748BL	PB164748BL	11/07/2024	12:54	PO107775.D	10.08	4.38
PB164748BS	PB164748BS	11/07/2024	13:10	PO107776.D	10.08	4.37
JC-701-COMP-01MS	P4720-01MS	11/07/2024	13:59	PO107779.D	10.08	4.37
JC-701-COMP-01MSD	P4720-01MSD	11/07/2024	14:15	PO107780.D	10.08	4.37
WC-1(0-6)	P4722-03	11/07/2024	14:31	PO107781.D	10.08	4.37
WC-2(0-6)	P4722-08	11/07/2024	14:47	PO107782.D	10.08	4.37
WC-3(0-6)	P4722-13	11/07/2024	15:04	PO107783.D	10.08	4.37
AR1660CCC500	AR1660CCC500	11/07/2024	16:52	PO107784.D	10.08	4.37
IBLK	IBLK	11/07/2024	17:56	PO107788.D	10.08	4.37

Analytical Sequence

Client: Walsh Construction Company II, LLC	SDG No.: P4722
Project: NYCDEP C547A - Shafts 17B-1 & 18B-1 Sta;	Instrument ID: ECD_O
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 10/15/2024 10/15/2024

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/15/2024	18:08	PO107183.D	8.64	3.64
AR1660ICC1000	AR1660ICC1000	10/15/2024	18:27	PO107184.D	8.64	3.64
AR1660ICC750	AR1660ICC750	10/15/2024	18:45	PO107185.D	8.64	3.64
AR1660ICC500	AR1660ICC500	10/15/2024	19:03	PO107186.D	8.64	3.65
AR1660ICC250	AR1660ICC250	10/15/2024	19:21	PO107187.D	8.64	3.64
AR1660ICC050	AR1660ICC050	10/15/2024	19:39	PO107188.D	8.64	3.64
AR1221ICC500	AR1221ICC500	10/15/2024	19:57	PO107189.D	8.64	3.64
AR1232ICC500	AR1232ICC500	10/15/2024	20:15	PO107190.D	8.64	3.64
AR1242ICC1000	AR1242ICC1000	10/15/2024	20:34	PO107191.D	8.64	3.64
AR1242ICC750	AR1242ICC750	10/15/2024	20:52	PO107192.D	8.64	3.64
AR1242ICC500	AR1242ICC500	10/15/2024	21:10	PO107193.D	8.64	3.64
AR1242ICC250	AR1242ICC250	10/15/2024	21:28	PO107194.D	8.64	3.65
AR1242ICC050	AR1242ICC050	10/15/2024	21:46	PO107195.D	8.64	3.64
AR1248ICC1000	AR1248ICC1000	10/15/2024	22:04	PO107196.D	8.64	3.64
AR1248ICC750	AR1248ICC750	10/15/2024	22:22	PO107197.D	8.64	3.64
AR1248ICC500	AR1248ICC500	10/15/2024	22:41	PO107198.D	8.64	3.64
AR1248ICC250	AR1248ICC250	10/15/2024	22:59	PO107199.D	8.64	3.64
AR1248ICC050	AR1248ICC050	10/15/2024	23:17	PO107200.D	8.64	3.64
AR1254ICC1000	AR1254ICC1000	10/15/2024	23:35	PO107201.D	8.64	3.65
AR1254ICC750	AR1254ICC750	10/15/2024	23:53	PO107202.D	8.64	3.64
AR1254ICC500	AR1254ICC500	10/16/2024	00:11	PO107203.D	8.64	3.64
AR1254ICC250	AR1254ICC250	10/16/2024	00:29	PO107204.D	8.64	3.64
AR1254ICC050	AR1254ICC050	10/16/2024	00:47	PO107205.D	8.64	3.65
AR1262ICC500	AR1262ICC500	10/16/2024	01:05	PO107206.D	8.64	3.64
AR1268ICC1000	AR1268ICC1000	10/16/2024	01:23	PO107207.D	8.64	3.65
AR1268ICC750	AR1268ICC750	10/16/2024	01:41	PO107208.D	8.64	3.64
AR1268ICC500	AR1268ICC500	10/16/2024	01:59	PO107209.D	8.64	3.65
AR1268ICC250	AR1268ICC250	10/16/2024	02:18	PO107210.D	8.64	3.64
AR1268ICC050	AR1268ICC050	10/16/2024	02:36	PO107211.D	8.64	3.65
AR1660CCC500	AR1660CCC500	11/07/2024	09:00	PO107769.D	8.64	3.64
IBLK	IBLK	11/07/2024	10:06	PO107773.D	8.64	3.64
PB164748BL	PB164748BL	11/07/2024	12:54	PO107775.D	8.64	3.64
PB164748BS	PB164748BS	11/07/2024	13:10	PO107776.D	8.64	3.64
JC-701-COMP-01MS	P4720-01MS	11/07/2024	13:59	PO107779.D	8.64	3.64
JC-701-COMP-01MSD	P4720-01MSD	11/07/2024	14:15	PO107780.D	8.64	3.64
WC-1(0-6)	P4722-03	11/07/2024	14:31	PO107781.D	8.64	3.64
WC-2(0-6)	P4722-08	11/07/2024	14:47	PO107782.D	8.64	3.64
WC-3(0-6)	P4722-13	11/07/2024	15:04	PO107783.D	8.64	3.64
AR1660CCC500	AR1660CCC500	11/07/2024	16:52	PO107784.D	8.64	3.64
IBLK	IBLK	11/07/2024	17:56	PO107788.D	8.64	3.64



QC SAMPLE DATA

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:		
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:		
Client Sample ID:	PB164748BL		SDG No.:	P4722	
Lab Sample ID:	PB164748BL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	100	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107775.D	1	11/07/24 08:40	11/07/24 12:54	PB164748

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.40	U	3.40	17.0	ug/kg
11104-28-2	Aroclor-1221	6.40	U	6.40	17.0	ug/kg
11141-16-5	Aroclor-1232	3.40	U	3.40	17.0	ug/kg
53469-21-9	Aroclor-1242	3.40	U	3.40	17.0	ug/kg
12672-29-6	Aroclor-1248	7.90	U	7.90	17.0	ug/kg
11097-69-1	Aroclor-1254	2.70	U	2.70	17.0	ug/kg
37324-23-5	Aroclor-1262	4.60	U	4.60	17.0	ug/kg
11100-14-4	Aroclor-1268	3.40	U	3.40	17.0	ug/kg
11096-82-5	Aroclor-1260	2.90	U	2.90	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.0		32 - 144	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.9		32 - 175	110%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	10/15/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	10/15/24
Client Sample ID:	PIBLK-PO107183.D	SDG No.:	P4722
Lab Sample ID:	I.BLK-PO107183.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107183.D	1		10/15/24	po101524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.15	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.23	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.37	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.16	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.12	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.15	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.12	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.1		60 - 140	110%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.8		60 - 140	114%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/07/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/07/24
Client Sample ID:	PIBLK-PO107773.D	SDG No.:	P4722
Lab Sample ID:	I.BLK-PO107773.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107773.D	1		11/07/24	PO110724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.15	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.23	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.37	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.16	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.12	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.15	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.12	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.8		60 - 140	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.9		60 - 140	109%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/07/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/07/24
Client Sample ID:	PIBLK-PO107788.D	SDG No.:	P4722
Lab Sample ID:	I.BLK-PO107788.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107788.D	1		11/07/24	PO110724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.15	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.23	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.37	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.16	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.12	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.15	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.12	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.4		60 - 140	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.6		60 - 140	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:		
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:		
Client Sample ID:	PB164748BS		SDG No.:	P4722	
Lab Sample ID:	PB164748BS		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	100	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107776.D	1	11/07/24 08:40	11/07/24 13:10	PB164748

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	163		3.40	17.0	ug/kg
11104-28-2	Aroclor-1221	6.40	U	6.40	17.0	ug/kg
11141-16-5	Aroclor-1232	3.40	U	3.40	17.0	ug/kg
53469-21-9	Aroclor-1242	3.40	U	3.40	17.0	ug/kg
12672-29-6	Aroclor-1248	7.90	U	7.90	17.0	ug/kg
11097-69-1	Aroclor-1254	2.70	U	2.70	17.0	ug/kg
37324-23-5	Aroclor-1262	4.60	U	4.60	17.0	ug/kg
11100-14-4	Aroclor-1268	3.40	U	3.40	17.0	ug/kg
11096-82-5	Aroclor-1260	167		2.90	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.3		32 - 144	107%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.2		32 - 175	116%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/05/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/05/24
Client Sample ID:	JC-701-COMP-01MS	SDG No.:	P4722
Lab Sample ID:	P4720-01MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.6
Sample Wt/Vol:	30.07 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107779.D	1	11/07/24 08:40	11/07/24 13:59	PB164748

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	203		3.70	18.7	ug/kg
11104-28-2	Aroclor-1221	7.10	U	7.10	18.7	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	18.7	ug/kg
53469-21-9	Aroclor-1242	3.70	U	3.70	18.7	ug/kg
12672-29-6	Aroclor-1248	8.70	U	8.70	18.7	ug/kg
11097-69-1	Aroclor-1254	3.00	U	3.00	18.7	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	18.7	ug/kg
11100-14-4	Aroclor-1268	3.80	U	3.80	18.7	ug/kg
11096-82-5	Aroclor-1260	232		3.20	18.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.1		32 - 144	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	29.5		32 - 175	148%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/05/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/05/24
Client Sample ID:	JC-701-COMP-01MSD	SDG No.:	P4722
Lab Sample ID:	P4720-01MSD	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.6
Sample Wt/Vol:	30.06 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO107780.D	1	11/07/24 08:40	11/07/24 14:15	PB164748

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	204		3.70	18.7	ug/kg
11104-28-2	Aroclor-1221	7.10	U	7.10	18.7	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	18.7	ug/kg
53469-21-9	Aroclor-1242	3.70	U	3.70	18.7	ug/kg
12672-29-6	Aroclor-1248	8.70	U	8.70	18.7	ug/kg
11097-69-1	Aroclor-1254	3.00	U	3.00	18.7	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	18.7	ug/kg
11100-14-4	Aroclor-1268	3.80	U	3.80	18.7	ug/kg
11096-82-5	Aroclor-1260	233		3.20	18.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.8		32 - 144	114%	SPK: 20
2051-24-3	Decachlorobiphenyl	29.8		32 - 175	149%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit