

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8082A

DataFile : PO107780.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JC-701-COMP-01MSD											
P4720-01MSD	AR1016	183.6	0	204	ug/kg	111		0		55	146	20
	AR1260	183.6	25.7	233	ug/kg	113		1		45	144	20