



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: **8082A** **Datafile :** PO107776.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB164748BS	AR1016	166.6	163	ug/kg	98				71	120		
	AR1260	166.6	167	ug/kg	100				65	130		