

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092888.D Time Analyzed: 11:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.713	6.610	6.810	49.380	50.000	-1.2
4,4'-DDE	6.195	6.093	6.293	46.780	50.000	-6.4
4,4'-DDT	7.026	6.923	7.123	46.720	50.000	-6.6
Aldrin	5.261	5.158	5.358	47.470	50.000	-5.1
alpha-BHC	3.998	3.896	4.096	48.980	50.000	-2.0
alpha-Chlordane	6.022	5.919	6.119	46.630	50.000	-6.7
beta-BHC	4.528	4.425	4.625	48.580	50.000	-2.8
Decachlorobiphenyl	9.059	8.954	9.154	46.020	50.000	-8.0
delta-BHC	4.775	4.672	4.872	48.390	50.000	-3.2
Dieldrin	6.347	6.245	6.445	45.990	50.000	-8.0
Endosulfan I	6.072	5.970	6.170	46.130	50.000	-7.7
Endosulfan II	6.797	6.694	6.894	44.890	50.000	-10.2
Endosulfan sulfate	7.161	7.058	7.258	45.940	50.000	-8.1
Endrin	6.577	6.474	6.674	45.550	50.000	-8.9
Endrin aldehyde	6.927	6.824	7.024	47.550	50.000	-4.9
Endrin ketone	7.647	7.543	7.743	45.940	50.000	-8.1
gamma-BHC (Lindane)	4.330	4.228	4.428	48.790	50.000	-2.4
gamma-Chlordane	5.943	5.841	6.041	46.570	50.000	-6.9
Heptachlor	4.919	4.817	5.017	47.760	50.000	-4.5
Heptachlor epoxide	5.686	5.584	5.784	47.690	50.000	-4.6
Methoxychlor	7.502	7.399	7.599	47.810	50.000	-4.4
Tetrachloro-m-xylene	3.542	3.440	3.640	48.470	50.000	-3.1

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092888.D Time Analyzed: 11:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.790	5.689	5.889	52.090	50.000	4.2
4,4'-DDE	5.235	5.135	5.335	50.820	50.000	1.6
4,4'-DDT	6.040	5.940	6.140	50.770	50.000	1.5
Aldrin	4.230	4.129	4.329	50.690	50.000	1.4
alpha-BHC	3.281	3.181	3.381	50.710	50.000	1.4
alpha-Chlordane	5.046	4.945	5.145	49.960	50.000	-0.1
beta-BHC	3.911	3.810	4.010	49.800	50.000	-0.4
Decachlorobiphenyl	7.917	7.816	8.016	46.840	50.000	-6.3
delta-BHC	4.140	4.039	4.239	51.070	50.000	2.1
Dieldrin	5.367	5.266	5.466	50.630	50.000	1.3
Endosulfan I	5.102	5.002	5.202	48.660	50.000	-2.7
Endosulfan II	5.936	5.836	6.036	51.680	50.000	3.4
Endosulfan sulfate	6.339	6.238	6.438	50.070	50.000	0.1
Endrin	5.642	5.541	5.741	51.530	50.000	3.1
Endrin aldehyde	6.116	6.015	6.215	49.630	50.000	-0.7
Endrin ketone	6.844	6.743	6.943	48.910	50.000	-2.2
gamma-BHC (Lindane)	3.611	3.511	3.711	50.380	50.000	0.8
gamma-Chlordane	4.982	4.882	5.082	50.080	50.000	0.2
Heptachlor	3.950	3.849	4.049	50.240	50.000	0.5
Heptachlor epoxide	4.732	4.632	4.832	50.240	50.000	0.5
Methoxychlor	6.615	6.515	6.715	49.970	50.000	-0.1
Tetrachloro-m-xylene	2.778	2.678	2.878	49.610	50.000	-0.8