

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092904.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.713	6.610	6.810	46.380	50.000	-7.2
4,4'-DDE	6.194	6.093	6.293	45.140	50.000	-9.7
4,4'-DDT	7.027	6.923	7.123	42.890	50.000	-14.2
Aldrin	5.259	5.158	5.358	45.730	50.000	-8.5
alpha-BHC	3.998	3.896	4.096	47.690	50.000	-4.6
alpha-Chlordane	6.022	5.919	6.119	45.000	50.000	-10.0
beta-BHC	4.528	4.425	4.625	47.930	50.000	-4.1
Decachlorobiphenyl	9.059	8.954	9.154	45.750	50.000	-8.5
delta-BHC	4.774	4.672	4.872	47.810	50.000	-4.4
Dieldrin	6.347	6.245	6.445	44.400	50.000	-11.2
Endosulfan I	6.072	5.970	6.170	44.450	50.000	-11.1
Endosulfan II	6.797	6.694	6.894	42.390	50.000	-15.2
Endosulfan sulfate	7.161	7.058	7.258	43.220	50.000	-13.6
Endrin	6.577	6.474	6.674	42.920	50.000	-14.2
Endrin aldehyde	6.927	6.824	7.024	44.140	50.000	-11.7
Endrin ketone	7.646	7.543	7.743	43.210	50.000	-13.6
gamma-BHC (Lindane)	4.330	4.228	4.428	48.200	50.000	-3.6
gamma-Chlordane	5.943	5.841	6.041	45.360	50.000	-9.3
Heptachlor	4.919	4.817	5.017	46.410	50.000	-7.2
Heptachlor epoxide	5.687	5.584	5.784	45.780	50.000	-8.4
Methoxychlor	7.503	7.399	7.599	43.270	50.000	-13.5
Tetrachloro-m-xylene	3.542	3.440	3.640	47.550	50.000	-4.9

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092904.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.790	5.689	5.889	49.190	50.000	-1.6
4,4'-DDE	5.235	5.135	5.335	50.300	50.000	0.6
4,4'-DDT	6.041	5.940	6.140	44.020	50.000	-12.0
Aldrin	4.230	4.129	4.329	50.680	50.000	1.4
alpha-BHC	3.281	3.181	3.381	51.060	50.000	2.1
alpha-Chlordane	5.046	4.945	5.145	48.510	50.000	-3.0
beta-BHC	3.911	3.810	4.010	49.650	50.000	-0.7
Decachlorobiphenyl	7.917	7.816	8.016	40.370	50.000	-19.3
delta-BHC	4.140	4.039	4.239	51.370	50.000	2.7
Dieldrin	5.367	5.266	5.466	48.730	50.000	-2.5
Endosulfan I	5.102	5.002	5.202	47.480	50.000	-5.0
Endosulfan II	5.937	5.836	6.036	47.060	50.000	-5.9
Endosulfan sulfate	6.339	6.238	6.438	46.300	50.000	-7.4
Endrin	5.642	5.541	5.741	46.330	50.000	-7.3
Endrin aldehyde	6.116	6.015	6.215	44.110	50.000	-11.8
Endrin ketone	6.845	6.743	6.943	43.580	50.000	-12.8
gamma-BHC (Lindane)	3.611	3.511	3.711	49.880	50.000	-0.2
gamma-Chlordane	4.983	4.882	5.082	48.040	50.000	-3.9
Heptachlor	3.950	3.849	4.049	49.860	50.000	-0.3
Heptachlor epoxide	4.732	4.632	4.832	47.410	50.000	-5.2
Methoxychlor	6.616	6.515	6.715	44.210	50.000	-11.6
Tetrachloro-m-xylene	2.778	2.678	2.878	50.400	50.000	0.8