

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL03 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092912.D Time Analyzed: 17:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.712	6.610	6.810	47.210	50.000	-5.6
4,4'-DDE	6.195	6.093	6.293	45.000	50.000	-10.0
4,4'-DDT	7.025	6.923	7.123	43.020	50.000	-14.0
Aldrin	5.259	5.158	5.358	45.590	50.000	-8.8
alpha-BHC	3.998	3.896	4.096	48.270	50.000	-3.5
alpha-Chlordane	6.021	5.919	6.119	44.800	50.000	-10.4
beta-BHC	4.528	4.425	4.625	47.790	50.000	-4.4
Decachlorobiphenyl	9.058	8.954	9.154	44.670	50.000	-10.7
delta-BHC	4.775	4.672	4.872	47.530	50.000	-4.9
Dieldrin	6.347	6.245	6.445	44.420	50.000	-11.2
Endosulfan I	6.072	5.970	6.170	44.500	50.000	-11.0
Endosulfan II	6.796	6.694	6.894	42.710	50.000	-14.6
Endosulfan sulfate	7.161	7.058	7.258	43.250	50.000	-13.5
Endrin	6.576	6.474	6.674	43.250	50.000	-13.5
Endrin aldehyde	6.926	6.824	7.024	43.230	50.000	-13.5
Endrin ketone	7.646	7.543	7.743	43.280	50.000	-13.4
gamma-BHC (Lindane)	4.330	4.228	4.428	48.250	50.000	-3.5
gamma-Chlordane	5.942	5.841	6.041	45.330	50.000	-9.3
Heptachlor	4.918	4.817	5.017	46.270	50.000	-7.5
Heptachlor epoxide	5.685	5.584	5.784	45.430	50.000	-9.1
Methoxychlor	7.502	7.399	7.599	43.320	50.000	-13.4
Tetrachloro-m-xylene	3.542	3.440	3.640	48.110	50.000	-3.8

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL03 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092912.D Time Analyzed: 17:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.790	5.689	5.889	51.020	50.000	2.0
4,4'-DDE	5.235	5.135	5.335	50.380	50.000	0.8
4,4'-DDT	6.041	5.940	6.140	45.260	50.000	-9.5
Aldrin	4.229	4.129	4.329	51.110	50.000	2.2
alpha-BHC	3.281	3.181	3.381	51.580	50.000	3.2
alpha-Chlordane	5.046	4.945	5.145	49.410	50.000	-1.2
beta-BHC	3.911	3.810	4.010	50.540	50.000	1.1
Decachlorobiphenyl	7.917	7.816	8.016	41.590	50.000	-16.8
delta-BHC	4.139	4.039	4.239	51.690	50.000	3.4
Dieldrin	5.366	5.266	5.466	48.970	50.000	-2.1
Endosulfan I	5.102	5.002	5.202	47.480	50.000	-5.0
Endosulfan II	5.937	5.836	6.036	47.730	50.000	-4.5
Endosulfan sulfate	6.339	6.238	6.438	46.530	50.000	-6.9
Endrin	5.642	5.541	5.741	48.020	50.000	-4.0
Endrin aldehyde	6.116	6.015	6.215	45.950	50.000	-8.1
Endrin ketone	6.845	6.743	6.943	45.130	50.000	-9.7
gamma-BHC (Lindane)	3.611	3.511	3.711	50.980	50.000	2.0
gamma-Chlordane	4.982	4.882	5.082	49.210	50.000	-1.6
Heptachlor	3.950	3.849	4.049	50.550	50.000	1.1
Heptachlor epoxide	4.732	4.632	4.832	49.550	50.000	-0.9
Methoxychlor	6.615	6.515	6.715	44.500	50.000	-11.0
Tetrachloro-m-xylene	2.778	2.678	2.878	50.440	50.000	0.9