

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL04 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092916.D Time Analyzed: 18:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.712	6.610	6.810	47.100	50.000	-5.8
4,4'-DDE	6.194	6.093	6.293	45.860	50.000	-8.3
4,4'-DDT	7.026	6.923	7.123	43.250	50.000	-13.5
Aldrin	5.259	5.158	5.358	46.420	50.000	-7.2
alpha-BHC	3.998	3.896	4.096	48.340	50.000	-3.3
alpha-Chlordane	6.021	5.919	6.119	45.360	50.000	-9.3
beta-BHC	4.528	4.425	4.625	47.410	50.000	-5.2
Decachlorobiphenyl	9.057	8.954	9.154	44.800	50.000	-10.4
delta-BHC	4.775	4.672	4.872	47.740	50.000	-4.5
Dieldrin	6.347	6.245	6.445	45.110	50.000	-9.8
Endosulfan I	6.072	5.970	6.170	45.100	50.000	-9.8
Endosulfan II	6.797	6.694	6.894	44.070	50.000	-11.9
Endosulfan sulfate	7.161	7.058	7.258	43.920	50.000	-12.2
Endrin	6.577	6.474	6.674	43.630	50.000	-12.7
Endrin aldehyde	6.926	6.824	7.024	45.190	50.000	-9.6
Endrin ketone	7.645	7.543	7.743	43.770	50.000	-12.5
gamma-BHC (Lindane)	4.330	4.228	4.428	47.980	50.000	-4.0
gamma-Chlordane	5.942	5.841	6.041	45.440	50.000	-9.1
Heptachlor	4.917	4.817	5.017	46.490	50.000	-7.0
Heptachlor epoxide	5.686	5.584	5.784	45.510	50.000	-9.0
Methoxychlor	7.501	7.399	7.599	44.260	50.000	-11.5
Tetrachloro-m-xylene	3.542	3.440	3.640	47.900	50.000	-4.2

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL04 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092916.D Time Analyzed: 18:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.790	5.689	5.889	51.850	50.000	3.7
4,4'-DDE	5.235	5.135	5.335	50.640	50.000	1.3
4,4'-DDT	6.040	5.940	6.140	46.830	50.000	-6.3
Aldrin	4.230	4.129	4.329	50.860	50.000	1.7
alpha-BHC	3.281	3.181	3.381	51.420	50.000	2.8
alpha-Chlordane	5.046	4.945	5.145	50.260	50.000	0.5
beta-BHC	3.910	3.810	4.010	50.530	50.000	1.1
Decachlorobiphenyl	7.917	7.816	8.016	43.460	50.000	-13.1
delta-BHC	4.139	4.039	4.239	51.360	50.000	2.7
Dieldrin	5.366	5.266	5.466	49.910	50.000	-0.2
Endosulfan I	5.102	5.002	5.202	47.240	50.000	-5.5
Endosulfan II	5.937	5.836	6.036	49.040	50.000	-1.9
Endosulfan sulfate	6.338	6.238	6.438	47.640	50.000	-4.7
Endrin	5.642	5.541	5.741	50.320	50.000	0.6
Endrin aldehyde	6.116	6.015	6.215	47.690	50.000	-4.6
Endrin ketone	6.844	6.743	6.943	47.930	50.000	-4.1
gamma-BHC (Lindane)	3.611	3.511	3.711	51.220	50.000	2.4
gamma-Chlordane	4.982	4.882	5.082	50.410	50.000	0.8
Heptachlor	3.950	3.849	4.049	50.110	50.000	0.2
Heptachlor epoxide	4.733	4.632	4.832	50.460	50.000	0.9
Methoxychlor	6.615	6.515	6.715	45.360	50.000	-9.3
Tetrachloro-m-xylene	2.778	2.678	2.878	50.430	50.000	0.9