

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

DataFile : PL092900.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JC-701-COMP-01MS											
P4720-01MS	alpha-BHC	18.35	0	16.4	ug/kg	89				60	144	
	beta-BHC	18.35	0	19.0	ug/kg	104				54	143	
	delta-BHC	18.35	0	4.60	ug/kg	25	*			47	144	
	gamma-BHC (Lindane)	18.35	0	16.7	ug/kg	91				61	140	
	Heptachlor	18.35	0	19.3	ug/kg	105				63	135	
	Aldrin	18.35	0	18.3	ug/kg	100				49	139	
	Heptachlor epoxide	18.35	0	18.6	ug/kg	101				32	180	
	Endosulfan I	18.35	0	17.9	ug/kg	98				56	142	
	Dieldrin	18.35	0	20.8	ug/kg	113				47	161	
	4,4'-DDE	18.35	7.3	25.7	ug/kg	100				55	136	
	Endrin	18.35	0	21.7	ug/kg	118				57	139	
	Endosulfan II	18.35	0	20.0	ug/kg	109				40	163	
	4,4'-DDD	18.35	0	20.3	ug/kg	111				37	192	
	Endosulfan sulfate	18.35	0	15.7	ug/kg	86				62	139	
	4,4'-DDT	18.35	14.8	31.9	ug/kg	93				51	146	
	Methoxychlor	18.35	0	18.1	ug/kg	99				54	136	
	Endrin ketone	18.35	0	18.0	ug/kg	98				60	129	
	Endrin aldehyde	18.35	0	18.5	ug/kg	101				59	132	
	alpha-Chlordane	18.35	0	19.2	ug/kg	105				30	192	
	gamma-Chlordane	18.35	0	20.2	ug/kg	110				44	175	