

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

DataFile : PL092901.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JC-701-COMP-01MSD											
P4720-01MSD	alpha-BHC	18.37	0	16.7	ug/kg	91		2		60	144	20
	beta-BHC	18.37	0	19.6	ug/kg	107		3		54	143	20
	delta-BHC	18.37	0	4.60	ug/kg	25	*	0		47	144	20
	gamma-BHC (Lindane)	18.37	0	17.1	ug/kg	93		2		61	140	20
	Heptachlor	18.37	0	19.7	ug/kg	107		2		63	135	20
	Aldrin	18.37	0	18.8	ug/kg	102		2		49	139	20
	Heptachlor epoxide	18.37	0	19.0	ug/kg	103		2		32	180	20
	Endosulfan I	18.37	0	18.1	ug/kg	99		1		56	142	20
	Dieldrin	18.37	0	21.4	ug/kg	116		3		47	161	20
	4,4'-DDE	18.37	7.3	26.2	ug/kg	103		3		55	136	20
	Endrin	18.37	0	22.0	ug/kg	120		2		57	139	20
	Endosulfan II	18.37	0	20.5	ug/kg	112		3		40	163	20
	4,4'-DDD	18.37	0	20.5	ug/kg	112		1		37	192	20
	Endosulfan sulfate	18.37	0	16.1	ug/kg	88		2		62	139	20
	4,4'-DDT	18.37	14.8	32.5	ug/kg	96		3		51	146	20
	Methoxychlor	18.37	0	19.4	ug/kg	106		7		54	136	20
	Endrin ketone	18.37	0	18.5	ug/kg	101		3		60	129	20
	Endrin aldehyde	18.37	0	18.8	ug/kg	102		1		59	132	20
	alpha-Chlordane	18.37	0	19.7	ug/kg	107		2		30	192	20
	gamma-Chlordane	18.37	0	20.8	ug/kg	113		3		44	175	20