

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Instrument ID: ECD_L Date(s) Analyzed: 10/28/2024 10/28/2024

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	23962400
		2	6.44	6.34	6.54	13823600
		3	7.06	6.96	7.16	79159800
		4	7.15	7.05	7.25	59803700
		5	7.93	7.83	8.03	45329200

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: WALS01

Lab Code: CHEM **Case No.:** P4722 **SAS No.:** P4722 **SDG NO.:** P4722

Instrument ID: ECD_L **Date(s) Analyzed:** 10/28/2024 10/28/2024

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.01	4.91	5.11	19952700
		2	5.33	5.23	5.43	19749600
		3	6.61	6.51	6.71	70222500
		4	6.73	6.63	6.83	98337700
		5	7.05	6.95	7.15	65479700