

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092973.D Time Analyzed: 20:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.715	6.610	6.810	50.780	50.000	1.6
4,4'-DDE	6.197	6.093	6.293	47.780	50.000	-4.4
4,4'-DDT	7.029	6.923	7.123	41.420	50.000	-17.2
Aldrin	5.262	5.158	5.358	47.210	50.000	-5.6
alpha-BHC	3.999	3.896	4.096	49.140	50.000	-1.7
alpha-Chlordane	6.023	5.919	6.119	46.880	50.000	-6.2
beta-BHC	4.529	4.425	4.625	48.080	50.000	-3.8
Decachlorobiphenyl	9.063	8.954	9.154	43.000	50.000	-14.0
delta-BHC	4.776	4.672	4.872	48.960	50.000	-2.1
Dieldrin	6.349	6.245	6.445	46.620	50.000	-6.8
Endosulfan I	6.074	5.970	6.170	46.400	50.000	-7.2
Endosulfan II	6.798	6.694	6.894	44.940	50.000	-10.1
Endosulfan sulfate	7.164	7.058	7.258	45.660	50.000	-8.7
Endrin	6.579	6.474	6.674	43.010	50.000	-14.0
Endrin aldehyde	6.929	6.824	7.024	46.140	50.000	-7.7
Endrin ketone	7.649	7.543	7.743	45.990	50.000	-8.0
gamma-BHC (Lindane)	4.331	4.228	4.428	48.520	50.000	-3.0
gamma-Chlordane	5.945	5.841	6.041	46.960	50.000	-6.1
Heptachlor	4.920	4.817	5.017	45.380	50.000	-9.2
Heptachlor epoxide	5.688	5.584	5.784	47.490	50.000	-5.0
Methoxychlor	7.505	7.399	7.599	42.350	50.000	-15.3
Tetrachloro-m-xylene	3.542	3.440	3.640	49.410	50.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092973.D Time Analyzed: 20:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.792	5.689	5.889	54.900	50.000	9.8
4,4'-DDE	5.236	5.135	5.335	51.590	50.000	3.2
4,4'-DDT	6.042	5.940	6.140	43.440	50.000	-13.1
Aldrin	4.231	4.129	4.329	51.190	50.000	2.4
alpha-BHC	3.282	3.181	3.381	51.560	50.000	3.1
alpha-Chlordane	5.047	4.945	5.145	50.130	50.000	0.3
beta-BHC	3.912	3.810	4.010	51.180	50.000	2.4
Decachlorobiphenyl	7.920	7.816	8.016	44.340	50.000	-11.3
delta-BHC	4.141	4.039	4.239	52.690	50.000	5.4
Dieldrin	5.368	5.266	5.466	50.660	50.000	1.3
Endosulfan I	5.103	5.002	5.202	47.530	50.000	-4.9
Endosulfan II	5.939	5.836	6.036	49.470	50.000	-1.1
Endosulfan sulfate	6.341	6.238	6.438	49.630	50.000	-0.7
Endrin	5.644	5.541	5.741	48.560	50.000	-2.9
Endrin aldehyde	6.118	6.015	6.215	49.070	50.000	-1.9
Endrin ketone	6.847	6.743	6.943	49.470	50.000	-1.1
gamma-BHC (Lindane)	3.612	3.511	3.711	50.790	50.000	1.6
gamma-Chlordane	4.984	4.882	5.082	50.590	50.000	1.2
Heptachlor	3.951	3.849	4.049	48.630	50.000	-2.7
Heptachlor epoxide	4.734	4.632	4.832	50.070	50.000	0.1
Methoxychlor	6.618	6.515	6.715	43.670	50.000	-12.7
Tetrachloro-m-xylene	2.779	2.678	2.878	50.370	50.000	0.7