

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Instrument ID: ECD_S Calibration Date(s): 11/08/2024 11/08/2024

Calibration Times: 15:13 16:58

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028356.D</u>	CF 500 =	<u>PS028357.D</u>		
CF 750 =		<u>PS028358.D</u>	CF 1000 =	<u>PS028359.D</u>	CF 1500 =	<u>PS028360.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	21008200000	18982300000	19068100000	18267500000	17318700000	18929000000	7
2,4-D	3997770000	3571450000	3468690000	3334460000	3198650000	3514210000	9
2,4-DCAA	3196830000	2740430000	2735520000	2619650000	2511350000	2760750000	9

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

Instrument ID: ECD_S Calibration Date(s): 11/08/2024 11/08/2024
Calibration Times: 15:13 16:58

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028356.D</u>	CF 500 =	<u>PS028357.D</u>		
CF 750 =		<u>PS028358.D</u>	CF 1000 =	<u>PS028359.D</u>	CF 1500 =	<u>PS028360.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	10970300000	10612300000	10768700000	10478800000	10130800000	10592200000	3
2,4-D	2051780000	1901200000	1935300000	1900220000	1880580000	1933820000	4
2,4-DCAA	1694540000	1575990000	1595150000	1561040000	1540400000	1593420000	4