

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028463.D Time Analyzed: 18:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.253	9.160	9.360	730.330	712.500	2.5
2,4-D	8.372	8.279	8.479	706.140	705.000	0.2
2,4-DCAA	7.243	7.150	7.350	746.010	750.000	-0.5

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM **Case No.:** P4722 **SAS No.:** P4722 **SDG NO.:** P4722

GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/08/2024 11/08/2024

Client Sample No.: CCAL02 **Date Analyzed:** 11/12/2024

Lab Sample No.: HSTDCCC750 **Data File :** PS028463.D **Time Analyzed:** 18:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.901	9.811	10.011	663.190	712.500	-6.9
2,4-D	8.994	8.905	9.105	643.750	705.000	-8.7
2,4-DCAA	7.746	7.655	7.855	675.910	750.000	-9.9