

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111224\
 Data File : PS028458.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2024 16:27
 Operator : AR\AJ
 Sample : P4718-03MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WB-307-SB02MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/13/2024
 Supervised By :Ankita Jodhani 11/13/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 23:53:43 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 08 17:46:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	7.746	1010.9E6	449.7E6	366.165	282.219
Target Compounds							
1) T	Dalapon	2.645	2.707	1085.7E6	883.8E6	304.269	311.686m
2) T	3,5-DICHL...	6.413	6.699	1637.7E6	909.6E6	409.275	383.951m
5) T	DICAMBA	7.431	7.947	4745.0E6	2525.0E6	394.077	352.182
6) T	MCPPE	7.612	8.048	268.8E6	197.0E6	32.870	33.531
7) T	MCPA	7.762	8.292	447.0E6	315.8E6	39.871	39.434
8) T	DICHLORPROP	8.141	8.664	1329.8E6	658.3E6	429.020	369.584
9) T	2,4-D	8.372	8.994	1679.5E6	824.8E6	477.904	426.508m
10) T	Pentachlo...	8.671	9.524	16218.0E6	7515.5E6	360.063	301.622
11) T	2,4,5-TP ...	9.253	9.905	8739.5E6	6210.7E6	461.699	586.342 #
12) T	2,4,5-T	9.546	10.323	8504.4E6	4019.3E6	438.799	387.361
13) T	2,4-DB	10.123	10.890	1029.9E6	419.1E6	333.195	316.242m
14) T	DINOSEB	11.337	11.269	4050.7E6	1558.4E6	249.839	217.126
15) T	Picloram	11.142	12.367	11606.3E6	4702.5E6	369.556	326.732
16) T	DCPA	11.629	12.315	11355.7E6	4797.6E6	424.885	427.974

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111224\
Data File : PS028458.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2024 16:27
Operator : AR\AJ
Sample : P4718-03MSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
WB-307-SB02MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/13/2024
Supervised By :Ankita Jodhani 11/13/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 23:53:43 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
Quant Title : 8080.M
QLast Update : Fri Nov 08 17:46:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

