

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110924

Review By	Abdul	Review On	11/11/2024 5:07:36 PM
Supervise By	Ankita	Supervise On	11/12/2024 8:59:40 AM
SubDirectory	PS110924	HP Acquire Method	HP Processing Method ps110924 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC	PP23462
Internal Standard/PEM	
ICV/I.BLK	PP23469
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028354.D	08 Nov 2024 14:00		AR\AJ	Ok
2	I.BLK	I.BLK	PS028355.D	08 Nov 2024 14:25		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS028356.D	08 Nov 2024 15:13		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS028357.D	08 Nov 2024 15:46		AR\AJ	Ok,M
5	HSTDICC750	HSTDICC750	PS028358.D	08 Nov 2024 16:10		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS028359.D	08 Nov 2024 16:34		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS028360.D	08 Nov 2024 16:58		AR\AJ	Ok
8	HSTDICV750	ICVPS110924	PS028361.D	08 Nov 2024 17:22		AR\AJ	Ok
9	I.BLK	I.BLK	PS028362.D	08 Nov 2024 17:46		AR\AJ	Ok,M
10	HSTDCCC750	HSTDCCC750	PS028363.D	08 Nov 2024 18:10		AR\AJ	Ok
11	P4700-01	MH-8	PS028364.D	08 Nov 2024 18:34		AR\AJ	Ok,M
12	P4701-01	BP-F3	PS028365.D	08 Nov 2024 18:58		AR\AJ	Ok,M
13	P4701-05	BP-F4	PS028366.D	08 Nov 2024 19:22		AR\AJ	Ok
14	PB164754BL	PB164754BL	PS028367.D	08 Nov 2024 19:46		AR\AJ	Ok
15	PB164754BS	PB164754BS	PS028368.D	08 Nov 2024 20:10		AR\AJ	Ok,M
16	P4722-03	WC-1(0-6)	PS028369.D	08 Nov 2024 20:34	2,4-DCAA low in both column	AR\AJ	Not Ok
17	P4722-08	WC-2(0-6)	PS028370.D	08 Nov 2024 20:58		AR\AJ	Ok,M
18	P4722-13	WC-3(0-6)	PS028371.D	08 Nov 2024 21:22	2,4-DCAA Low in both column	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110924

Review By	Abdul	Review On	11/11/2024 5:07:36 PM
Supervise By	Ankita	Supervise On	11/12/2024 8:59:40 AM
SubDirectory	PS110924	HP Acquire Method	HP Processing Method ps110924 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM	PP23469		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P4739-01	TP-14	PS028372.D	08 Nov 2024 21:46		AR\AJ	Ok,M
20	P4739-01MS	TP-14MS	PS028373.D	08 Nov 2024 22:10	Some compound recovery fail	AR\AJ	Ok,M
21	P4739-01MSD	TP-14MSD	PS028374.D	08 Nov 2024 22:34	Some compound recovery fail	AR\AJ	Ok,M
22	P4739-05	BP-G2	PS028375.D	08 Nov 2024 22:58		AR\AJ	Ok
23	P4739-09	BP-B2	PS028376.D	08 Nov 2024 23:22		AR\AJ	Ok,M
24	P4739-13	TP-11	PS028377.D	08 Nov 2024 23:46		AR\AJ	Ok,M
25	I.BLK	I.BLK	PS028378.D	09 Nov 2024 00:10		AR\AJ	Ok
26	HSTDCCC750	HSTDCCC750	PS028379.D	09 Nov 2024 00:34		AR\AJ	Ok
27	P4601-24RE	CC0P1RE	PS028380.D	09 Nov 2024 00:58	Closing CCC fail,	AR\AJ	Not Ok
28	P4601-25RE	CC0P3RE	PS028381.D	09 Nov 2024 01:22	Closing CCC fail,	AR\AJ	Not Ok
29	P4601-26RE	CC0P5RE	PS028382.D	09 Nov 2024 01:46	Closing CCC fail,	AR\AJ	Not Ok
30	P4602-43	CC0R8	PS028383.D	09 Nov 2024 02:10	, Closing CCC fail	AR\AJ	Not Ok
31	P4602-44MS	CC0R8MS	PS028384.D	09 Nov 2024 02:34	Some compound recovery fail, Closing CCC fail	AR\AJ	Not Ok
32	P4602-45MSD	CC0R8MSD	PS028385.D	09 Nov 2024 02:58	Some compound recovery fail, Closing CCC fail	AR\AJ	Not Ok
33	P4602-40	CC0R6	PS028386.D	09 Nov 2024 03:22	Closing CCC fail	AR\AJ	Not Ok
34	P4602-41MS	CC0R6MS	PS028387.D	09 Nov 2024 03:46	Some compound recovery fail, Closing CCC fail	AR\AJ	Not Ok
35	P4602-42MSD	CC0R6MSD	PS028388.D	09 Nov 2024 04:10	Some compound recovery fail, RPD fail, Closing CCC fail	AR\AJ	Not Ok
36	PB164393TB	PB164393TB	PS028389.D	09 Nov 2024 04:34	Closing CCC fail	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110924

Review By	Abdul	Review On	11/11/2024 5:07:36 PM
Supervise By	Ankita	Supervise On	11/12/2024 8:59:40 AM
SubDirectory	PS110924	HP Acquire Method	HP Processing Method ps110924 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC Internal Standard/PEM	PP23462
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469

37	I.BLK	I.BLK	PS028390.D	09 Nov 2024 04:58	Closing CCC fail	AR\AJ	Not Ok
38	HSTDCCC750	HSTDCCC750	PS028391.D	09 Nov 2024 05:22	CCC fail	AR\AJ	Not Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111224

Review By	Abdul	Review On	11/13/2024 9:35:41 AM
Supervise By	Ankita	Supervise On	11/13/2024 10:39:11 AM
SubDirectory	PS111224	HP Acquire Method	HP Processing Method ps110924 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC	PP23462
Internal Standard/PEM	
ICV/I.BLK	PP23469
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028449.D	12 Nov 2024 12:36		AR\AJ	Ok
2	I.BLK	I.BLK	PS028450.D	12 Nov 2024 13:00		AR\AJ	Ok,M
3	HSTDCCC750	HSTDCCC750	PS028451.D	12 Nov 2024 13:25		AR\AJ	Ok,M
4	PB164850BL	PB164850BL	PS028452.D	12 Nov 2024 13:53		AR\AJ	Ok
5	PB164850BS	PB164850BS	PS028453.D	12 Nov 2024 14:17		AR\AJ	Ok,M
6	PB164694TB	PB164694TB	PS028454.D	12 Nov 2024 14:49		AR\AJ	Ok
7	PB164793TB	PB164793TB	PS028455.D	12 Nov 2024 15:14		AR\AJ	Ok
8	P4718-03	WB-307-SB02	PS028456.D	12 Nov 2024 15:38		AR\AJ	Ok
9	P4718-03MS	WB-307-SB02MS	PS028457.D	12 Nov 2024 16:03	Some compound recovery fail	AR\AJ	Ok,M
10	P4718-03MSD	WB-307-SB02MSD	PS028458.D	12 Nov 2024 16:27	Some compound recovery fail	AR\AJ	Ok,M
11	P4722-04	WC-1(0-6)	PS028459.D	12 Nov 2024 16:51		AR\AJ	Ok,M
12	P4722-09	WC-2(0-6)	PS028460.D	12 Nov 2024 17:16		AR\AJ	Ok
13	P4722-14	WC-3(0-6)	PS028461.D	12 Nov 2024 17:41		AR\AJ	Ok,M
14	I.BLK	I.BLK	PS028462.D	12 Nov 2024 18:04		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS028463.D	12 Nov 2024 18:29		AR\AJ	Ok,M
16	P4732-02	PPE-COMP	PS028464.D	12 Nov 2024 19:18		AR\AJ	Ok
17	P4771-06	P-1	PS028465.D	12 Nov 2024 19:43		AR\AJ	Ok
18	P4771-08	P-2	PS028466.D	12 Nov 2024 20:08		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111224

Review By	Abdul	Review On	11/13/2024 9:35:41 AM
Supervise By	Ankita	Supervise On	11/13/2024 10:39:11 AM
SubDirectory	PS111224	HP Acquire Method	HP Processing Method ps110924 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB164908BL	PB164908BL	PS028467.D	12 Nov 2024 20:32		AR\AJ	Ok
20	PB164908BS	PB164908BS	PS028468.D	12 Nov 2024 20:56		AR\AJ	Ok,M
21	PB164904TB	PB164904TB	PS028469.D	12 Nov 2024 21:21		AR\AJ	Ok
22	P4722-05	WC-1(0-6)	PS028470.D	12 Nov 2024 21:45		AR\AJ	Ok
23	P4722-10	WC-2(0-6)	PS028471.D	12 Nov 2024 22:09		AR\AJ	Ok
24	I.BLK	I.BLK	PS028472.D	12 Nov 2024 22:34		AR\AJ	Ok,M
25	HSTDCCC750	HSTDCCC750	PS028473.D	12 Nov 2024 22:57		AR\AJ	Ok,M
26	P4722-15	WC-3(0-6)	PS028474.D	12 Nov 2024 23:46		AR\AJ	Ok
27	P4722-10MS	WC-2(0-6)MS	PS028475.D	13 Nov 2024 00:10		AR\AJ	Ok,M
28	P4722-10MSD	WC-2(0-6)MSD	PS028476.D	13 Nov 2024 00:34		AR\AJ	Ok,M
29	PB164910BL	PB164910BL	PS028477.D	13 Nov 2024 00:59		AR\AJ	Ok
30	PB164910BS	PB164910BS	PS028478.D	13 Nov 2024 01:23		AR\AJ	Ok,M
31	PB164910BSD	PB164910BSD	PS028479.D	13 Nov 2024 01:48		AR\AJ	Ok,M
32	P4791-02	HINCHMAN-OILY-WAT	PS028480.D	13 Nov 2024 02:12		AR\AJ	Ok
33	P4792-04	MANHOLE-WASTE-DR	PS028481.D	13 Nov 2024 02:36		AR\AJ	Ok,M
34	I.BLK	I.BLK	PS028482.D	13 Nov 2024 03:01		AR\AJ	Ok
35	HSTDCCC750	HSTDCCC750	PS028483.D	13 Nov 2024 04:14		AR\AJ	Ok,M

M : Manual Integration