

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL03 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092973.D Time Analyzed: 20:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.063	8.954	9.154	43.000	50.000	-14.0
Endrin	6.579	6.474	6.674	43.010	50.000	-14.0
gamma-BHC (Lindane)	4.331	4.228	4.428	48.520	50.000	-3.0
Heptachlor	4.920	4.817	5.017	45.380	50.000	-9.2
Heptachlor epoxide	5.688	5.584	5.784	47.490	50.000	-5.0
Methoxychlor	7.505	7.399	7.599	42.350	50.000	-15.3
Tetrachloro-m-xylene	3.542	3.440	3.640	49.410	50.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL03 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092973.D Time Analyzed: 20:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.920	7.816	8.016	44.340	50.000	-11.3
Endrin	5.644	5.541	5.741	48.560	50.000	-2.9
gamma-BHC (Lindane)	3.612	3.511	3.711	50.790	50.000	1.6
Heptachlor	3.951	3.849	4.049	48.630	50.000	-2.7
Heptachlor epoxide	4.734	4.632	4.832	50.070	50.000	0.1
Methoxychlor	6.618	6.515	6.715	43.670	50.000	-12.7
Tetrachloro-m-xylene	2.779	2.678	2.878	50.370	50.000	0.7