

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112124\
 Data File : BF140536.D
 Acq On : 21 Nov 2024 15:07
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF112124

Quant Time: Nov 21 16:11:44 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	39.036	2.4	98	-0.01
3	Pyridine	40.000	44.701	-11.8	101	-0.01
4	n-Nitrosodimethylamine	40.000	41.184	-3.0	101	-0.01
5 S	2-Fluorophenol	80.000	79.009	1.2	98	0.00
6	Aniline	40.000	47.067	-17.7	102	0.00
7 S	Phenol-d6	80.000	81.459	-1.8	98	0.00
8	2-Chlorophenol	40.000	40.191	-0.5	98	0.00
9	Benzaldehyde	40.000	36.939	7.7	102	0.00
10 C	Phenol	40.000	41.030	-2.6	97	0.00
11	bis(2-Chloroethyl)ether	40.000	39.796	0.5	96	0.00
12	1,3-Dichlorobenzene	40.000	39.732	0.7	100	0.00
13 C	1,4-Dichlorobenzene	40.000	39.668	0.8	99	0.00
14	1,2-Dichlorobenzene	40.000	39.406	1.5	97	0.00
15	Benzyl Alcohol	40.000	41.993	-5.0	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.183	-3.0	100	0.00
17	2-Methylphenol	40.000	41.606	-4.0	100	0.00
18	Hexachloroethane	40.000	40.108	-0.3	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.038	-5.1	101	0.00
20	3+4-Methylphenols	40.000	41.536	-3.8	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	101	0.00
22	Acetophenone	40.000	38.963	2.6	100	0.00
23 S	Nitrobenzene-d5	80.000	78.907	1.4	99	0.00
24	Nitrobenzene	40.000	39.127	2.2	99	0.00
25	Isophorone	40.000	40.332	-0.8	100	0.00
26 C	2-Nitrophenol	40.000	40.372	-0.9	101	0.00
27	2,4-Dimethylphenol	40.000	39.568	1.1	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.938	0.2	101	0.00
29 C	2,4-Dichlorophenol	40.000	39.952	0.1	102	0.00
30	1,2,4-Trichlorobenzene	40.000	38.901	2.7	101	0.00
31	Naphthalene	40.000	39.152	2.1	101	0.00
32	Benzoic acid	40.000	41.429	-3.6	107	0.00
33	4-Chloroaniline	40.000	43.551	-8.9	104	0.00
34 C	Hexachlorobutadiene	40.000	39.406	1.5	101	0.00
35	Caprolactam	40.000	40.381	-1.0	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.203	-0.5	99	0.00
37	2-Methylnaphthalene	40.000	39.987	0.0	102	0.00
38	1-Methylnaphthalene	40.000	39.688	0.8	101	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.824	2.9	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.711	5.7	99	0.00
42 S	2,4,6-Tribromophenol	80.000	79.677	0.4	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.390	1.5	101	0.00
44	2,4,5-Trichlorophenol	40.000	40.302	-0.8	102	0.00
45 S	2-Fluorobiphenyl	80.000	76.390	4.5	101	0.00
46	1,1'-Biphenyl	40.000	38.769	3.1	102	0.00
47	2-Chloronaphthalene	40.000	39.197	2.0	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112124\
 Data File : BF140536.D
 Acq On : 21 Nov 2024 15:07
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF112124

Quant Time: Nov 21 16:11:44 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.037	-0.1	101	0.00
49	Acenaphthylene	40.000	38.850	2.9	102	0.00
50	Dimethylphthalate	40.000	39.591	1.0	103	0.00
51	2,6-Dinitrotoluene	40.000	39.356	1.6	100	0.00
52 C	Acenaphthene	40.000	38.578	3.6	101	0.00
53	3-Nitroaniline	40.000	40.901	-2.3	102	0.00
54 P	2,4-Dinitrophenol	40.000	36.341	9.1	89	0.00
55	Dibenzofuran	40.000	38.644	3.4	101	0.00
56 P	4-Nitrophenol	40.000	42.263	-5.7	104	0.00
57	2,4-Dinitrotoluene	40.000	41.025	-2.6	103	0.00
58	Fluorene	40.000	39.115	2.2	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.797	-4.5	105	0.00
60	Diethylphthalate	40.000	39.078	2.3	102	0.00
61	4-Chlorophenyl-phenylether	40.000	39.536	1.2	103	0.00
62	4-Nitroaniline	40.000	40.369	-0.9	101	0.00
63	Azobenzene	40.000	39.396	1.5	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.134	-2.8	98	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.723	3.2	102	0.00
67	4-Bromophenyl-phenylether	40.000	38.977	2.6	103	0.00
68	Hexachlorobenzene	40.000	38.974	2.6	102	0.00
69	Atrazine	40.000	34.643	13.4	106	0.00
70 C	Pentachlorophenol	40.000	42.779	-6.9	99	0.00
71	Phenanthrene	40.000	38.861	2.8	101	0.00
72	Anthracene	40.000	38.829	2.9	101	0.00
73	Carbazole	40.000	39.207	2.0	102	0.00
74	Di-n-butylphthalate	40.000	39.765	0.6	104	0.00
75 C	Fluoranthene	40.000	40.002	-0.0	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
77	Benzydine	40.000	39.351	1.6	111	0.00
78	Pyrene	40.000	38.219	4.5	106	0.00
79 S	Terphenyl-d14	80.000	75.460	5.7	105	0.00
80	Butylbenzylphthalate	40.000	40.542	-1.4	111	0.00
81	Benzo(a)anthracene	40.000	39.723	0.7	114	0.00
82	3,3'-Dichlorobenzidine	40.000	41.011	-2.5	110	0.00
83	Chrysene	40.000	39.406	1.5	111	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.823	-4.6	118	0.00
85 c	Di-n-octyl phthalate	40.000	43.965	-9.9	124	0.01
86 I	Perylene-d12	20.000	20.000	0.0	102	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	38.995	2.5	99	0.01
88	Benzo(b)fluoranthene	40.000	40.850	-2.1	110	0.00
89	Benzo(k)fluoranthene	40.000	40.110	-0.3	102	0.01
90 C	Benzo(a)pyrene	40.000	40.041	-0.1	103	0.01
91	Dibenzo(a,h)anthracene	40.000	40.052	-0.1	102	0.01
92	Benzo(g,h,i)perylene	40.000	39.591	1.0	101	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112124\
Data File : BF140536.D
Acq On : 21 Nov 2024 15:07
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF112124

Quant Time: Nov 21 16:11:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0