

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722
Instrument ID: BNA_E Calibration Date(s): 11/06/2024 11/06/2024
Calibration Time(s): 13:51 18:02

LAB FILE ID:		RRF2.5 = BE101493.D			RRF005 = BE101494.D		RRF010 = BE101495.D	
		RRF020 = BE101496.D			RRF040 = BE101497.D		RRF050 = BE101498.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.161	1.135	1.120	1.101	1.114	1.131	2.0
Benzaldehyde		0.902	0.930	0.875	0.767	0.654	0.759	19.8
Phenol-d6		1.494	1.525	1.558	1.540	1.571	1.551	2.5
Phenol		1.646	1.682	1.702	1.664	1.709	1.688	2.5
bis(2-Chloroethyl) ether		1.362	1.474	1.392	1.165	1.446	1.397	8.0
2-Chlorophenol		1.339	1.343	1.359	1.316	1.336	1.340	1.3
2-Methylphenol		1.015	1.041	1.073	1.113	1.123	1.093	4.8
2,2-oxybis(1-Chloropropane)		1.733	1.690	1.700	1.628	1.625	1.654	3.4
Acetophenone		0.462	0.460	0.461	0.443	0.447	0.453	1.9
3+4-Methylphenols		1.376	1.483	1.529	1.524	1.558	1.515	4.7
n-Nitroso-di-n-propylamine	0.958	0.996	1.040	1.058	1.027	1.046	1.023	3.5
Nitrobenzene-d5		0.313	0.316	0.318	0.312	0.318	0.317	1.5
Hexachloroethane		0.524	0.509	0.507	0.484	0.494	0.500	3.0
Nitrobenzene		0.325	0.329	0.330	0.324	0.329	0.329	1.4
Isophorone		0.599	0.619	0.632	0.611	0.618	0.616	2.0
2-Nitrophenol		0.159	0.167	0.175	0.174	0.182	0.175	5.4
2,4-Dimethylphenol		0.196	0.199	0.204	0.198	0.205	0.203	2.8
bis(2-Chloroethoxy) methane		0.376	0.383	0.386	0.371	0.375	0.377	1.7
2,4-Dichlorophenol		0.275	0.279	0.289	0.283	0.294	0.288	3.4
Naphthalene		1.071	1.033	1.027	0.983	0.983	1.010	3.6
4-Chloroaniline		0.342	0.364	0.371	0.358	0.356	0.355	3.7
Hexachlorobutadiene		0.209	0.198	0.202	0.192	0.195	0.198	3.0
Caprolactam		0.097	0.107	0.107	0.107	0.106	0.106	4.1
4-Chloro-3-methylphenol		0.293	0.318	0.316	0.312	0.318	0.314	3.3
2-Methylnaphthalene		0.744	0.733	0.733	0.706	0.705	0.717	2.9
Hexachlorocyclopentadiene		0.106	0.128	0.153	0.167	0.175	0.154	17.7
2,4,6-Trichlorophenol		0.358	0.352	0.360	0.358	0.361	0.362	2.2
2-Fluorobiphenyl		1.353	1.295	1.299	1.204	1.160	1.225	7.5
2,4,5-Trichlorophenol		0.390	0.381	0.402	0.400	0.411	0.403	3.7
1,1-Biphenyl		1.451	1.416	1.424	1.336	1.344	1.380	3.7
2-Chloronaphthalene		1.124	1.104	1.115	1.060	1.065	1.088	2.5
2-Nitroaniline		0.248	0.273	0.292	0.288	0.300	0.288	7.4
Dimethylphthalate		1.500	1.479	1.478	1.389	1.368	1.424	4.2
Acenaphthylene		1.664	1.660	1.661	1.584	1.586	1.622	2.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722
Instrument ID: BNA_E Calibration Date(s): 11/06/2024 11/06/2024
Calibration Time(s): 13:51 18:02

LAB FILE ID:		RRF2.5 = BE101493.D			RRF005 = BE101494.D		RRF010 = BE101495.D	
		RRF020 = BE101496.D			RRF040 = BE101497.D		RRF050 = BE101498.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.316	0.330	0.336	0.324	0.327	0.329	2.2
3-Nitroaniline		0.282	0.320	0.335	0.328	0.323	0.319	5.7
Acenaphthene		1.116	1.075	1.080	1.007	0.996	1.035	5.4
2,4-Dinitrophenol			0.139	0.178	0.194	0.208	0.193	15.8
4-Nitrophenol		0.197	0.235	0.278	0.275	0.274	0.264	13.3
Dibenzofuran		1.799	1.736	1.724	1.615	1.591	1.666	5.2
2,4-Dinitrotoluene		0.426	0.456	0.480	0.460	0.463	0.462	4.0
Diethylphthalate		1.600	1.579	1.577	1.457	1.442	1.502	5.4
4-Chlorophenyl-phenylether		0.757	0.736	0.733	0.689	0.680	0.706	5.1
Fluorene		1.490	1.475	1.458	1.363	1.328	1.391	6.0
4-Nitroaniline		0.285	0.334	0.359	0.349	0.355	0.344	8.1
4,6-Dinitro-2-methylphenol		0.084	0.104	0.117	0.123	0.131	0.119	16.0
n-Nitrosodiphenylamine		0.550	0.534	0.545	0.519	0.526	0.531	2.7
2,4,6-Tribromophenol		0.388	0.391	0.391	0.369	0.364	0.376	3.6
4-Bromophenyl-phenylether		0.226	0.218	0.221	0.214	0.221	0.221	1.9
Hexachlorobenzene		0.301	0.290	0.295	0.282	0.287	0.290	2.3
Atrazine		0.202	0.187	0.151	0.171	0.121	0.157	19.7
Pentachlorophenol		0.129	0.139	0.155	0.162	0.167	0.158	11.4
Phenanthrene		1.067	1.000	1.016	0.942	0.936	0.973	5.9
Anthracene		1.030	0.985	0.998	0.943	0.932	0.961	4.9
Carbazole		1.030	1.003	1.001	0.940	0.924	0.963	5.0
Di-n-butylphthalate		1.312	1.269	1.261	1.170	1.128	1.190	7.7
Fluoranthene		1.442	1.353	1.316	1.208	1.157	1.247	10.0
Pyrene		1.201	1.158	1.199	1.130	1.125	1.138	4.9
Terphenyl-d14		1.065	1.023	1.028	0.902	0.819	0.904	15.4
Butylbenzylphthalate		0.541	0.522	0.531	0.507	0.500	0.511	4.3
3,3-Dichlorobenzidine		0.466	0.474	0.480	0.469	0.478	0.468	2.9
Benzo (a) anthracene		1.334	1.258	1.258	1.170	1.138	1.188	8.5
Chrysene		1.261	1.215	1.207	1.116	1.074	1.129	9.1
Bis (2-ethylhexyl) phthalate		0.857	0.816	0.822	0.760	0.738	0.773	7.8
Di-n-octyl phthalate		1.510	1.430	1.383	1.284	1.226	1.308	10.6
Benzo (b) fluoranthene		1.220	1.158	1.107	1.056	1.053	1.097	6.5
Benzo (k) fluoranthene		1.119	1.045	1.080	0.977	0.966	0.996	9.2
Benzo (a) pyrene		1.023	0.974	0.977	0.926	0.923	0.947	5.0
Indeno (1,2,3-cd) pyrene		1.480	1.394	1.411	1.347	1.342	1.374	4.4

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722
Instrument ID: BNA_E Calibration Date(s): 11/06/2024 11/06/2024
Calibration Time(s): 13:51 18:02

LAB FILE ID:		RRF2.5 = BE101493.D		RRF005 = BE101494.D		RRF010 = BE101495.D		
		RRF020 = BE101496.D		RRF040 = BE101497.D		RRF050 = BE101498.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.250	1.169	1.189	1.128	1.111	1.145	5.7
Benzo (g,h,i) perylene		1.239	1.164	1.182	1.133	1.148	1.163	3.5
1,2,4,5-Tetrachlorobenzene		0.546	0.518	0.535	0.513	0.519	0.527	2.2
1,4-Dioxane		0.495	0.458	0.438	0.406	0.378	0.421	10.5
2,3,4,6-Tetrachlorophenol		0.379	0.366	0.377	0.369	0.370	0.373	1.5