

SOP ID : M1312-SPLP-10
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-4
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : WC
Tumbler ID : T-2
TCLP Filter ID : 114771

Start Prep Date : 11/08/2024 Time : 14:00
End Prep Date : 11/09/2024 Time : 07:15
Combination Ratio : 20
ZHE Cleaning Batch : N/A
Initial Room Temperature: 23 °C
Final Room Temperature: 22 °C
TCLP Technician Signature : 
Supervisor By : 

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
SPLP FLUID	N/A	WP110482
N/A	N/A	N/A
HNO3-TCLP,1N	N/A	WP108585
pH Strips	N/A	W1931,W1934,W2350,W2755
pH Strips	N/A	W1937,W1938,W1939,W1940,W1941,W1942
1 Liter Amber	N/A	23091
120ml Plastic bottle	N/A	21029
1:1 HNO3	MP81119	N/A

Extraction Conformance/Non-Conformance Comments:

TUMBLER T-2 checked,30 rpm. Matrix spikes are added after fofostration and before preservation. p4722-15 is used for MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/09/24 11:30	 / CLP room	 / Met sig
	Preparation Group	Analysis Group

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
P4722-05	WC-1(0-6)	11	100.03	2000	N/A	N/A	N/A	7.2	1.5	T-2
P4722-10	WC-2(0-6)	12	100.04	2000	N/A	N/A	N/A	7.0	1.0	T-2
P4722-15	WC-3(0-6)	13	100.02	2000	N/A	N/A	N/A	7.2	1.5	T-2
PB164792TB	LEB792	14	N/A	2000	N/A	N/A	N/A	4.24	1.0	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
P4722-05	WC-1(0-6)	N/A	N/A	N/A	N/A	100	N/A
P4722-10	WC-2(0-6)	N/A	N/A	N/A	N/A	100	N/A
P4722-15	WC-3(0-6)	N/A	N/A	N/A	N/A	100	N/A
PB164792TB	LEB792	N/A	N/A	N/A	N/A	N/A	N/A

Hot Block ID : WC S-1 /WC S-2

Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
P4722-05	WC-1(0-6)	N/A	N/A	N/A	N/A	#1	4.24
P4722-10	WC-2(0-6)	N/A	N/A	N/A	N/A	#1	4.24
P4722-15	WC-3(0-6)	N/A	N/A	N/A	N/A	#1	4.24
PB164792TB	LEB792	N/A	N/A	N/A	N/A	#1	4.24