

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/12/2024

Lab File ID: BE101575.D Time Analyzed: 10:18

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	74030	7.553	315608	10.33	204281	14.17
UPPER LIMIT	148060	8.053	631216	10.826	408562	14.669
LOWER LIMIT	37015	7.053	157804	9.826	102141	13.669
EPA SAMPLE NO.						
01 PB164886BL	43722	7.56	176416	10.32	107077	14.17
02 WC-1(0-6)	37247	7.55	175455	10.32	135773	14.16
03 WC-3(0-6)MS	40085	7.55	193126	10.32	136804	14.16
04 WC-3(0-6)MSD	43626	7.55	201868	10.32	141265	14.17

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.