

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: WALS01

Lab Code: CHEM

SAS No.: P4722

SDG NO.: P4722

Lab File ID: BF140589.D

DFTPP Injection Date: 11/25/2024

Instrument ID: BNA\_F

DFTPP Injection Time: 09:07

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.8
68	Less than 2.0% of mass 69	0.7 ( 1.9 ) 1
69	Mass 69 relative abundance	36.7
70	Less than 2.0% of mass 69	0.2 ( 0.6 ) 1
127	10.0 - 80.0% of mass 198	48.2
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
275	10.0 - 60.0% of mass 198	28.6
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	98.2
443	15.0 - 24.0% of mass 442	18.5 ( 18.9 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140590.D	11/25/2024	09:33
PB164886TB	PB164886TB	BF140599.D	11/25/2024	13:27