

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722
Instrument ID: BNA_F Calibration Date(s): 11/05/2024 11/05/2024
Calibration Time(s): 12:26 15:30

LAB FILE ID:		RRF2.5 = BF140231.D			RRF005 = BF140232.D		RRF010 = BF140233.D	
		RRF020 = BF140234.D			RRF040 = BF140235.D		RRF050 = BF140236.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.359	1.298	1.261	1.159	1.138	1.207	8.3
Benzaldehyde		1.159	1.137	1.085	0.949	0.839	0.956	18.5
Phenol-d6		1.765	1.679	1.604	1.495	1.462	1.552	8.7
Phenol		1.863	1.779	1.752	1.593	1.575	1.649	9.2
bis(2-Chloroethyl) ether		1.417	1.354	1.285	1.209	1.201	1.252	8.6
2-Chlorophenol		1.441	1.385	1.311	1.214	1.187	1.263	9.4
2-Methylphenol		1.155	1.105	1.086	1.028	1.005	1.053	6.1
2,2-oxybis(1-Chloropropane)		2.305	2.213	2.143	1.958	1.910	2.027	9.7
Acetophenone		0.564	0.535	0.512	0.474	0.461	0.496	8.4
3+4-Methylphenols		1.503	1.446	1.403	1.277	1.240	1.323	9.7
n-Nitroso-di-n-propylamine	1.035	1.117	1.062	1.015	0.949	0.923	0.990	8.0
Nitrobenzene-d5		0.435	0.420	0.415	0.392	0.387	0.402	5.2
Hexachloroethane		0.612	0.603	0.576	0.533	0.524	0.556	7.4
Nitrobenzene		0.452	0.442	0.436	0.414	0.407	0.422	5.0
Isophorone		0.754	0.722	0.699	0.675	0.659	0.693	5.1
2-Nitrophenol		0.160	0.167	0.170	0.171	0.167	0.168	2.5
2,4-Dimethylphenol		0.240	0.237	0.234	0.224	0.221	0.228	3.9
bis(2-Chloroethoxy) methane		0.459	0.441	0.428	0.400	0.390	0.415	6.9
2,4-Dichlorophenol		0.297	0.299	0.291	0.278	0.270	0.283	4.7
Naphthalene		1.206	1.119	1.071	0.992	0.956	1.033	9.9
4-Chloroaniline		0.372	0.367	0.354	0.330	0.322	0.340	7.1
Hexachlorobutadiene		0.251	0.241	0.238	0.222	0.218	0.229	6.2
Caprolactam		0.090	0.087	0.088	0.085	0.084	0.086	2.9
4-Chloro-3-methylphenol		0.337	0.326	0.319	0.312	0.303	0.315	4.3
2-Methylnaphthalene		0.772	0.743	0.702	0.644	0.626	0.674	9.6
Hexachlorocyclopentadiene		0.124	0.145	0.169	0.183	0.177	0.164	13.3
2,4,6-Trichlorophenol		0.377	0.369	0.363	0.361	0.354	0.363	2.2
2-Fluorobiphenyl		1.481	1.381	1.312	1.183	1.148	1.253	11.3
2,4,5-Trichlorophenol		0.391	0.398	0.396	0.373	0.374	0.381	4.0
1,1-Biphenyl		1.663	1.590	1.521	1.376	1.342	1.446	10.0
2-Chloronaphthalene		1.195	1.171	1.135	1.050	1.036	1.088	7.2
2-Nitroaniline		0.357	0.372	0.371	0.369	0.363	0.364	2.0
Dimethylphthalate		1.363	1.310	1.262	1.191	1.182	1.232	6.7
Acenaphthylene		1.730	1.668	1.618	1.486	1.465	1.540	8.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722
Instrument ID: BNA_F Calibration Date(s): 11/05/2024 11/05/2024
Calibration Time(s): 12:26 15:30

LAB FILE ID:		RRF2.5 = BF140231.D			RRF005 = BF140232.D		RRF010 = BF140233.D	
		RRF020 = BF140234.D			RRF040 = BF140235.D		RRF050 = BF140236.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.302	0.297	0.299	0.286	0.286	0.290	3.6
3-Nitroaniline		0.294	0.286	0.282	0.269	0.267	0.273	5.8
Acenaphthene		1.204	1.153	1.134	1.059	1.055	1.093	6.6
2,4-Dinitrophenol			0.071	0.101	0.128	0.135	0.120	23.2
4-Nitrophenol		0.160	0.178	0.196	0.203	0.206	0.192	8.8
Dibenzofuran		1.773	1.683	1.603	1.467	1.424	1.528	10.6
2,4-Dinitrotoluene		0.384	0.385	0.388	0.374	0.373	0.376	3.2
Diethylphthalate		1.360	1.319	1.271	1.185	1.170	1.224	7.7
4-Chlorophenyl-phenylether		0.715	0.673	0.638	0.581	0.573	0.613	10.3
Fluorene		1.446	1.345	1.267	1.154	1.127	1.216	11.6
4-Nitroaniline		0.278	0.276	0.282	0.270	0.267	0.271	3.5
4,6-Dinitro-2-methylphenol			0.080	0.100	0.110	0.112	0.105	12.6
n-Nitrosodiphenylamine		0.646	0.608	0.600	0.551	0.543	0.575	7.5
2,4,6-Tribromophenol		0.199	0.200	0.197	0.194	0.197	0.198	1.4
4-Bromophenyl-phenylether		0.230	0.219	0.223	0.209	0.208	0.215	4.1
Hexachlorobenzene		0.261	0.255	0.251	0.237	0.235	0.245	4.3
Atrazine		0.204	0.189	0.155	0.156	0.127	0.157	18.5
Pentachlorophenol		0.105	0.121	0.133	0.138	0.139	0.132	10.6
Phenanthrene		1.082	1.044	1.004	0.915	0.885	0.954	9.3
Anthracene		1.071	1.014	0.984	0.894	0.881	0.936	9.3
Carbazole		0.973	0.919	0.901	0.822	0.805	0.855	8.9
Di-n-butylphthalate		1.119	1.096	1.062	0.975	0.961	1.012	7.9
Fluoranthene		1.157	1.101	1.044	0.950	0.928	0.998	10.5
Pyrene		1.851	1.733	1.728	1.590	1.572	1.651	7.5
Terphenyl-d14		1.383	1.309	1.258	1.159	1.164	1.221	8.2
Butylbenzylphthalate		0.580	0.596	0.613	0.597	0.601	0.595	2.9
3,3-Dichlorobenzidine		0.387	0.409	0.410	0.417	0.406	0.403	2.6
Benzo (a) anthracene		1.386	1.344	1.345	1.232	1.232	1.284	5.7
Chrysene		1.320	1.264	1.206	1.154	1.177	1.202	5.8
Bis (2-ethylhexyl) phthalate		0.868	0.858	0.858	0.819	0.825	0.833	3.8
Di-n-octyl phthalate		1.201	1.234	1.283	1.267	1.265	1.254	2.5
Benzo (b) fluoranthene		1.300	1.184	1.214	1.245	1.125	1.210	6.0
Benzo (k) fluoranthene		1.214	1.230	1.206	1.027	1.085	1.110	9.5
Benzo (a) pyrene		1.023	0.999	1.022	0.993	0.975	0.997	2.2
Indeno (1,2,3-cd) pyrene		1.107	1.110	1.213	1.195	1.172	1.174	4.1

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722
Instrument ID: BNA_F Calibration Date(s): 11/05/2024 11/05/2024
Calibration Time(s): 12:26 15:30

LAB FILE ID:		RRF2.5 = BF140231.D		RRF005 = BF140232.D		RRF010 = BF140233.D		
		RRF020 = BF140234.D		RRF040 = BF140235.D		RRF050 = BF140236.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		0.917	0.923	0.998	0.979	0.968	0.966	3.4
Benzo (g,h,i) perylene		0.949	0.941	1.017	0.994	0.977	0.985	3.1
1,2,4,5-Tetrachlorobenzene		0.657	0.619	0.600	0.562	0.559	0.586	7.2
1,4-Dioxane		0.613	0.582	0.576	0.550	0.526	0.555	6.6
2,3,4,6-Tetrachlorophenol		0.323	0.324	0.321	0.305	0.309	0.313	3.3

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1