

d. Acid Fraction

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

SequenceID: bf110724 NA NO YES  1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)  2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)  3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series  4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis of 600 series and 12 hours for 8000 series  5. GC/MS Calibration Met:  a. Initial calibration Meet Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable range.  b. Continuous Calibration(CCC) Meet Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable range.  CCC is marginally biased high for compound #77 which is not present in parameter list of the associated samples.	CHEMTECH PROJECT NU	MBER:	bf110724	_			
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)  3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series  4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series  5. GC/MS Calibration Met:  a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.  b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.  CCC is marginally biased high for compound #77 which is not present in parameter list of the associated samples.	SequenceID:	bf110724			NA	NO	YES
NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)  3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series  4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series  5. GC/MS Calibration Met:  a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.  b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.  CCC is marginally biased high for compound #77 which is not present in parameter list of the associated samples.	1. Chromatograms Labeled/O	Compounds Identifi	ed. (Field samples and Met	hod Blanks)			_
series  4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series  5. GC/MS Calibration Met:  a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.  b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.  CCC is marginally biased high for compound #77 which is not present in parameter list of the associated samples.	· .			CLP AND NJ)			_
analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series  5. GC/MS Calibration Met:  a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.  b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.  CCC is marginally biased high for compound #77 which is not present in parameter list of the associated samples.  6. Blank Contamination - If yes, list compounds and concentrations in each blank:		- Performed every	24 hours for 600 series and	12 hours for 8000			<b>✓</b>
a. Initial calibration Meet Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable range.  b. Continuous Calibration(CCC) Meet Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable range.  CCC is marginally biased high for compound #77 which is not present in parameter list of the associated samples.  6. Blank Contamination - If yes, list compounds and concentrations in each blank:	analysis and continuing ca	llibration performed	-	•			_
If not met, list those compounds and their recoveries which fall outside the acceptable range.  b. Continuous Calibration(CCC) Meet Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable range.  CCC is marginally biased high for compound #77 which is not present in parameter list of the associated samples.  6. Blank Contamination - If yes, list compounds and concentrations in each blank:	5. GC/MS Calibration Met:						<b>✓</b>
If not met, list those compounds and their recoveries which fall outside the acceptable range.  CCC is marginally biased high for compound #77 which is not present in parameter list of the associated samples.  6. Blank Contamination - If yes, list compounds and concentrations in each blank:			eries which fall outside the a	acceptable range.			_
6. Blank Contamination - If yes, list compounds and concentrations in each blank:		` '		acceptable range.		<u> </u>	
<del></del>	CCC is marginally biased	l high for compound	d #77 which is not present in	n parameter list of the associated san	nples.		
	-	yes, list compounds	and concentrations in each	blank:		<u> </u>	

7. Surrogate Recoveries Meet Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<del></del>
a. B/N Fraction	
d. Acid Fraction	
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u> </u>
a. B/N Fraction  Recovery failed for some compound in P4693-05MS/MSD and P4718-02MS/MSD due to matrix interference hence no contact action is required.	orrective
d. Acid Fraction	
9. Internal Standard Area/Retention Time Shift Meet Criteria  Comments:	<u> </u>
Internal standard failed in 05, P4695-01 DL these samples will be re-analyzed again. Internal standard failed even in diluted sample of P4722-13 and P4700-01 proving matrix interference. Hence no corrective required. internal standard failed in the samples P4694-01, P4720-01 due to matrix interference hence this analyses is final however these samples will be analyzed 10. Extraction Holding Time Met with their required dilutions.  If not met, list number of days exceeded for each sample:	
11. Analysis Holding Time Met  If not met, list number of days exceeded for each sample:  ———————————————————————————————————	<u> </u>
ADDITIONAL COMMENTS:  Recovery failed for some compound in PB164702BS but their Hit is not present in any associated samples. The data will be use for hardcopies. The samples P4722-13, P4734-01, and P4700-01 had to be analyzed with 5X dilution due to dirty and viscous matrix. Hence no corrective action is required.	ed

krunal	11/08/2024
Analyst	Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER:		bf110724			
SequenceID :	bf110724		J	NA NO	O YES