

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/08/2024

Lab File ID: BE101544.D Time Analyzed: 10:07

Instrument ID: BNA\_E GC Column: ZB-GR ID: 0.25 (mm)

|                | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #   | IS3 (ANT)<br>AREA # | RT #   |
|----------------|---------------------|-------|---------------------|--------|---------------------|--------|
| 12 HOUR STD    | 51174               | 7.558 | 237509              | 10.33  | 179651              | 14.17  |
| UPPER LIMIT    | 102348              | 8.058 | 475018              | 10.826 | 359302              | 14.674 |
| LOWER LIMIT    | 25587               | 7.058 | 118755              | 9.826  | 89825.5             | 13.674 |
| EPA SAMPLE NO. |                     |       |                     |        |                     |        |
| 01 PB164750BL  | 34636               | 7.56  | 145179              | 10.33  | 93474               | 14.17  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.