

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/07/2024

Lab File ID: BF140261.D Time Analyzed: 08:59

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	752975	11.41	445081	14.069	403199	15.574
UPPER LIMIT	1505950	11.91	890162	14.569	806398	16.074
LOWER LIMIT	376488	10.91	222541	13.569	201600	15.074
EPA SAMPLE NO.						
01 WC-3 (0-6)	319073 *	11.40	257317	14.05	207144	15.54

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.