

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110524\
 Data File : BF140239.D
 Acq On : 05 Nov 2024 16:07
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF110524

Quant Time: Nov 05 16:49:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 05 16:47:56 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	39.464	1.3	99	0.00
3	Pyridine	40.000	39.452	1.4	100	0.00
4	n-Nitrosodimethylamine	40.000	39.257	1.9	98	0.00
5 S	2-Fluorophenol	80.000	78.362	2.0	102	0.00
6	Aniline	40.000	40.014	-0.0	103	0.00
7 S	Phenol-d6	80.000	77.432	3.2	100	0.00
8	2-Chlorophenol	40.000	38.692	3.3	101	0.00
9	Benzaldehyde	40.000	39.783	0.5	100	0.00
10 C	Phenol	40.000	38.933	2.7	101	0.00
11	bis(2-Chloroethyl)ether	40.000	39.355	1.6	102	0.00
12	1,3-Dichlorobenzene	40.000	38.882	2.8	102	0.00
13 C	1,4-Dichlorobenzene	40.000	38.398	4.0	101	0.00
14	1,2-Dichlorobenzene	40.000	38.196	4.5	102	0.00
15	Benzyl Alcohol	40.000	39.247	1.9	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.401	4.0	99	0.00
17	2-Methylphenol	40.000	39.246	1.9	100	0.00
18	Hexachloroethane	40.000	39.497	1.3	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.494	3.8	100	0.00
20	3+4-Methylphenols	40.000	39.115	2.2	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	37.721	5.7	101	0.00
23 S	Nitrobenzene-d5	80.000	76.816	4.0	101	0.00
24	Nitrobenzene	40.000	37.948	5.1	99	0.00
25	Isophorone	40.000	38.273	4.3	101	0.00
26 C	2-Nitrophenol	40.000	40.349	-0.9	102	0.00
27	2,4-Dimethylphenol	40.000	39.198	2.0	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.011	5.0	101	0.00
29 C	2,4-Dichlorophenol	40.000	39.176	2.1	102	0.00
30	1,2,4-Trichlorobenzene	40.000	38.282	4.3	102	0.00
31	Naphthalene	40.000	37.544	6.1	100	0.00
32	Benzoic acid	40.000	43.437	-8.6	108	0.00
33	4-Chloroaniline	40.000	37.948	5.1	100	0.00
34 C	Hexachlorobutadiene	40.000	38.514	3.7	102	0.00
35	Caprolactam	40.000	39.367	1.6	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.305	1.7	102	0.00
37	2-Methylnaphthalene	40.000	37.768	5.6	101	0.00
38	1-Methylnaphthalene	40.000	37.949	5.1	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.952	2.6	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.739	-4.3	96	0.00
42 S	2,4,6-Tribromophenol	80.000	81.047	-1.3	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.578	1.1	102	0.00
44	2,4,5-Trichlorophenol	40.000	40.052	-0.1	105	0.00
45 S	2-Fluorobiphenyl	80.000	74.802	6.5	102	0.00
46	1,1'-Biphenyl	40.000	37.852	5.4	102	0.00
47	2-Chloronaphthalene	40.000	37.970	5.1	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110524\
 Data File : BF140239.D
 Acq On : 05 Nov 2024 16:07
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF110524

Quant Time: Nov 05 16:49:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 05 16:47:56 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.981	0.0	101	0.00
49	Acenaphthylene	40.000	38.713	3.2	103	0.00
50	Dimethylphthalate	40.000	38.809	3.0	103	0.00
51	2,6-Dinitrotoluene	40.000	40.169	-0.4	104	0.00
52 C	Acenaphthene	40.000	38.663	3.3	102	0.00
53	3-Nitroaniline	40.000	40.203	-0.5	105	0.00
54 P	2,4-Dinitrophenol	40.000	39.421	1.4	102	0.00
55	Dibenzofuran	40.000	38.079	4.8	102	0.00
56 P	4-Nitrophenol	40.000	43.155	-7.9	104	0.00
57	2,4-Dinitrotoluene	40.000	39.922	0.2	103	0.00
58	Fluorene	40.000	37.745	5.6	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.370	-0.9	106	0.00
60	Diethylphthalate	40.000	39.080	2.3	103	0.00
61	4-Chlorophenyl-phenylether	40.000	38.133	4.7	103	0.00
62	4-Nitroaniline	40.000	40.392	-1.0	104	0.00
63	Azobenzene	40.000	37.917	5.2	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.784	-4.5	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.296	4.3	104	0.00
67	4-Bromophenyl-phenylether	40.000	38.966	2.6	104	0.00
68	Hexachlorobenzene	40.000	39.175	2.1	105	0.00
69	Atrazine	40.000	39.237	1.9	102	0.00
70 C	Pentachlorophenol	40.000	43.075	-7.7	106	0.00
71	Phenanthrene	40.000	38.048	4.9	103	0.00
72	Anthracene	40.000	37.919	5.2	103	0.00
73	Carbazole	40.000	38.509	3.7	104	0.00
74	Di-n-butylphthalate	40.000	39.432	1.4	106	0.00
75 C	Fluoranthene	40.000	38.360	4.1	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzidine	40.000	62.287	-55.7#	127	0.00
78	Pyrene	40.000	39.081	2.3	104	0.00
79 S	Terphenyl-d14	80.000	77.169	3.5	104	0.00
80	Butylbenzylphthalate	40.000	41.595	-4.0	106	0.00
81	Benzo(a)anthracene	40.000	40.097	-0.2	107	0.00
82	3,3'-Dichlorobenzidine	40.000	42.445	-6.1	105	0.01
83	Chrysene	40.000	38.388	4.0	102	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	40.534	-1.3	105	0.01
85 c	Di-n-octyl phthalate	40.000	41.861	-4.7	106	0.02
86 I	Perylene-d12	20.000	20.000	0.0	100	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	40.852	-2.1	101	0.02
88	Benzo(b)fluoranthene	40.000	39.022	2.4	95	0.02
89	Benzo(k)fluoranthene	40.000	39.994	0.0	109	0.02
90 C	Benzo(a)pyrene	40.000	40.159	-0.4	101	0.02
91	Dibenzo(a,h)anthracene	40.000	40.753	-1.9	101	0.02
92	Benzo(g,h,i)perylene	40.000	40.205	-0.5	100	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110524\
Data File : BF140239.D
Acq On : 05 Nov 2024 16:07
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF110524

Quant Time: Nov 05 16:49:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 05 16:47:56 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0