

LAB CHRONICLE

OrderID:	P4722	OrderDate:	11/5/2024 3:33:08 PM
Client:	Walsh Construction Company II, LLC	Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2
Contact:	Kayla Timony	Location:	L23,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4722-04	WC-1(0-6)	TCLP	TCLP BNA	8270E	11/05/24	11/07/24	11/08/24	11/05/24
P4722-09	WC-2(0-6)	TCLP	TCLP BNA	8270E	11/05/24	11/07/24	11/08/24	11/05/24
P4722-14	WC-3(0-6)	TCLP	TCLP BNA	8270E	11/05/24	11/07/24	11/08/24	11/05/24



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: P4722
Client: Walsh Construction Company II, LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



SAMPLE DATA

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	11/05/24	
Client Sample ID:	WC-1(0-6)		SDG No.:	P4722	
Lab Sample ID:	P4722-04		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101562.D	1	11/07/24 11:30	11/08/24 21:06	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	132		10 - 139	88%	SPK: 150
13127-88-3	Phenol-d6	107		10 - 134	71%	SPK: 150
4165-60-0	Nitrobenzene-d5	96.3		49 - 133	96%	SPK: 100
321-60-8	2-Fluorobiphenyl	94.4		52 - 132	94%	SPK: 100
118-79-6	2,4,6-Tribromophenol	155		44 - 137	103%	SPK: 150
1718-51-0	Terphenyl-d14	111		48 - 125	111%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	43600		7.559		
1146-65-2	Naphthalene-d8	181000		10.326		
15067-26-2	Acenaphthene-d10	121000		14.163		
1517-22-2	Phenanthrene-d10	294000		16.901		
1719-03-5	Chrysene-d12	363000		21.067		
1520-96-3	Perylene-d12	463000		23.346		

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	11/05/24	
Client Sample ID:	WC-1(0-6)		SDG No.:	P4722	
Lab Sample ID:	P4722-04		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101562.D	1	11/07/24 11:30	11/08/24 21:06	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/05/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/05/24
Client Sample ID:	WC-2(0-6)	SDG No.:	P4722
Lab Sample ID:	P4722-09	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101561.D	1	11/07/24 11:30	11/08/24 20:30	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	155		10 - 139	103%	SPK: 150
13127-88-3	Phenol-d6	129		10 - 134	86%	SPK: 150
4165-60-0	Nitrobenzene-d5	106		49 - 133	106%	SPK: 100
321-60-8	2-Fluorobiphenyl	103		52 - 132	103%	SPK: 100
118-79-6	2,4,6-Tribromophenol	162		44 - 137	108%	SPK: 150
1718-51-0	Terphenyl-d14	121		48 - 125	121%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	40500	7.559			
1146-65-2	Naphthalene-d8	171000	10.326			
15067-26-2	Acenaphthene-d10	113000	14.168			
1517-22-2	Phenanthrene-d10	271000	16.906			
1719-03-5	Chrysene-d12	325000	21.066			
1520-96-3	Perylene-d12	419000	23.352			

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/05/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/05/24
Client Sample ID:	WC-2(0-6)	SDG No.:	P4722
Lab Sample ID:	P4722-09	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101561.D	1	11/07/24 11:30	11/08/24 20:30	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	11/05/24	
Client Sample ID:	WC-3(0-6)		SDG No.:	P4722	
Lab Sample ID:	P4722-14		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101560.D	1	11/07/24 11:30	11/08/24 19:54	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	140		10 - 139	94%	SPK: 150
13127-88-3	Phenol-d6	123		10 - 134	82%	SPK: 150
4165-60-0	Nitrobenzene-d5	98.7		49 - 133	99%	SPK: 100
321-60-8	2-Fluorobiphenyl	97.4		52 - 132	97%	SPK: 100
118-79-6	2,4,6-Tribromophenol	159		44 - 137	106%	SPK: 150
1718-51-0	Terphenyl-d14	114		48 - 125	114%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	43400		7.559		
1146-65-2	Naphthalene-d8	177000		10.326		
15067-26-2	Acenaphthene-d10	120000		14.169		
1517-22-2	Phenanthrene-d10	291000		16.907		
1719-03-5	Chrysene-d12	362000		21.066		
1520-96-3	Perylene-d12	470000		23.346		

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/05/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/05/24
Client Sample ID:	WC-3(0-6)	SDG No.:	P4722
Lab Sample ID:	P4722-14	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101560.D	1	11/07/24 11:30	11/08/24 19:54	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/07/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/07/24
Client Sample ID:	PB164694TB	SDG No.:	P4722
Lab Sample ID:	PB164694TB	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101547.D	1	11/07/24 11:30	11/08/24 12:03	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	177		10 - 139	118%	SPK: 150
13127-88-3	Phenol-d6	166		10 - 134	111%	SPK: 150
4165-60-0	Nitrobenzene-d5	110		49 - 133	110%	SPK: 100
321-60-8	2-Fluorobiphenyl	110		52 - 132	110%	SPK: 100
118-79-6	2,4,6-Tribromophenol	157		44 - 137	104%	SPK: 150
1718-51-0	Terphenyl-d14	112		48 - 125	112%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	34900	7.555			
1146-65-2	Naphthalene-d8	150000	10.329			
15067-26-2	Acenaphthene-d10	99900	14.171			
1517-22-2	Phenanthrene-d10	241000	16.909			
1719-03-5	Chrysene-d12	325000	21.069			
1520-96-3	Perylene-d12	415000	23.349			

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/07/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/07/24
Client Sample ID:	PB164694TB	SDG No.:	P4722
Lab Sample ID:	PB164694TB	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101547.D	1	11/07/24 11:30	11/08/24 12:03	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4722-04	WC-1(0-6)	2-Fluorophenol	150	132	88		10	139
		Phenol-d6	150	107	71		10	134
		Nitrobenzene-d5	100	96.3	96		49	133
		2-Fluorobiphenyl	100	94.4	94		52	132
		2,4,6-Tribromophenol	150	155	103		44	137
		Terphenyl-d14	100	111	111		48	125
P4722-09	WC-2(0-6)	2-Fluorophenol	150	155	103		10	139
		Phenol-d6	150	129	86		10	134
		Nitrobenzene-d5	100	106	106		49	133
		2-Fluorobiphenyl	100	103	103		52	132
		2,4,6-Tribromophenol	150	162	108		44	137
		Terphenyl-d14	100	121	121		48	125
P4722-14	WC-3(0-6)	2-Fluorophenol	150	140	94		10	139
		Phenol-d6	150	123	82		10	134
		Nitrobenzene-d5	100	98.7	99		49	133
		2-Fluorobiphenyl	100	97.4	97		52	132
		2,4,6-Tribromophenol	150	159	106		44	137
		Terphenyl-d14	100	114	114		48	125
P4739-04MS	TP-14MS	2-Fluorophenol	150	175	117		10	139
		Phenol-d6	150	155	103		10	134
		Nitrobenzene-d5	100	112	112		49	133
		2-Fluorobiphenyl	100	110	110		52	132
		2,4,6-Tribromophenol	150	166	110		44	137
		Terphenyl-d14	100	115	115		48	125
P4739-04MSD	TP-14MSD	2-Fluorophenol	150	175	117		10	139
		Phenol-d6	150	156	104		10	134
		Nitrobenzene-d5	100	114	114		49	133
		2-Fluorobiphenyl	100	109	109		52	132
		2,4,6-Tribromophenol	150	168	112		44	137
		Terphenyl-d14	100	115	115		48	125
PB164694TB	PB164694TB	2-Fluorophenol	150	177	118		10	139
		Phenol-d6	150	166	111		10	134
		Nitrobenzene-d5	100	110	110		49	133
		2-Fluorobiphenyl	100	110	110		52	132
		2,4,6-Tribromophenol	150	157	104		44	137
		Terphenyl-d14	100	112	112		48	125
PB164765BL	PB164765BL	2-Fluorophenol	150	148	99		10	139
		Phenol-d6	150	146	97		10	134
		Nitrobenzene-d5	100	101	101		49	133
		2-Fluorobiphenyl	100	94.3	94		52	132
		2,4,6-Tribromophenol	150	184	122		44	137
		Terphenyl-d14	100	96.4	96		48	125
PB164765BS	PB164765BS	2-Fluorophenol	150	175	116		10	139
		Phenol-d6	150	164	109		10	134
		Nitrobenzene-d5	100	103	103		49	133
		2-Fluorobiphenyl	100	103	103		52	132
		2,4,6-Tribromophenol	150	149	99		44	137
		Terphenyl-d14	100	104	104		48	125

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4739-04MS	Client Sample ID:	TP-14MS					DataFile:	BE101557.D		
Pyridine	500	0	340	ug/L	68				10	109	
1,4-Dichlorobenzene	500	0	520	ug/L	104				55	125	
2-Methylphenol	500	0	560	ug/L	112				37	126	
3+4-Methylphenols	500	0	540	ug/L	108				31	127	
Hexachloroethane	500	0	520	ug/L	104				49	110	
Nitrobenzene	500	0	540	ug/L	108				62	112	
Hexachlorobutadiene	500	0	510	ug/L	102				52	125	
2,4,6-Trichlorophenol	500	0	590	ug/L	118	*			78	112	
2,4,5-Trichlorophenol	500	0	580	ug/L	116	*			71	111	
2,4-Dinitrotoluene	500	0	550	ug/L	110				50	142	
Hexachlorobenzene	500	0	550	ug/L	110				72	115	
Pentachlorophenol	1000	0	1300	ug/L	130				25	139	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4739-04MSD	Client Sample ID:	TP-14MSD					DataFile:	BE101558.D		
Pyridine	500	0	330	ug/L	66		3		10	109	20
1,4-Dichlorobenzene	500	0	510	ug/L	102		2		55	125	20
2-Methylphenol	500	0	550	ug/L	110		2		37	126	20
3+4-Methylphenols	500	0	550	ug/L	110		2		31	127	20
Hexachloroethane	500	0	510	ug/L	102		2		49	110	20
Nitrobenzene	500	0	550	ug/L	110		2		62	112	20
Hexachlorobutadiene	500	0	510	ug/L	102		0		52	125	20
2,4,6-Trichlorophenol	500	0	590	ug/L	118	*	0		78	112	20
2,4,5-Trichlorophenol	500	0	580	ug/L	116	*	0		71	111	20
2,4-Dinitrotoluene	500	0	560	ug/L	112		2		50	142	20
Hexachlorobenzene	500	0	540	ug/L	108		2		72	115	20
Pentachlorophenol	1000	0	1200	ug/L	120		8		25	139	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: 8270E DataFile: BE101546.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB164765BS	Pyridine	50	38.5	ug/L	77				29	97	
	1,4-Dichlorobenzene	50	48.0	ug/L	96				76	103	
	2-Methylphenol	50	53.8	ug/L	108				69	109	
	3+4-Methylphenols	50	52.7	ug/L	105				67	106	
	Hexachloroethane	50	47.3	ug/L	95				76	118	
	Nitrobenzene	50	49.6	ug/L	99				58	106	
	Hexachlorobutadiene	50	44.9	ug/L	90				69	101	
	2,4,6-Trichlorophenol	50	52.2	ug/L	104				61	110	
	2,4,5-Trichlorophenol	50	51.0	ug/L	102				70	106	
	2,4-Dinitrotoluene	50	50.2	ug/L	100				60	115	
	Hexachlorobenzene	50	48.3	ug/L	97				73	106	
	Pentachlorophenol	100	110	ug/L	110				47	114	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB164765BL

Lab Name: CHEMTECH

Contract: WALS01

Lab Code: CHEM Case No.: P4722

SAS No.: P4722 SDG NO.: P4722

Lab File ID: BF140287.D

Lab Sample ID: PB164765BL

Instrument ID: BNA_F

Date Extracted: 11/07/2024

Matrix: (soil/water) water

Date Analyzed: 11/08/2024

Level: (low/med) LOW

Time Analyzed: 10:02

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
WC-1 (0-6)	P4722-04	BE101562.D	11/08/2024
WC-2 (0-6)	P4722-09	BE101561.D	11/08/2024
PB164694TB	PB164694TB	BE101547.D	11/08/2024
TP-14MS	P4739-04MS	BE101557.D	11/08/2024
TP-14MSD	P4739-04MSD	BE101558.D	11/08/2024
WC-3 (0-6)	P4722-14	BE101560.D	11/08/2024
PB164765BS	PB164765BS	BE101546.D	11/08/2024

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: WALS01

Lab Code: CHEM

SAS No.: P4722 SDG NO.: P4722

Lab File ID: BE101491.D

DFTPP Injection Date: 11/06/2024

Instrument ID: BNA_E

DFTPP Injection Time: 11:40

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	14.8
68	Less than 2.0% of mass 69	0.2 (1.3) 1
69	Mass 69 relative abundance	16
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	23.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	3.6
275	10.0 - 60.0% of mass 198	18.3
365	Greater than 1% of mass 198	2.9
441	Present, but less than mass 443	16.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BE101493.D	11/06/2024	13:51
SSTDICC005	SSTDICC005	BE101494.D	11/06/2024	14:27
SSTDICC010	SSTDICC010	BE101495.D	11/06/2024	15:03
SSTDICC020	SSTDICC020	BE101496.D	11/06/2024	15:39
SSTDICCC040	SSTDICCC040	BE101497.D	11/06/2024	16:14
SSTDICC050	SSTDICC050	BE101498.D	11/06/2024	16:50
SSTDICC060	SSTDICC060	BE101499.D	11/06/2024	17:26
SSTDICC080	SSTDICC080	BE101500.D	11/06/2024	18:02

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: WALS01

Lab Code: CHEM

SAS No.: P4722 SDG NO.: P4722

Lab File ID: BE101543.D

DFTPP Injection Date: 11/08/2024

Instrument ID: BNA_E

DFTPP Injection Time: 09:34

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	16.2
68	Less than 2.0% of mass 69	0.2 (1) 1
69	Mass 69 relative abundance	17.2
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	24.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	3.8
275	10.0 - 60.0% of mass 198	19.1
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	16.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BE101544.D	11/08/2024	10:07
PB164765BS	PB164765BS	BE101546.D	11/08/2024	11:27
PB164694TB	PB164694TB	BE101547.D	11/08/2024	12:03
TP-14MS	P4739-04MS	BE101557.D	11/08/2024	18:07
TP-14MSD	P4739-04MSD	BE101558.D	11/08/2024	18:43
WC-3(0-6)	P4722-14	BE101560.D	11/08/2024	19:54
WC-2(0-6)	P4722-09	BE101561.D	11/08/2024	20:30
WC-1(0-6)	P4722-04	BE101562.D	11/08/2024	21:06

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: WALS01

Lab Code: CHEM

SAS No.: P4722 SDG NO.: P4722

Lab File ID: BF140230.D

DFTPP Injection Date: 11/05/2024

Instrument ID: BNA_F

DFTPP Injection Time: 11:56

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	38.4
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	36.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	45.5
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	27.8
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF140231.D	11/05/2024	12:26
SSTDICC005	SSTDICC005	BF140232.D	11/05/2024	12:52
SSTDICC010	SSTDICC010	BF140233.D	11/05/2024	13:18
SSTDICC020	SSTDICC020	BF140234.D	11/05/2024	13:45
SSTDICCC040	SSTDICCC040	BF140235.D	11/05/2024	14:11
SSTDICC050	SSTDICC050	BF140236.D	11/05/2024	14:37
SSTDICC060	SSTDICC060	BF140237.D	11/05/2024	15:03
SSTDICC080	SSTDICC080	BF140238.D	11/05/2024	15:30

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: WALS01

Lab Code: CHEM

SAS No.: P4722 SDG NO.: P4722

Lab File ID: BF140285.D

DFTPP Injection Date: 11/08/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:09

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	28.4
68	Less than 2.0% of mass 69	0.5 (2) 1
69	Mass 69 relative abundance	27.7
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	36.3
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.8
275	10.0 - 60.0% of mass 198	25.2
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140286.D	11/08/2024	09:35
PB164765BL	PB164765BL	BF140287.D	11/08/2024	10:02

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/08/2024

Lab File ID: BE101544.D Time Analyzed: 10:07

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	51174	7.558	237509	10.33	179651	14.17
UPPER LIMIT	102348	8.058	475018	10.826	359302	14.674
LOWER LIMIT	25587	7.058	118755	9.826	89825.5	13.674
EPA SAMPLE NO.						
01 PB164765BS	39669	7.56	182397	10.33	121151	14.17
02 PB164694TB	34932	7.56	149600	10.33	99908	14.17
03 WC-1 (0-6)	43597	7.56	181077	10.33	121098	14.16
04 WC-2 (0-6)	40529	7.56	171441	10.33	113413	14.17
05 WC-3 (0-6)	43377	7.56	177204	10.33	120124	14.17
06 TP-14MS	47543	7.56	202399	10.33	130758	14.17
07 TP-14MSD	43513	7.56	182419	10.33	120731	14.17

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/08/2024

Lab File ID: BE101544.D Time Analyzed: 10:07

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	420948	16.912	472751	21.072	602462	23.357
UPPER LIMIT	841896	17.412	945502	21.572	1204920	23.857
LOWER LIMIT	210474	16.412	236376	20.572	301231	22.857
EPA SAMPLE NO.						
01 PB164765BS	268998	16.91	348750	21.07	477959	23.35
02 PB164694TB	241198	16.91	324971	21.07	414995	23.35
03 WC-1 (0-6)	293866	16.90	362629	21.07	463143	23.35
04 WC-2 (0-6)	271101	16.91	324956	21.07	419473	23.35
05 WC-3 (0-6)	290802	16.91	361520	21.07	469692	23.35
06 TP-14MS	298419	16.91	369646	21.07	476783	23.35
07 TP-14MSD	285084	16.91	361953	21.07	460616	23.35

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/08/2024

Lab File ID: BF140286.D Time Analyzed: 09:35

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	203786	6.881	760595	8.16	473134	9.92
UPPER LIMIT	407572	7.381	1521190	8.663	946268	10.422
LOWER LIMIT	101893	6.381	380298	7.663	236567	9.422
EPA SAMPLE NO.						
01 PB164765BL	180593	6.88	699300	8.16	441390	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/08/2024

Lab File ID: BF140286.D Time Analyzed: 09:35

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	883136	11.41	554697	14.063	449723	15.551
UPPER LIMIT	1766270	11.91	1109390	14.563	899446	16.051
LOWER LIMIT	441568	10.91	277349	13.563	224862	15.051
EPA SAMPLE NO.						
01 PB164765BL	869591	11.40	647570	14.05	498484	15.55

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	PB164765BL		SDG No.:	P4722
Lab Sample ID:	PB164765BL		Matrix:	TCLP
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140287.D	1	11/07/24 11:30	11/08/24 10:02	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.60	U	1.60	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.84	U	0.84	5.00	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.20	U	1.20	10.0	ug/L
67-72-1	Hexachloroethane	1.00	U	1.00	5.00	ug/L
98-95-3	Nitrobenzene	1.30	U	1.30	5.00	ug/L
87-68-3	Hexachlorobutadiene	1.30	U	1.30	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.89	U	0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	1.00	U	1.00	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.50	U	1.50	5.00	ug/L
118-74-1	Hexachlorobenzene	1.10	U	1.10	5.00	ug/L
87-86-5	Pentachlorophenol	1.90	U	1.90	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	148		10 - 139	99%	SPK: 150
13127-88-3	Phenol-d6	146		10 - 134	97%	SPK: 150
4165-60-0	Nitrobenzene-d5	101		49 - 133	101%	SPK: 100
321-60-8	2-Fluorobiphenyl	94.3		52 - 132	94%	SPK: 100
118-79-6	2,4,6-Tribromophenol	184		44 - 137	122%	SPK: 150
1718-51-0	Terphenyl-d14	96.4		48 - 125	96%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	181000	6.875			
1146-65-2	Naphthalene-d8	699000	8.157			
15067-26-2	Acenaphthene-d10	441000	9.916			
1517-22-2	Phenanthrene-d10	870000	11.404			
1719-03-5	Chrysene-d12	648000	14.051			
1520-96-3	Perylene-d12	498000	15.545			

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	PB164765BL		SDG No.:	P4722
Lab Sample ID:	PB164765BL		Matrix:	TCLP
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140287.D	1	11/07/24 11:30	11/08/24 10:02	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	PB164765BS		SDG No.:	P4722
Lab Sample ID:	PB164765BS		Matrix:	TCLP
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101546.D	1	11/07/24 11:30	11/08/24 11:27	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	38.5		1.60	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	48.0		0.84	5.00	ug/L
95-48-7	2-Methylphenol	53.8		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	52.7		1.20	10.0	ug/L
67-72-1	Hexachloroethane	47.3		1.00	5.00	ug/L
98-95-3	Nitrobenzene	49.6		1.30	5.00	ug/L
87-68-3	Hexachlorobutadiene	44.9		1.30	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	52.2		0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	51.0		1.00	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	50.2		1.50	5.00	ug/L
118-74-1	Hexachlorobenzene	48.3		1.10	5.00	ug/L
87-86-5	Pentachlorophenol	110	E	1.90	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	175		10 - 139	116%	SPK: 150
13127-88-3	Phenol-d6	164		10 - 134	109%	SPK: 150
4165-60-0	Nitrobenzene-d5	103		49 - 133	103%	SPK: 100
321-60-8	2-Fluorobiphenyl	103		52 - 132	103%	SPK: 100
118-79-6	2,4,6-Tribromophenol	149		44 - 137	99%	SPK: 150
1718-51-0	Terphenyl-d14	104		48 - 125	104%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	39700		7.559		
1146-65-2	Naphthalene-d8	182000		10.326		
15067-26-2	Acenaphthene-d10	121000		14.169		
1517-22-2	Phenanthrene-d10	269000		16.907		
1719-03-5	Chrysene-d12	349000		21.072		
1520-96-3	Perylene-d12	478000		23.352		

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	PB164765BS		SDG No.:	P4722
Lab Sample ID:	PB164765BS		Matrix:	TCLP
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101546.D	1	11/07/24 11:30	11/08/24 11:27	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/06/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/06/24
Client Sample ID:	TP-14MS	SDG No.:	P4722
Lab Sample ID:	P4739-04MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101557.D	1	11/07/24 11:30	11/08/24 18:07	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	340		15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	520		8.40	50.0	ug/L
95-48-7	2-Methylphenol	560		11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	540		11.5	100	ug/L
67-72-1	Hexachloroethane	520		10.1	50.0	ug/L
98-95-3	Nitrobenzene	540		12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	510		12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	590		8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	580		10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	550		15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	550		11.4	50.0	ug/L
87-86-5	Pentachlorophenol	1300	E	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	175		10 - 139	117%	SPK: 150
13127-88-3	Phenol-d6	155		10 - 134	103%	SPK: 150
4165-60-0	Nitrobenzene-d5	112		49 - 133	112%	SPK: 100
321-60-8	2-Fluorobiphenyl	110		52 - 132	110%	SPK: 100
118-79-6	2,4,6-Tribromophenol	166		44 - 137	110%	SPK: 150
1718-51-0	Terphenyl-d14	115		48 - 125	115%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	47500		7.558		
1146-65-2	Naphthalene-d8	202000		10.326		
15067-26-2	Acenaphthene-d10	131000		14.168		
1517-22-2	Phenanthrene-d10	298000		16.906		
1719-03-5	Chrysene-d12	370000		21.072		
1520-96-3	Perylene-d12	477000		23.351		

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/06/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/06/24
Client Sample ID:	TP-14MS	SDG No.:	P4722
Lab Sample ID:	P4739-04MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101557.D	1	11/07/24 11:30	11/08/24 18:07	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/06/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/06/24
Client Sample ID:	TP-14MSD	SDG No.:	P4722
Lab Sample ID:	P4739-04MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101558.D	1	11/07/24 11:30	11/08/24 18:43	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	330		15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	510		8.40	50.0	ug/L
95-48-7	2-Methylphenol	550		11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	550		11.5	100	ug/L
67-72-1	Hexachloroethane	510		10.1	50.0	ug/L
98-95-3	Nitrobenzene	550		12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	510		12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	590		8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	580		10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	560		15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	540		11.4	50.0	ug/L
87-86-5	Pentachlorophenol	1200	E	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	175		10 - 139	117%	SPK: 150
13127-88-3	Phenol-d6	156		10 - 134	104%	SPK: 150
4165-60-0	Nitrobenzene-d5	114		49 - 133	114%	SPK: 100
321-60-8	2-Fluorobiphenyl	109		52 - 132	109%	SPK: 100
118-79-6	2,4,6-Tribromophenol	168		44 - 137	112%	SPK: 150
1718-51-0	Terphenyl-d14	115		48 - 125	115%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	43500		7.558		
1146-65-2	Naphthalene-d8	182000		10.326		
15067-26-2	Acenaphthene-d10	121000		14.168		
1517-22-2	Phenanthrene-d10	285000		16.906		
1719-03-5	Chrysene-d12	362000		21.072		
1520-96-3	Perylene-d12	461000		23.352		

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/06/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/06/24
Client Sample ID:	TP-14MSD	SDG No.:	P4722
Lab Sample ID:	P4739-04MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101558.D	1	11/07/24 11:30	11/08/24 18:43	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_E\Methods\
 Method File : 8270-BE110624.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Nov 07 00:01:20 2024
 Response Via : Initial Calibration

Calibration Files

2.5 =BE101493.D 5 =BE101494.D 10 =BE101495.D 20 =BE101496.D 40 =BE101497.D 50 =BE101498.D 60 =BE101499.D 80 =BE101500.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.495	0.458	0.438	0.406	0.378	0.394	0.378	0.421	10.54	
3)	Pyridine	1.154	1.183	1.143	1.154	1.192	1.230	1.186	1.177	2.53	
4)	n-Nitrosodimet...	0.509	0.465	0.446	0.445	0.461	0.473	0.464	0.466	4.58	
5) S	2-Fluorophenol	1.161	1.135	1.120	1.101	1.114	1.159	1.127	1.131	1.99	
6)	Aniline	1.323	1.251	1.316	1.368	1.026	1.044	0.819	1.164	17.52	
7) S	Phenol-d6	1.494	1.525	1.558	1.540	1.571	1.617	1.551	1.551	2.49	
8)	2-Chlorophenol	1.339	1.343	1.359	1.316	1.336	1.364	1.324	1.340	1.30	
9)	Benzaldehyde	0.902	0.930	0.875	0.767	0.654	0.640	0.542	0.759	19.80	
10) C	Phenol	1.646	1.682	1.702	1.664	1.709	1.764	1.649	1.688	2.45	
11)	bis(2-Chloroet...	1.362	1.474	1.392	1.165	1.446	1.447	1.497	1.397	8.03	
12)	1,3-Dichlorobe...	1.573	1.526	1.491	1.424	1.410	1.437	1.379	1.463	4.74	
13) C	1,4-Dichlorobe...	1.565	1.543	1.508	1.444	1.437	1.464	1.412	1.482	3.89	
14)	1,2-Dichlorobe...	1.559	1.496	1.481	1.418	1.416	1.435	1.378	1.455	4.20	
15)	Benzyl Alcohol	0.791	0.861	0.910	0.919	0.958	0.977	0.926	0.906	6.94	
16)	2,2'-oxybis(1-...	1.733	1.690	1.700	1.628	1.625	1.636	1.567	1.654	3.41	
17)	2-Methylphenol	1.015	1.041	1.073	1.113	1.123	1.163	1.123	1.093	4.76	
18)	Hexachloroethane	0.524	0.509	0.507	0.484	0.494	0.501	0.483	0.500	2.95	
19) P	n-Nitroso-di-n...	0.958	0.996	1.040	1.058	1.027	1.046	1.057	0.998	1.023	3.48
20)	3+4-Methylphenols	1.376	1.483	1.529	1.524	1.558	1.606	1.530	1.515	4.74	
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.462	0.460	0.461	0.443	0.447	0.455	0.443	0.453	1.91	
23) S	Nitrobenzene-d5	0.313	0.316	0.318	0.312	0.318	0.325	0.313	0.317	1.46	
24)	Nitrobenzene	0.325	0.329	0.330	0.324	0.329	0.338	0.330	0.329	1.41	
25)	Isophorone	0.599	0.619	0.632	0.611	0.618	0.630	0.605	0.616	1.98	
26) C	2-Nitrophenol	0.159	0.167	0.175	0.174	0.182	0.187	0.183	0.175	5.41	
27)	2,4-Dimethylph...	0.196	0.199	0.204	0.198	0.205	0.212	0.206	0.203	2.83	
28)	bis(2-Chloroet...	0.376	0.383	0.386	0.371	0.375	0.381	0.369	0.377	1.65	
29) C	2,4-Dichloroph...	0.275	0.279	0.289	0.283	0.294	0.302	0.294	0.288	3.35	
30)	1,2,4-Trichlor...	0.344	0.326	0.323	0.311	0.313	0.321	0.312	0.321	3.63	
31)	Naphthalene	1.071	1.033	1.027	0.983	0.983	1.003	0.968	1.010	3.58	
32)	Benzoic acid		0.112	0.137	0.166	0.191	0.202	0.202	0.168	22.18	
33)	4-Chloroaniline	0.342	0.364	0.371	0.358	0.356	0.361	0.334	0.355	3.65	
34) C	Hexachlorobuta...	0.209	0.198	0.202	0.192	0.195	0.198	0.193	0.198	2.95	
35)	Caprolactam	0.097	0.107	0.107	0.107	0.106	0.111	0.105	0.106	4.09	
36) C	4-Chloro-3-met...	0.293	0.318	0.316	0.312	0.318	0.327	0.317	0.314	3.27	
37)	2-Methylnaphth...	0.744	0.733	0.733	0.706	0.705	0.715	0.684	0.717	2.93	
38)	1-Methylnaphth...	0.752	0.738	0.734	0.696	0.697	0.711	0.679	0.715	3.72	

Method Path : Z:\svoasrv\HPCHEM1\BNA_E\Methods\
 Method File : 8270-BE110624.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.546	0.518	0.535	0.513	0.519	0.535	0.525	0.527	2.24	
41) P	Hexachlorocycl...	0.106	0.128	0.153	0.167	0.175	0.178	0.171	0.154	17.72	
42) S	2,4,6-Tribromo...	0.388	0.391	0.391	0.369	0.364	0.375	0.357	0.376	3.64	
43) C	2,4,6-Trichlor...	0.358	0.352	0.360	0.358	0.361	0.376	0.369	0.362	2.21	
44)	2,4,5-Trichlor...	0.390	0.381	0.402	0.400	0.411	0.425	0.415	0.403	3.71	
45) S	2-Fluorobiphenyl	1.353	1.295	1.299	1.204	1.160	1.163	1.100	1.225	7.52	
46)	1,1'-Biphenyl	1.451	1.416	1.424	1.336	1.344	1.366	1.320	1.380	3.66	
47)	2-Chloronaphth...	1.124	1.104	1.115	1.060	1.065	1.091	1.060	1.088	2.49	
48)	2-Nitroaniline	0.248	0.273	0.292	0.288	0.300	0.312	0.302	0.288	7.44	
49)	Acenaphthylene	1.664	1.660	1.661	1.584	1.586	1.629	1.565	1.622	2.62	
50)	Dimethylphthalate	1.500	1.479	1.478	1.389	1.368	1.403	1.352	1.424	4.22	
51)	2,6-Dinitrotol...	0.316	0.330	0.336	0.324	0.327	0.338	0.329	0.329	2.23	
52) C	Acenaphthene	1.116	1.075	1.080	1.007	0.996	1.007	0.962	1.035	5.38	
53)	3-Nitroaniline	0.282	0.320	0.335	0.328	0.323	0.333	0.314	0.319	5.67	
54) P	2,4-Dinitrophenol		0.139	0.178	0.194	0.208	0.220	0.217	0.193	15.82	
55)	Dibenzofuran	1.799	1.736	1.724	1.615	1.591	1.629	1.568	1.666	5.20	
56) P	4-Nitrophenol	0.197	0.235	0.278	0.275	0.274	0.295	0.290	0.264	13.26	
57)	2,4-Dinitrotol...	0.426	0.456	0.480	0.460	0.463	0.482	0.466	0.462	3.99	
58)	Fluorene	1.490	1.475	1.458	1.363	1.328	1.347	1.272	1.391	6.03	
59)	2,3,4,6-Tetrac...	0.379	0.366	0.377	0.369	0.370	0.381	0.372	0.373	1.53	
60)	Diethylphthalate	1.600	1.579	1.577	1.457	1.442	1.460	1.397	1.502	5.41	
61)	4-Chlorophenyl...	0.757	0.736	0.733	0.689	0.680	0.692	0.657	0.706	5.12	
62)	4-Nitroaniline	0.285	0.334	0.359	0.349	0.355	0.366	0.359	0.344	8.07	
63)	Azobenzene	1.283	1.269	1.289	1.196	1.183	1.203	1.153	1.225	4.44	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...	0.084	0.104	0.117	0.123	0.131	0.136	0.135	0.119	15.97	
66) c	n-Nitrosodiphe...	0.550	0.534	0.545	0.519	0.526	0.533	0.509	0.531	2.68	
67)	4-Bromophenyl-...	0.226	0.218	0.221	0.214	0.221	0.225	0.219	0.221	1.85	
68)	Hexachlorobenzene	0.301	0.290	0.295	0.282	0.287	0.294	0.283	0.290	2.33	
69)	Atrazine	0.202	0.187	0.151	0.171	0.121	0.135	0.130	0.157	19.70	
70) C	Pentachlorophenol	0.129	0.139	0.155	0.162	0.167	0.176	0.176	0.158	11.39	
71)	Phenanthrene	1.067	1.000	1.016	0.942	0.936	0.947	0.901	0.973	5.87	
72)	Anthracene	1.030	0.985	0.998	0.943	0.932	0.949	0.887	0.961	4.94	
73)	Carbazole	1.030	1.003	1.001	0.940	0.924	0.947	0.899	0.963	5.00	
74)	Di-n-butylphth...	1.312	1.269	1.261	1.170	1.128	1.121	1.068	1.190	7.68	
75) C	Fluoranthene	1.442	1.353	1.316	1.208	1.157	1.158	1.096	1.247	10.04	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.324	0.433	0.250	0.403	0.656	0.503	0.440	0.430	30.17	
78)	Pyrene	1.201	1.158	1.199	1.130	1.125	1.117	1.036	1.138	4.95	
79) S	Terphenyl-d14	1.065	1.023	1.028	0.902	0.819	0.780	0.707	0.904	15.44	
80)	Butylbenzylpht...	0.541	0.522	0.531	0.507	0.500	0.505	0.474	0.511	4.31	
81)	Benzo(a)anthra...	1.334	1.258	1.258	1.170	1.138	1.118	1.040	1.188	8.47	
82)	3,3'-Dichlorob...	0.466	0.474	0.480	0.469	0.478	0.467	0.439	0.468	2.92	
83)	Chrysene	1.261	1.215	1.207	1.116	1.074	1.050	0.979	1.129	9.06	
84)	Bis(2-ethylhex...	0.857	0.816	0.822	0.760	0.738	0.733	0.686	0.773	7.81	
85) c	Di-n-octyl pht...	1.510	1.430	1.383	1.284	1.226	1.203	1.119	1.308	10.60	

Method Path : Z:\svoasrv\HPCHEM1\BNA_E\Methods\
Method File : 8270-BE110624.M

86) I	Perylene-d12	-----ISTD-----									
87)	Indeno(1,2,3-c...	1.480	1.394	1.411	1.347	1.342	1.357	1.290	1.374	4.40	
88)	Benzo(b)fluora...	1.220	1.158	1.107	1.056	1.053	1.073	1.014	1.097	6.46	
89)	Benzo(k)fluora...	1.119	1.045	1.080	0.977	0.966	0.937	0.849	0.996	9.23	
90) C	Benzo(a)pyrene	1.023	0.974	0.977	0.926	0.923	0.931	0.877	0.947	5.02	
91)	Dibenzo(a,h)an...	1.250	1.169	1.189	1.128	1.111	1.119	1.047	1.145	5.68	
92)	Benzo(g,h,i)pe...	1.239	1.164	1.182	1.133	1.148	1.162	1.113	1.163	3.46	

(#) = Out of Range

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF110524.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 05 16:47:56 2024
 Response Via : Initial Calibration

Calibration Files

2.5 =BF140231.D 5 =BF140232.D 10 =BF140233.D 20 =BF140234.D 40 =BF140235.D 50 =BF140236.D 60 =BF140237.D 80 =BF140238.D

Compound		2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.613	0.582	0.576	0.550	0.526	0.527	0.511	0.555	6.64	
3)	Pyridine	1.464	1.445	1.402	1.340	1.296	1.298	1.233	1.354	6.33	
4)	n-Nitrosodimet...	0.799	0.792	0.802	0.780	0.767	0.786	0.746	0.782	2.52	
5) S	2-Fluorophenol	1.359	1.298	1.261	1.159	1.138	1.151	1.083	1.207	8.29	
6)	Aniline	1.609	1.513	1.499	1.365	1.316	1.298	1.218	1.402	9.99	
7) S	Phenol-d6	1.765	1.679	1.604	1.495	1.462	1.468	1.392	1.552	8.66	
8)	2-Chlorophenol	1.441	1.385	1.311	1.214	1.187	1.182	1.121	1.263	9.38	
9)	Benzaldehyde	1.159	1.137	1.085	0.949	0.839	0.823	0.701	0.956	18.46	
10) C	Phenol	1.863	1.779	1.752	1.593	1.575	1.544	1.434	1.649	9.23	
11)	bis(2-Chloroet...	1.417	1.354	1.285	1.209	1.201	1.201	1.100	1.252	8.57	
12)	1,3-Dichlorobe...	1.690	1.605	1.555	1.413	1.387	1.388	1.322	1.480	9.22	
13) C	1,4-Dichlorobe...	1.720	1.636	1.551	1.423	1.393	1.401	1.322	1.492	9.78	
14)	1,2-Dichlorobe...	1.622	1.553	1.494	1.309	1.291	1.286	1.201	1.394	11.49	
15)	Benzyl Alcohol	1.282	1.280	1.258	1.182	1.147	1.154	1.075	1.197	6.59	
16)	2,2'-oxybis(1-...	2.305	2.213	2.143	1.958	1.910	1.898	1.764	2.027	9.65	
17)	2-Methylphenol	1.155	1.105	1.086	1.028	1.005	1.022	0.973	1.053	6.05	
18)	Hexachloroethane	0.612	0.603	0.576	0.533	0.524	0.537	0.507	0.556	7.37	
19) P	n-Nitroso-di-n...	1.035	1.117	1.062	1.015	0.949	0.923	0.936	0.886	0.990	8.01
20)	3+4-Methylphenols	1.503	1.446	1.403	1.277	1.240	1.242	1.150	1.323	9.73	
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.564	0.535	0.512	0.474	0.461	0.472	0.453	0.496	8.43	
23) S	Nitrobenzene-d5	0.435	0.420	0.415	0.392	0.387	0.391	0.377	0.402	5.21	
24)	Nitrobenzene	0.452	0.442	0.436	0.414	0.407	0.407	0.396	0.422	5.04	
25)	Isophorone	0.754	0.722	0.699	0.675	0.659	0.680	0.660	0.693	5.05	
26) C	2-Nitrophenol	0.160	0.167	0.170	0.171	0.167	0.173	0.168	0.168	2.47	
27)	2,4-Dimethylph...	0.240	0.237	0.234	0.224	0.221	0.227	0.216	0.228	3.88	
28)	bis(2-Chloroet...	0.459	0.441	0.428	0.400	0.390	0.399	0.383	0.415	6.93	
29) C	2,4-Dichloroph...	0.297	0.299	0.291	0.278	0.270	0.277	0.266	0.283	4.67	
30)	1,2,4-Trichlor...	0.373	0.362	0.349	0.328	0.318	0.327	0.314	0.339	6.72	
31)	Naphthalene	1.206	1.119	1.071	0.992	0.956	0.965	0.921	1.033	9.94	
32)	Benzoic acid		0.136	0.161	0.193	0.207	0.211	0.216	0.187	17.07	
33)	4-Chloroaniline	0.372	0.367	0.354	0.330	0.322	0.323	0.311	0.340	7.13	
34) C	Hexachlorobuta...	0.251	0.241	0.238	0.222	0.218	0.223	0.212	0.229	6.22	
35)	Caprolactam	0.090	0.087	0.088	0.085	0.084	0.086	0.082	0.086	2.94	
36) C	4-Chloro-3-met...	0.337	0.326	0.319	0.312	0.303	0.307	0.298	0.315	4.33	
37)	2-Methylnaphth...	0.772	0.743	0.702	0.644	0.626	0.627	0.605	0.674	9.64	
38)	1-Methylnaphth...	0.765	0.720	0.691	0.634	0.612	0.619	0.583	0.661	10.00	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF110524.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.657	0.619	0.600	0.562	0.559	0.562	0.538	0.586	7.15	
41) P	Hexachlorocycl...	0.124	0.145	0.169	0.183	0.177	0.181	0.169	0.164	13.27	
42) S	2,4,6-Tribromo...	0.199	0.200	0.197	0.194	0.197	0.202	0.195	0.198	1.40	
43) C	2,4,6-Trichlor...	0.377	0.369	0.363	0.361	0.354	0.356	0.359	0.363	2.16	
44)	2,4,5-Trichlor...	0.391	0.398	0.396	0.373	0.374	0.383	0.355	0.381	4.02	
45) S	2-Fluorobiphenyl	1.481	1.381	1.312	1.183	1.148	1.166	1.097	1.253	11.27	
46)	1,1'-Biphenyl	1.663	1.590	1.521	1.376	1.342	1.347	1.281	1.446	10.02	
47)	2-Chloronaphth...	1.195	1.171	1.135	1.050	1.036	1.040	0.990	1.088	7.17	
48)	2-Nitroaniline	0.357	0.372	0.371	0.369	0.363	0.364	0.352	0.364	1.99	
49)	Acenaphthylene	1.730	1.668	1.618	1.486	1.465	1.452	1.364	1.540	8.62	
50)	Dimethylphthalate	1.363	1.310	1.262	1.191	1.182	1.180	1.135	1.232	6.65	
51)	2,6-Dinitrotol...	0.302	0.297	0.299	0.286	0.286	0.284	0.272	0.290	3.57	
52) C	Acenaphthene	1.204	1.153	1.134	1.059	1.055	1.041	1.003	1.093	6.60	
53)	3-Nitroaniline	0.294	0.286	0.282	0.269	0.267	0.262	0.248	0.273	5.83	
54) P	2,4-Dinitrophenol		0.071	0.101	0.128	0.135	0.139	0.142	0.120	23.23	
55)	Dibenzofuran	1.773	1.683	1.603	1.467	1.424	1.421	1.324	1.528	10.59	
56) P	4-Nitrophenol	0.160	0.178	0.196	0.203	0.206	0.201	0.197	0.192	8.78	
57)	2,4-Dinitrotol...	0.384	0.385	0.388	0.374	0.373	0.374	0.352	0.376	3.23	
58)	Fluorene	1.446	1.345	1.267	1.154	1.127	1.115	1.059	1.216	11.55	
59)	2,3,4,6-Tetrac...	0.323	0.324	0.321	0.305	0.309	0.308	0.298	0.313	3.26	
60)	Diethylphthalate	1.360	1.319	1.271	1.185	1.170	1.165	1.097	1.224	7.72	
61)	4-Chlorophenyl...	0.715	0.673	0.638	0.581	0.573	0.569	0.542	0.613	10.33	
62)	4-Nitroaniline	0.278	0.276	0.282	0.270	0.267	0.270	0.254	0.271	3.45	
63)	Azobenzene	1.460	1.373	1.319	1.219	1.200	1.190	1.122	1.269	9.38	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.080	0.100	0.110	0.112	0.116	0.113	0.105	12.60	
66) c	n-Nitrosodiphe...	0.646	0.608	0.600	0.551	0.543	0.549	0.530	0.575	7.48	
67)	4-Bromophenyl-...	0.230	0.219	0.223	0.209	0.208	0.212	0.207	0.215	4.15	
68)	Hexachlorobenzene	0.261	0.255	0.251	0.237	0.235	0.242	0.237	0.245	4.26	
69)	Atrazine	0.204	0.189	0.155	0.156	0.127	0.138	0.133	0.157	18.48	
70) C	Pentachlorophenol	0.105	0.121	0.133	0.138	0.139	0.144	0.142	0.132	10.59	
71)	Phenanthrene	1.082	1.044	1.004	0.915	0.885	0.902	0.849	0.954	9.28	
72)	Anthracene	1.071	1.014	0.984	0.894	0.881	0.876	0.832	0.936	9.34	
73)	Carbazole	0.973	0.919	0.901	0.822	0.805	0.809	0.759	0.855	8.93	
74)	Di-n-butylphth...	1.119	1.096	1.062	0.975	0.961	0.966	0.907	1.012	7.87	
75) C	Fluoranthene	1.157	1.101	1.044	0.950	0.928	0.935	0.870	0.998	10.48	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.507	0.466	0.380	0.490	0.365	0.249	0.281	0.391	25.96	
78)	Pyrene	1.851	1.733	1.728	1.590	1.572	1.596	1.490	1.651	7.49	
79) S	Terphenyl-d14	1.383	1.309	1.258	1.159	1.164	1.180	1.093	1.221	8.25	
80)	Butylbenzylphth...	0.580	0.596	0.613	0.597	0.601	0.612	0.564	0.595	2.95	
81)	Benzo(a)anthra...	1.386	1.344	1.345	1.232	1.232	1.254	1.193	1.284	5.71	
82)	3,3'-Dichlorob...	0.387	0.409	0.410	0.417	0.406	0.401	0.392	0.403	2.57	
83)	Chrysene	1.320	1.264	1.206	1.154	1.177	1.180	1.113	1.202	5.81	
84)	Bis(2-ethylhex...	0.868	0.858	0.858	0.819	0.825	0.830	0.775	0.833	3.82	
85) c	Di-n-octyl pht...	1.201	1.234	1.283	1.267	1.265	1.291	1.240	1.254	2.52	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF110524.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.107	1.110	1.213	1.195	1.172	1.230	1.191	1.174	4.11
88)	Benzo(b)fluora...	1.300	1.184	1.214	1.245	1.125	1.286	1.115	1.210	6.05
89)	Benzo(k)fluora...	1.214	1.230	1.206	1.027	1.085	0.975	1.034	1.110	9.46
90) C	Benzo(a)pyrene	1.023	0.999	1.022	0.993	0.975	1.006	0.964	0.997	2.23
91)	Dibenzo(a,h)an...	0.917	0.923	0.998	0.979	0.968	1.000	0.976	0.966	3.44
92)	Benzo(g,h,i)pe...	0.949	0.941	1.017	0.994	0.977	1.017	1.002	0.985	3.13

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722

Instrument ID: BNA_E Calibration Date/Time: 11/08/2024 10:07

Lab File ID: BE101544.D Init. Calib. Date(s): 11/06/2024 11/06/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:51 18:02

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.177	1.299		10.4	
2-Fluorophenol	1.131	1.156		2.2	
Phenol-d6	1.551	1.606		3.5	
1,4-Dichlorobenzene	1.482	1.484		0.1	20.0
2-Methylphenol	1.093	1.113		1.8	
3+4-Methylphenols	1.515	1.581		4.4	
Nitrobenzene-d5	0.317	0.329		3.8	
Hexachloroethane	0.500	0.497		-0.6	
Nitrobenzene	0.329	0.344		4.6	
Hexachlorobutadiene	0.198	0.190		-4.0	20.0
2,4,6-Trichlorophenol	0.362	0.358		-1.1	20.0
2-Fluorobiphenyl	1.225	1.192		-2.7	
2,4,5-Trichlorophenol	0.403	0.404		0.2	
2,4-Dinitrotoluene	0.462	0.497		7.6	
2,4,6-Tribromophenol	0.376	0.380		1.1	
Hexachlorobenzene	0.290	0.292		0.7	
Pentachlorophenol	0.158	0.168		6.3	20.0
Terphenyl-d14	0.904	0.945		4.5	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722

Instrument ID: BNA_F Calibration Date/Time: 11/08/2024 09:35

Lab File ID: BF140286.D Init. Calib. Date(s): 11/05/2024 11/05/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:26 15:30

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.354	1.280		-5.5	
2-Fluorophenol	1.207	1.151		-4.6	
Phenol-d6	1.552	1.455		-6.3	
1,4-Dichlorobenzene	1.492	1.440		-3.5	20.0
2-Methylphenol	1.053	1.025		-2.7	
3+4-Methylphenols	1.323	1.267		-4.2	
Nitrobenzene-d5	0.402	0.378		-6.0	
Hexachloroethane	0.556	0.530		-4.7	
Nitrobenzene	0.422	0.398		-5.7	
Hexachlorobutadiene	0.229	0.225		-1.7	20.0
2,4,6-Trichlorophenol	0.363	0.362		-0.3	20.0
2-Fluorobiphenyl	1.253	1.125		-10.1	
2,4,5-Trichlorophenol	0.381	0.388		1.8	
2,4-Dinitrotoluene	0.376	0.395		5.1	
2,4,6-Tribromophenol	0.198	0.223		12.6	
Hexachlorobenzene	0.245	0.245		0.0	
Pentachlorophenol	0.132	0.154		16.7	20.0
Terphenyl-d14	1.221	1.161		-4.9	

All other compounds must meet a minimum RRF of 0.010.