

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4739-04MS	Client Sample ID:	TP-14MS					DataFile:	BE101557.D		
Pyridine	500	0	340	ug/L	68				10	109	
1,4-Dichlorobenzene	500	0	520	ug/L	104				55	125	
2-Methylphenol	500	0	560	ug/L	112				37	126	
3+4-Methylphenols	500	0	540	ug/L	108				31	127	
Hexachloroethane	500	0	520	ug/L	104				49	110	
Nitrobenzene	500	0	540	ug/L	108				62	112	
Hexachlorobutadiene	500	0	510	ug/L	102				52	125	
2,4,6-Trichlorophenol	500	0	590	ug/L	118	*			78	112	
2,4,5-Trichlorophenol	500	0	580	ug/L	116	*			71	111	
2,4-Dinitrotoluene	500	0	550	ug/L	110				50	142	
Hexachlorobenzene	500	0	550	ug/L	110				72	115	
Pentachlorophenol	1000	0	1300	ug/L	130				25	139	