

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4739-04MSD	Client Sample ID:	TP-14MSD					DataFile:	BE101558.D		
Pyridine	500	0	330	ug/L	66		3		10	109	20
1,4-Dichlorobenzene	500	0	510	ug/L	102		2		55	125	20
2-Methylphenol	500	0	550	ug/L	110		2		37	126	20
3+4-Methylphenols	500	0	550	ug/L	110		2		31	127	20
Hexachloroethane	500	0	510	ug/L	102		2		49	110	20
Nitrobenzene	500	0	550	ug/L	110		2		62	112	20
Hexachlorobutadiene	500	0	510	ug/L	102		0		52	125	20
2,4,6-Trichlorophenol	500	0	590	ug/L	118	*	0		78	112	20
2,4,5-Trichlorophenol	500	0	580	ug/L	116	*	0		71	111	20
2,4-Dinitrotoluene	500	0	560	ug/L	112		2		50	142	20
Hexachlorobenzene	500	0	540	ug/L	108		2		72	115	20
Pentachlorophenol	1000	0	1200	ug/L	120		8		25	139	20