

SOP ID :	M1311-TCLP-15	Start Prep Date :	11/06/2024	Time :	16:45
SDG No :	N/A	End Prep Date :	11/07/2024	Time :	09:35
Weigh By :	JP	Combination Ratio :	20		
Balance ID :	WC SC-4	ZHE Cleaning Batch :	N/A		
pH Meter ID :	WC PH METER-1	Initial Room Temperature:	24 °C		
Extraction By :	JP	Final Room Temperature:	23 °C		
Filter By :	JP	TCLP Technician Signature :	TO		
Pipette ID :	WC	Supervisor By :	SJ		
Tumbler ID :	ZHE-1 / ZHE-2				
TCLP Filter ID :	50223706				

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP108622
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
40ml VOA Vials	22437	N/A

Extraction Conformance/Non-Conformance Comments:

TUMBLER ZHE-1 / ZHE-2 checked, 30 rpm. ALL ZHE samples are extracted and given as vial A & B. Leak checked after 10 minutes of tumbling.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/07/24 11:00	TO Preparation Group	SJ VOC Lab Analysis Group

Sample ID	ClientID	ZHE Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
P4718-03	WB-307-SB02	01	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4719-03	BAYAVE-STOCKPILE	02	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4720-05	JC-701-COMP-01	03	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4722-02	WC-1(3.5)	04	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4722-07	WC-2(4)	05	25.04	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4722-12	WC-3(3)	06	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4722-16	WC-1(5.5)	07	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4722-18	WC-2(2)	08	25.04	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4722-20	WC-3(5)	09	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4732-05	PPE-Grab	10	25.01	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
P4738-02	72-12018	11	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
P4739-04	TP-14	12	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
P4739-08	BP-G2	13	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
P4739-12	BP-B2	14	25.01	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
P4739-16	TP-11	15	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
PB164695TB	LEB695	16	N/A	500	N/A	N/A	N/A	4.94	N/A	ZHE-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
P4718-03	WB-307-SB02	N/A	N/A	N/A	N/A	100	N/A
P4719-03	BAYAVE-STOCKPILE	N/A	N/A	N/A	N/A	100	N/A
P4720-05	JC-701-COMP-01	N/A	N/A	N/A	N/A	100	N/A
P4722-02	WC-1(3.5)	N/A	N/A	N/A	N/A	100	N/A
P4722-07	WC-2(4)	N/A	N/A	N/A	N/A	100	N/A
P4722-12	WC-3(3)	N/A	N/A	N/A	N/A	100	N/A
P4722-16	WC-1(5.5)	N/A	N/A	N/A	N/A	100	N/A
P4722-18	WC-2(2)	N/A	N/A	N/A	N/A	100	N/A
P4722-20	WC-3(5)	N/A	N/A	N/A	N/A	100	N/A
P4732-05	PPE-Grab	N/A	N/A	N/A	N/A	100	N/A
P4738-02	72-12018	N/A	N/A	N/A	N/A	100	N/A
P4739-04	TP-14	N/A	N/A	N/A	N/A	100	N/A
P4739-08	BP-G2	N/A	N/A	N/A	N/A	100	N/A
P4739-12	BP-B2	N/A	N/A	N/A	N/A	100	N/A
P4739-16	TP-11	N/A	N/A	N/A	N/A	100	N/A
PB164695TB	LEB695	N/A	N/A	N/A	N/A	N/A	N/A