

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN103024

Review By	Semsettin Yesilyurt	Review On	11/1/2024 2:56:11 PM
Supervise By	Maresh Dadoda	Supervise On	11/4/2024 1:18:30 AM
SubDirectory	VN103024	HP Acquire Method	MSVOA_N HP Processing Method 82n103024w.m

STD. NAME	STD REF.#
Tune/Reschk	VP131194,VP131195
Initial Calibration Stds	VP131185,VP131186,VP131187,VP131188,VP131189,VP131190
CCC	VP131191,VP131192
Internal Standard/PEM	VP128298
ICV/I.BLK	VP131193
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084569.D	30 Oct 2024 10:42		JC\MD	Ok
2	VSTDICC100	VSTDICC100	VN084570.D	30 Oct 2024 11:46	Comp #03 fail for %D	JC\MD	Ok,M
3	VSTDICCC050	VSTDICCC050	VN084571.D	30 Oct 2024 12:09	LR-03,06,16,18,43	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN084572.D	30 Oct 2024 12:33	QR-88	JC\MD	Ok,M
5	VSTDICC010	VSTDICC010	VN084573.D	30 Oct 2024 12:57		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN084574.D	30 Oct 2024 13:21		JC\MD	Ok,M
7	VSTDICC001	VSTDICC001	VN084575.D	30 Oct 2024 13:45		JC\MD	Ok,M
8	VIBLK	VIBLK	VN084576.D	30 Oct 2024 14:42		JC\MD	Ok
9	VSTDICV050	ICVVN103024	VN084577.D	30 Oct 2024 15:06		JC\MD	Ok,M
10	BFB	BFB	VN084578.D	30 Oct 2024 19:24		JC\MD	Ok
11	VSTDCCC050	VSTDCCC050	VN084579.D	30 Oct 2024 20:12		JC\MD	Ok,M
12	VN1030MBL01	VN1030MBL01	VN084580.D	30 Oct 2024 20:48		JC\MD	Ok
13	VN1030WBL01	VN1030WBL01	VN084581.D	30 Oct 2024 21:12		JC\MD	Ok
14	VN1030WBS01	VN1030WBS01	VN084582.D	30 Oct 2024 21:46		JC\MD	Ok,M
15	VN1030WBSD01	VN1030WBSD01	VN084583.D	30 Oct 2024 22:10		JC\MD	Ok,M
16	P4594-04	TP-4	VN084584.D	30 Oct 2024 22:34	pH#5.0 A	JC\MD	Ok,M
17	P4594-08	BP-F17	VN084585.D	30 Oct 2024 22:58	pH#5.0 A	JC\MD	Ok
18	P4594-12	BP-F16	VN084586.D	30 Oct 2024 23:22	pH#5.0 A	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN103024

Review By	Semsettin Yesilyurt	Review On	11/1/2024 2:56:11 PM		
Supervise By	Mahesh Dadoda	Supervise On	11/4/2024 1:18:30 AM		
SubDirectory	VN103024	HP Acquire Method	MSVOA_N	HP Processing Method	82n103024w.m
STD. NAME		STD REF.#			
Tune/Reschk		VP131194.VP131195			
Initial Calibration Stds		VP131185,VP131186,VP131187,VP131188,VP131189,VP131190			
CCC		VP131191,VP131192			
Internal Standard/PEM		VP128298			
ICV/I.BLK		VP131193			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4594-16	TP-5	VN084587.D	30 Oct 2024 23:46	pH#7.0 A	JC\MD	Ok
20	P4594-20	BP-F15	VN084588.D	31 Oct 2024 00:09	pH#6.0 A	JC\MD	Ok
21	P4597-03	RED-1-1	VN084589.D	31 Oct 2024 00:34	pH#6.0 A	JC\MD	Ok
22	P4597-06	RED-1-2	VN084590.D	31 Oct 2024 00:57	pH#5.0 A	JC\MD	Ok,M
23	P4597-09	BLUE-2-1	VN084591.D	31 Oct 2024 01:21	pH#5.0 A	JC\MD	Ok
24	P4597-12	BLUE-2-2	VN084592.D	31 Oct 2024 01:45	pH#5.0 A	JC\MD	Ok,M
25	P4598-04	BP-F12	VN084593.D	31 Oct 2024 02:09	pH#7.0 A	JC\MD	Ok
26	P4598-08	BP-F11	VN084594.D	31 Oct 2024 02:33	pH#7.0 A	JC\MD	Ok
27	P4598-12	TP-8	VN084595.D	31 Oct 2024 02:57	pH#6.0 A	JC\MD	Ok
28	P4611-03	TP-1	VN084596.D	31 Oct 2024 03:21	pH#5.0 A	JC\MD	Ok
29	P4611-06	TP-2	VN084597.D	31 Oct 2024 03:45	pH#5.0 A	JC\MD	Ok
30	P4611-09	TP-3	VN084598.D	31 Oct 2024 04:09	pH#5.0 A	JC\MD	Ok
31	P4611-12	TP-4	VN084599.D	31 Oct 2024 04:33	pH#5.0 A	JC\MD	Ok
32	P4611-15	TP-5	VN084600.D	31 Oct 2024 04:57	pH#5.0 A	JC\MD	Ok
33	P4611-18	TP-6	VN084601.D	31 Oct 2024 05:21	pH#5.0 A	JC\MD	Ok
34	P4612-04	MOO-24-00335	VN084602.D	31 Oct 2024 05:45	pH#5.0 A	JC\MD	Ok
35	P4613-02	ARS20-0001	VN084603.D	31 Oct 2024 06:09	pH#5.0 A	JC\MD	Ok
36	PB164501ZHE#13	PB164501ZHE#13	VN084604.D	31 Oct 2024 06:33	pH#5.0 A	JC\MD	Ok,M
37	PB164501ZHE#14	PB164501ZHE#14	VN084605.D	31 Oct 2024 06:57	pH#5.0 A	JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN103024

Review By	Semsettin Yesilyurt	Review On	11/1/2024 2:56:11 PM		
Supervise By	Mahesh Dadoda	Supervise On	11/4/2024 1:18:30 AM		
SubDirectory	VN103024	HP Acquire Method	MSVOA_N	HP Processing Method	82n103024w.m
STD. NAME		STD REF.#			
Tune/Reschk		VP131194.VP131195			
Initial Calibration Stds		VP131185,VP131186,VP131187,VP131188,VP131189,VP131190			
CCC		VP131191,VP131192			
Internal Standard/PEM		VP128298			
ICV/I.BLK		VP131193			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

38	PB164501ZHE#15	PB164501ZHE#15	VN084606.D	31 Oct 2024 07:22	pH#5.0 A	JC\MD	Ok,M
39	PB164501ZHE#16	PB164501ZHE#16	VN084607.D	31 Oct 2024 07:46	pH#5.0 A	JC\MD	Ok
40	PB164501ZHE#17	PB164501ZHE#17	VN084608.D	31 Oct 2024 08:10	pH#5.0 A	JC\MD	Ok,M
41	PB164501ZHE#18	PB164501ZHE#18	VN084609.D	31 Oct 2024 08:34	pH#5.0 A	JC\MD	Ok,M
42	PB164501ZHE#19	PB164501ZHE#19	VN084610.D	31 Oct 2024 08:58	pH#5.0 A	JC\MD	Ok,M
43	PB164501ZHE#20	PB164501ZHE#20	VN084611.D	31 Oct 2024 09:22	pH#5.0 A	JC\MD	Ok

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN111324

Review By	John Carlone	Review On	11/14/2024 9:44:38 AM
Supervise By	Maresh Dadoda	Supervise On	11/14/2024 3:18:46 PM
SubDirectory	VN111324	HP Acquire Method	HP Processing Method 82N103024W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131432
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131433,VP131434

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084814.D	13 Nov 2024 08:53		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN084815.D	13 Nov 2024 09:52		JC\MD	Ok,M
3	VN1113MBL01	VN1113MBL01	VN084816.D	13 Nov 2024 10:26	Contaminated	JC\MD	Not Ok
4	VN1113WBL01	VN1113WBL01	VN084817.D	13 Nov 2024 10:50		JC\MD	Ok
5	VN1113MBL02	VN1113MBL02	VN084818.D	13 Nov 2024 11:13		JC\MD	Ok
6	VN1113MBS01	VN1113MBS01	VN084819.D	13 Nov 2024 11:37	BS failed low for comp. #93	JC\MD	Ok,M
7	P4793-01	M00-24-00345	VN084820.D	13 Nov 2024 12:01	Need 5X	JC\MD	Dilution
8	P4793-01DL	M00-24-00345DL	VN084821.D	13 Nov 2024 13:00		JC\MD	Ok
9	VN1113WBS01	VN1113WBS01	VN084822.D	13 Nov 2024 13:24		JC\MD	Ok,M
10	VN1113WBSD01	VN1113WBSD01	VN084823.D	13 Nov 2024 14:00		JC\MD	Ok,M
11	P4722-05	WC-1(0-6)	VN084824.D	13 Nov 2024 14:24	vial B pH#6.0	JC\MD	Ok
12	P4722-10	WC-2(0-6)	VN084825.D	13 Nov 2024 14:48	vial B pH#6.0	JC\MD	Ok
13	P4722-15	WC-3(0-6)	VN084826.D	13 Nov 2024 15:12	vial B pH#6.0	JC\MD	Ok
14	P4780-01	Storage-Blank-SOIL-RE	VN084827.D	13 Nov 2024 15:35	vial B pH<2	JC\MD	Ok
15	P4739-11DL	BP-B2-VOC DL	VN084828.D	13 Nov 2024 15:59		JC\MD	Ok
16	IBLK	IBLK	VN084829.D	13 Nov 2024 16:23		JC\MD	Ok
17	P4770-04	FIELD-BLANK	VN084830.D	13 Nov 2024 16:47	vial A pH<2 FB	JC\MD	Ok
18	P4780-04	Storage-Blank-SAMPLE	VN084831.D	13 Nov 2024 17:11	vial B pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN111324

Review By	John Carlone	Review On	11/14/2024 9:44:38 AM
Supervise By	Maresh Dadoda	Supervise On	11/14/2024 3:18:46 PM
SubDirectory	VN111324	HP Acquire Method	HP Processing Method 82N103024W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP131432		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131433,VP131434		

19	P4770-05	TRIP-BLANK	VN084832.D	13 Nov 2024 17:35	vial A pH<2 TB	JC\MD	Ok
20	P4780-02	Storage-Blank-WATER	VN084833.D	13 Nov 2024 18:00	vial B pH<2	JC\MD	Ok
21	P4780-03	Storage-Blank-WATER	VN084834.D	13 Nov 2024 18:23	vial B pH<2	JC\MD	Ok
22	P4791-02	HINCHMAN-OILY-WAT	VN084835.D	13 Nov 2024 18:47	vial B pH<2	JC\MD	Ok
23	P4792-04	MANHOLE-WASTE-DR	VN084836.D	13 Nov 2024 19:11	vial B pH<2	JC\MD	Ok
24	P4799-17	WC-TA1-03-G	VN084837.D	13 Nov 2024 19:35	vial B pH#5.0	JC\MD	Ok
25	P4732-05	PPE-Grab	VN084838.D	13 Nov 2024 19:59	vial B pH#5.0	JC\MD	Ok
26	P4792-02	DIESLE-DRUM	VN084839.D	13 Nov 2024 20:23		JC\MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VN084840.D	13 Nov 2024 20:47		JC\MD	Ok,M

M : Manual Integration