

LAB CHRONICLE

OrderID:	P4722	OrderDate:	11/5/2024 3:33:08 PM
Client:	Walsh Construction Company II, LLC	Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2
Contact:	Kayla Timony	Location:	L23,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4722-02	WC-1(3.5)	TCLP	TCLP VOA	8260D	11/05/24		11/08/24	11/05/24
P4722-05	WC-1(0-6)	Water	SPLP VOA	8260D	11/05/24		11/13/24	11/05/24
P4722-07	WC-2(4)	TCLP	TCLP VOA	8260D	11/05/24		11/11/24	11/05/24
P4722-10	WC-2(0-6)	Water	SPLP VOA	8260D	11/05/24		11/13/24	11/05/24
P4722-12	WC-3(3)	TCLP	TCLP VOA	8260D	11/05/24		11/11/24	11/05/24
P4722-15	WC-3(0-6)	Water	SPLP VOA	8260D	11/05/24		11/13/24	11/05/24
P4722-16	WC-1(5.5)	TCLP	TCLP VOA	8260D	11/05/24		11/11/24	11/05/24
P4722-18	WC-2(2)	TCLP	TCLP VOA	8260D	11/05/24		11/11/24	11/05/24
P4722-20	WC-3(5)	TCLP	TCLP VOA	8260D	11/05/24		11/11/24	11/05/24

Hit Summary Sheet
SW-846

SDG No.: P4722
Client: Walsh Construction Company II, LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	WC-1(0-6)							
P4722-05	WC-1(0-6)	WATER	Acetone	7.40	J	1.40	25.0	ug/L
P4722-05	WC-1(0-6)	WATER	Methylene Chloride	3.20	J	0.32	5.00	ug/L
P4722-05	WC-1(0-6)	WATER	1,2,4-Trichlorobenzene	1.00	J	0.42	5.00	ug/L
P4722-05	WC-1(0-6)	WATER	1,2,3-Trichlorobenzene	1.10	J	0.51	5.00	ug/L
			Total Voc :	12.7				
			Total Concentration:	12.7				
Client ID:	WC-2(0-6)							
P4722-10	WC-2(0-6)	WATER	Acetone	6.10	J	1.40	25.0	ug/L
P4722-10	WC-2(0-6)	WATER	Methylene Chloride	15.5		0.32	5.00	ug/L
			Total Voc :	21.6				
			Total Concentration:	21.6				
Client ID:	WC-3(0-6)							
P4722-15	WC-3(0-6)	WATER	Methylene Chloride	1.30	J	0.32	5.00	ug/L
			Total Voc :	1.30				
			Total Concentration:	1.30				



SAMPLE DATA

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/05/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/05/24
Client Sample ID:	WC-1(0-6)	SDG No.:	P4722
Lab Sample ID:	P4722-05	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084824.D	1		11/13/24 14:24	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	7.40	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	3.20	J	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	11/05/24	
Client Sample ID:	WC-1(0-6)		SDG No.:	P4722	
Lab Sample ID:	P4722-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084824.D	1		11/13/24 14:24	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	UQ	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	0.14	UQ	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	1.00	J	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	1.10	J	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.3		74 - 125	97%	SPK: 50
1868-53-7	Dibromofluoromethane	47.8		75 - 124	96%	SPK: 50
2037-26-5	Toluene-d8	45.8		86 - 113	92%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.8		77 - 121	94%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	159000	8.218			
540-36-3	1,4-Difluorobenzene	275000	9.094			
3114-55-4	Chlorobenzene-d5	235000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	107000	13.788			

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/05/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/05/24
Client Sample ID:	WC-2(0-6)	SDG No.:	P4722
Lab Sample ID:	P4722-10	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084825.D	1		11/13/24 14:48	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	6.10	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	15.5		0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	Walsh Construction Company II, LLC			Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2			Date Received:	11/05/24	
Client Sample ID:	WC-2(0-6)			SDG No.:	P4722	
Lab Sample ID:	P4722-10			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	SPLP VOA	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084825.D	1		11/13/24 14:48	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	UQ	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	0.14	UQ	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	47.9		74 - 125	96%	SPK: 50
1868-53-7	Dibromofluoromethane	48.4		75 - 124	97%	SPK: 50
2037-26-5	Toluene-d8	45.9		86 - 113	92%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.3		77 - 121	93%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	172000	8.218			
540-36-3	1,4-Difluorobenzene	295000	9.094			
3114-55-4	Chlorobenzene-d5	255000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	118000	13.788			

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	11/05/24	
Client Sample ID:	WC-2(0-6)		SDG No.:	P4722	
Lab Sample ID:	P4722-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084825.D	1		11/13/24 14:48	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	11/05/24
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2	Date Received:	11/05/24
Client Sample ID:	WC-3(0-6)	SDG No.:	P4722
Lab Sample ID:	P4722-15	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084826.D	1		11/13/24 15:12	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	1.40	U	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	1.30	J	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	11/05/24	
Client Sample ID:	WC-3(0-6)		SDG No.:	P4722	
Lab Sample ID:	P4722-15		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084826.D	1		11/13/24 15:12	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	UQ	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	0.14	UQ	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.5		74 - 125	93%	SPK: 50
1868-53-7	Dibromofluoromethane	48.3		75 - 124	97%	SPK: 50
2037-26-5	Toluene-d8	45.9		86 - 113	92%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.8		77 - 121	90%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	181000	8.218			
540-36-3	1,4-Difluorobenzene	309000	9.1			
3114-55-4	Chlorobenzene-d5	269000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	122000	13.788			

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	11/05/24	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	11/05/24	
Client Sample ID:	WC-3(0-6)		SDG No.:	P4722	
Lab Sample ID:	P4722-15		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084826.D	1		11/13/24 15:12	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4722-05	WC-1(0-6)	1,2-Dichloroethane-d4	50	48.3	97	74	125
		Dibromofluoromethane	50	47.8	96	75	124
		Toluene-d8	50	45.8	92	86	113
		4-Bromofluorobenzene	50	46.8	94	77	121
P4722-10	WC-2(0-6)	1,2-Dichloroethane-d4	50	47.9	96	74	125
		Dibromofluoromethane	50	48.4	97	75	124
		Toluene-d8	50	45.9	92	86	113
		4-Bromofluorobenzene	50	46.3	93	77	121
P4722-15	WC-3(0-6)	1,2-Dichloroethane-d4	50	46.5	93	74	125
		Dibromofluoromethane	50	48.3	97	75	124
		Toluene-d8	50	45.9	92	86	113
		4-Bromofluorobenzene	50	44.8	90	77	121

Surrogate Summary

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN1113WBL01	VN1113WBL01	1,2-Dichloroethane-d4	50	49.5	99	74	125
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	46.6	93	86	113
		4-Bromofluorobenzene	50	45.8	92	77	121
VN1113WBS01	VN1113WBS01	1,2-Dichloroethane-d4	50	46.4	93	74	125
		Dibromofluoromethane	50	49.9	100	75	124
		Toluene-d8	50	49.2	98	86	113
		4-Bromofluorobenzene	50	51.2	102	77	121
VN1113WBSD0	VN1113WBSD01	1,2-Dichloroethane-d4	50	45.4	91	74	125
		Dibromofluoromethane	50	48.6	97	75	124
		Toluene-d8	50	47.8	96	86	113
		4-Bromofluorobenzene	50	51.0	102	77	121

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Datafile : VN084822.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1113WBS01	Dichlorodifluoromethane	20	17.7	ug/L	89			69	116	
	Chloromethane	20	15.0	ug/L	75			65	116	
	Vinyl chloride	20	20.1	ug/L	101			65	117	
	Bromomethane	20	19.6	ug/L	98			58	125	
	Chloroethane	20	20.9	ug/L	104			56	128	
	Trichlorofluoromethane	20	18.3	ug/L	92			73	115	
	1,1,2-Trichlorotrifluoroethane	20	18.4	ug/L	92			80	112	
	1,1-Dichloroethene	20	17.5	ug/L	88			74	110	
	Acetone	100	89.8	ug/L	90			60	125	
	Carbon disulfide	20	15.6	ug/L	78			64	112	
	Methyl tert-butyl Ether	20	18.8	ug/L	94			78	114	
	Methyl Acetate	20	13.3	ug/L	67			67	125	
	Methylene Chloride	20	17.5	ug/L	88			72	114	
	trans-1,2-Dichloroethene	20	17.5	ug/L	88			75	108	
	1,1-Dichloroethane	20	18.0	ug/L	90			78	112	
	Cyclohexane	20	18.1	ug/L	91			75	110	
	2-Butanone	100	91.6	ug/L	92			65	122	
	Carbon Tetrachloride	20	20.1	ug/L	101			77	113	
	cis-1,2-Dichloroethene	20	18.3	ug/L	92			77	110	
	Bromochloromethane	20	20.9	ug/L	104			70	124	
	Chloroform	20	18.9	ug/L	95			79	113	
	1,1,1-Trichloroethane	20	18.9	ug/L	95			80	108	
	Methylcyclohexane	20	19.8	ug/L	99			72	115	
	Benzene	20	18.5	ug/L	93			82	109	
	1,2-Dichloroethane	20	19.0	ug/L	95			80	115	
	Trichloroethene	20	19.2	ug/L	96			77	113	
	1,2-Dichloropropane	20	19.3	ug/L	97			83	111	
	Bromodichloromethane	20	19.3	ug/L	97			83	110	
	4-Methyl-2-Pentanone	100	98.0	ug/L	98			74	118	
	Toluene	20	20.0	ug/L	100			82	110	
	t-1,3-Dichloropropene	20	18.2	ug/L	91			79	110	
	cis-1,3-Dichloropropene	20	19.2	ug/L	96			82	110	
	1,1,2-Trichloroethane	20	19.5	ug/L	98			83	112	
	2-Hexanone	100	99.3	ug/L	99			73	117	
	Dibromochloromethane	20	19.6	ug/L	98			82	110	
	1,2-Dibromoethane	20	19.1	ug/L	96			81	110	
	Tetrachloroethene	20	20.6	ug/L	103			67	123	
	Chlorobenzene	20	18.9	ug/L	95			82	109	
	Ethyl Benzene	20	19.7	ug/L	99			83	109	
	m/p-Xylenes	40	39.9	ug/L	100			82	110	
	o-Xylene	20	20.9	ug/L	104			83	109	
	Styrene	20	19.8	ug/L	99			80	111	
	Bromoform	20	19.5	ug/L	98			79	109	
	Isopropylbenzene	20	19.1	ug/L	96			83	112	
	1,1,2,2-Tetrachloroethane	20	17.5	ug/L	88			76	118	
	1,3-Dichlorobenzene	20	16.9	ug/L	85			82	108	
	1,4-Dichlorobenzene	20	18.5	ug/L	93			82	107	
	1,2-Dichlorobenzene	20	17.8	ug/L	89			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Datafile : VN084822.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1113WBS01	1,2-Dibromo-3-Chloropropane	20	16.2	ug/L	81			68	112	
	1,2,4-Trichlorobenzene	20	16.1	ug/L	81			75	113	
	1,2,3-Trichlorobenzene	20	15.7	ug/L	79			76	114	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Datafile : VN084823.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1113WBSD01	Dichlorodifluoromethane	20	20.5	ug/L	103	15		69	116	20
	Chloromethane	20	17.0	ug/L	85	13		65	116	20
	Vinyl chloride	20	21.9	ug/L	110	9		65	117	20
	Bromomethane	20	21.4	ug/L	107	9		58	125	20
	Chloroethane	20	22.6	ug/L	113	8		56	128	20
	Trichlorofluoromethane	20	20.9	ug/L	104	12		73	115	20
	1,1,2-Trichlorotrifluoroethane	20	21.3	ug/L	106	14		80	112	20
	1,1-Dichloroethene	20	19.8	ug/L	99	12		74	110	20
	Acetone	100	100	ug/L	100	11		60	125	20
	Carbon disulfide	20	18.2	ug/L	91	15		64	112	20
	Methyl tert-butyl Ether	20	21.4	ug/L	107	13		78	114	20
	Methyl Acetate	20	16.3	ug/L	81	19		67	125	20
	Methylene Chloride	20	20.5	ug/L	103	16		72	114	20
	trans-1,2-Dichloroethene	20	20.1	ug/L	101	14		75	108	20
	1,1-Dichloroethane	20	20.1	ug/L	101	12		78	112	20
	Cyclohexane	20	20.4	ug/L	102	11		75	110	20
	2-Butanone	100	100	ug/L	100	8		65	122	20
	Carbon Tetrachloride	20	21.9	ug/L	110	9		77	113	20
	cis-1,2-Dichloroethene	20	20.5	ug/L	103	11		77	110	20
	Bromochloromethane	20	20.9	ug/L	104	0		70	124	20
	Chloroform	20	21.5	ug/L	108	13		79	113	20
	1,1,1-Trichloroethane	20	21.0	ug/L	105	10		80	108	20
	Methylcyclohexane	20	21.8	ug/L	109	10		72	115	20
	Benzene	20	20.7	ug/L	104	11		82	109	20
	1,2-Dichloroethane	20	21.5	ug/L	108	13		80	115	20
	Trichloroethene	20	22.3	ug/L	112	15		77	113	20
	1,2-Dichloropropane	20	21.0	ug/L	105	8		83	111	20
	Bromodichloromethane	20	21.4	ug/L	107	10		83	110	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	12		74	118	20
	Toluene	20	21.4	ug/L	107	7		82	110	20
	t-1,3-Dichloropropene	20	20.5	ug/L	103	12		79	110	20
	cis-1,3-Dichloropropene	20	21.0	ug/L	105	9		82	110	20
	1,1,2-Trichloroethane	20	21.2	ug/L	106	8		83	112	20
	2-Hexanone	100	110	ug/L	110	11		73	117	20
	Dibromochloromethane	20	23.0	ug/L	115	16	*	82	110	20
	1,2-Dibromoethane	20	20.8	ug/L	104	8		81	110	20
	Tetrachloroethene	20	22.0	ug/L	110	7		67	123	20
	Chlorobenzene	20	20.5	ug/L	103	8		82	109	20
	Ethyl Benzene	20	21.0	ug/L	105	6		83	109	20
	m/p-Xylenes	40	42.9	ug/L	107	7		82	110	20
	o-Xylene	20	22.0	ug/L	110	6	*	83	109	20
	Styrene	20	21.5	ug/L	108	9		80	111	20
	Bromoform	20	21.7	ug/L	109	11		79	109	20
	Isopropylbenzene	20	19.9	ug/L	100	4		83	112	20
	1,1,2,2-Tetrachloroethane	20	18.8	ug/L	94	7		76	118	20
	1,3-Dichlorobenzene	20	18.4	ug/L	92	8		82	108	20
	1,4-Dichlorobenzene	20	19.7	ug/L	99	6		82	107	20
	1,2-Dichlorobenzene	20	18.6	ug/L	93	4		82	109	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Datafile : VN084823.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN1113WBSD01	1,2-Dibromo-3-Chloropropane	20	19.1	ug/L	96	17		68	112	20
	1,2,4-Trichlorobenzene	20	17.8	ug/L	89	9		75	113	20
	1,2,3-Trichlorobenzene	20	17.3	ug/L	86	8		76	114	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN1113WBL01

Lab Name: CHEMTECH

Contract: WALS01

Lab Code: CHEM Case No.: P4722

SAS No.: P4722 SDG NO.: P4722

Lab File ID: VN084817.D

Lab Sample ID: VN1113WBL01

Date Analyzed: 11/13/2024

Time Analyzed: 10:50

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN1113WBS01	VN1113WBS01	VN084822.D	11/13/2024
VN1113WBSD01	VN1113WBSD01	VN084823.D	11/13/2024
WC-1 (0-6)	P4722-05	VN084824.D	11/13/2024
WC-2 (0-6)	P4722-10	VN084825.D	11/13/2024
WC-3 (0-6)	P4722-15	VN084826.D	11/13/2024

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722
Lab File ID: VN084569.D BFB Injection Date: 10/30/2024
Instrument ID: MSVOA_N BFB Injection Time: 10:42
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.5
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.2
173	Less than 2.0% of mass 174	1.2 (1.6) 1
174	50.0 - 100.0% of mass 95	73.5
175	5.0 - 9.0% of mass 174	5.7 (7.7) 1
176	95.0 - 101.0% of mass 174	70.1 (95.4) 1
177	5.0 - 9.0% of mass 176	4.8 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN084570.D	10/30/2024	11:46
VSTDICCC050	VSTDICCC050	VN084571.D	10/30/2024	12:09
VSTDICC020	VSTDICC020	VN084572.D	10/30/2024	12:33
VSTDICC010	VSTDICC010	VN084573.D	10/30/2024	12:57
VSTDICC005	VSTDICC005	VN084574.D	10/30/2024	13:21
VSTDICC001	VSTDICC001	VN084575.D	10/30/2024	13:45

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722
Lab File ID: VN084814.D BFB Injection Date: 11/13/2024
Instrument ID: MSVOA_N BFB Injection Time: 08:53
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.2
75	30.0 - 60.0% of mass 95	50.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.7 (0.9) 1
174	50.0 - 100.0% of mass 95	76.3
175	5.0 - 9.0% of mass 174	5.7 (7.5) 1
176	95.0 - 101.0% of mass 174	75.8 (99.3) 1
177	5.0 - 9.0% of mass 176	5 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN084815.D	11/13/2024	09:52
VN1113WBL01	VN1113WBL01	VN084817.D	11/13/2024	10:50
VN1113WBS01	VN1113WBS01	VN084822.D	11/13/2024	13:24
VN1113WBSD01	VN1113WBSD01	VN084823.D	11/13/2024	14:00
WC-1 (0-6)	P4722-05	VN084824.D	11/13/2024	14:24
WC-2 (0-6)	P4722-10	VN084825.D	11/13/2024	14:48
WC-3 (0-6)	P4722-15	VN084826.D	11/13/2024	15:12

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
 Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722
 Lab File ID: VN084815.D Date Analyzed: 11/13/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:52
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	176295	8.22	281546	9.09	243058	11.87
UPPER LIMIT	352590	8.718	563092	9.594	486116	12.365
LOWER LIMIT	88147.5	7.718	140773	8.594	121529	11.365
EPA SAMPLE NO.						
WC-1 (0-6)	158761	8.22	275207	9.09	235054	11.87
WC-2 (0-6)	172472	8.22	295434	9.09	255475	11.87
WC-3 (0-6)	181188	8.22	309126	9.10	268522	11.87
VN1113WBL01	177026	8.22	303081	9.09	266497	11.87
VN1113WBS01	193540	8.22	311172	9.10	273249	11.87
VN1113WBSD01	144835	8.22	235271	9.10	211677	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
 Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722
 Lab File ID: VN084815.D Date Analyzed: 11/13/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:52
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	126719	13.788				
UPPER LIMIT	253438	14.288				
LOWER LIMIT	63359.5	13.288				
EPA SAMPLE NO.						
WC-1 (0-6)	107145	13.79				
WC-2 (0-6)	118407	13.79				
WC-3 (0-6)	121632	13.79				
VN1113WBL01	115573	13.79				
VN1113WBS01	141208	13.79				
VN1113WBSD01	110546	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	VN1113WBL01		SDG No.:	P4722
Lab Sample ID:	VN1113WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084817.D	1		11/13/24 10:50	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	1.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	1.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	1.00	ug/L
67-64-1	Acetone	1.40	U	1.40	5.00	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	1.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	1.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	1.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	1.00	ug/L
71-43-2	Benzene	0.16	U	0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	1.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	1.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	5.00	ug/L
108-88-3	Toluene	0.18	U	0.18	1.00	ug/L

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	VN1113WBL01		SDG No.:	P4722
Lab Sample ID:	VN1113WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084817.D	1		11/13/24 10:50	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	1.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	1.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	1.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	2.00	ug/L
95-47-6	o-Xylene	0.14	U	0.14	1.00	ug/L
100-42-5	Styrene	0.16	U	0.16	1.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	1.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.5		74 - 125	99%	SPK: 50
1868-53-7	Dibromofluoromethane	50.0		75 - 124	100%	SPK: 50
2037-26-5	Toluene-d8	46.6		86 - 113	93%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.8		77 - 121	92%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	177000	8.218			
540-36-3	1,4-Difluorobenzene	303000	9.094			
3114-55-4	Chlorobenzene-d5	266000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	116000	13.788			

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	VN1113WBL01		SDG No.:	P4722
Lab Sample ID:	VN1113WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084817.D	1		11/13/24 10:50	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	VN1113WBS01		SDG No.:	P4722
Lab Sample ID:	VN1113WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084822.D	1		11/13/24 13:24	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	17.7		0.21	1.00	ug/L
74-87-3	Chloromethane	15.0		0.35	1.00	ug/L
75-01-4	Vinyl Chloride	20.1		0.34	1.00	ug/L
74-83-9	Bromomethane	19.6		1.40	5.00	ug/L
75-00-3	Chloroethane	20.9		0.56	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.3		0.34	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.4		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.5		0.26	1.00	ug/L
67-64-1	Acetone	89.8		1.40	5.00	ug/L
75-15-0	Carbon Disulfide	15.6		0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.8		0.16	1.00	ug/L
79-20-9	Methyl Acetate	13.3		0.60	1.00	ug/L
75-09-2	Methylene Chloride	17.5		0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	17.5		0.25	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.0		0.23	1.00	ug/L
110-82-7	Cyclohexane	18.1		1.60	5.00	ug/L
78-93-3	2-Butanone	91.6		1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	20.1		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.3		0.25	1.00	ug/L
74-97-5	Bromochloromethane	20.9		0.18	1.00	ug/L
67-66-3	Chloroform	18.9		0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.9		0.19	1.00	ug/L
108-87-2	Methylcyclohexane	19.8		0.19	1.00	ug/L
71-43-2	Benzene	18.5		0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.0		0.24	1.00	ug/L
79-01-6	Trichloroethene	19.2		0.32	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.3		0.19	1.00	ug/L
75-27-4	Bromodichloromethane	19.3		0.24	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	98.0		0.75	5.00	ug/L
108-88-3	Toluene	20.0		0.18	1.00	ug/L

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	VN1113WBS01	SDG No.:	P4722	
Lab Sample ID:	VN1113WBS01	Matrix:	Water	
Analytical Method:	SW8260	% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084822.D	1		11/13/24 13:24	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.2		0.21	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.2		0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.5		0.21	1.00	ug/L
591-78-6	2-Hexanone	99.3		1.10	5.00	ug/L
124-48-1	Dibromochloromethane	19.6		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.1		0.16	1.00	ug/L
127-18-4	Tetrachloroethene	20.6		0.25	1.00	ug/L
108-90-7	Chlorobenzene	18.9		0.13	1.00	ug/L
100-41-4	Ethyl Benzene	19.7		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	39.9		0.31	2.00	ug/L
95-47-6	o-Xylene	20.9		0.14	1.00	ug/L
100-42-5	Styrene	19.8		0.16	1.00	ug/L
75-25-2	Bromoform	19.5		0.21	1.00	ug/L
98-82-8	Isopropylbenzene	19.1		0.13	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	17.5		0.27	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	16.9		0.24	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.5		0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	17.8		0.19	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.2		0.46	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	16.1		0.42	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	15.7		0.51	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.4		74 - 125	93%	SPK: 50
1868-53-7	Dibromofluoromethane	49.9		75 - 124	100%	SPK: 50
2037-26-5	Toluene-d8	49.2		86 - 113	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.2		77 - 121	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	194000	8.224			
540-36-3	1,4-Difluorobenzene	311000	9.1			
3114-55-4	Chlorobenzene-d5	273000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	141000	13.788			

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	VN1113WBS01		SDG No.:	P4722
Lab Sample ID:	VN1113WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084822.D	1		11/13/24 13:24	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	VN1113WBSD01	SDG No.:	P4722	
Lab Sample ID:	VN1113WBSD01	Matrix:	Water	
Analytical Method:	SW8260	% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084823.D	1		11/13/24 14:00	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	20.5		0.21	1.00	ug/L
74-87-3	Chloromethane	17.0		0.35	1.00	ug/L
75-01-4	Vinyl Chloride	21.9		0.34	1.00	ug/L
74-83-9	Bromomethane	21.4		1.40	5.00	ug/L
75-00-3	Chloroethane	22.6		0.56	1.00	ug/L
75-69-4	Trichlorofluoromethane	20.9		0.34	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	21.3		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.8		0.26	1.00	ug/L
67-64-1	Acetone	100		1.40	5.00	ug/L
75-15-0	Carbon Disulfide	18.2		0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	21.4		0.16	1.00	ug/L
79-20-9	Methyl Acetate	16.3		0.60	1.00	ug/L
75-09-2	Methylene Chloride	20.5		0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	20.1		0.25	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.1		0.23	1.00	ug/L
110-82-7	Cyclohexane	20.4		1.60	5.00	ug/L
78-93-3	2-Butanone	100		1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.9		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.5		0.25	1.00	ug/L
74-97-5	Bromochloromethane	20.9		0.18	1.00	ug/L
67-66-3	Chloroform	21.5		0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	21.0		0.19	1.00	ug/L
108-87-2	Methylcyclohexane	21.8		0.19	1.00	ug/L
71-43-2	Benzene	20.7		0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.5		0.24	1.00	ug/L
79-01-6	Trichloroethene	22.3		0.32	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.0		0.19	1.00	ug/L
75-27-4	Bromodichloromethane	21.4		0.24	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.75	5.00	ug/L
108-88-3	Toluene	21.4		0.18	1.00	ug/L

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	VN1113WBSD01		SDG No.:	P4722
Lab Sample ID:	VN1113WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084823.D	1		11/13/24 14:00	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.5		0.21	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	21.0		0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.2		0.21	1.00	ug/L
591-78-6	2-Hexanone	110		1.10	5.00	ug/L
124-48-1	Dibromochloromethane	23.0		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.8		0.16	1.00	ug/L
127-18-4	Tetrachloroethene	22.0		0.25	1.00	ug/L
108-90-7	Chlorobenzene	20.5		0.13	1.00	ug/L
100-41-4	Ethyl Benzene	21.0		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	42.9		0.31	2.00	ug/L
95-47-6	o-Xylene	22.0		0.14	1.00	ug/L
100-42-5	Styrene	21.5		0.16	1.00	ug/L
75-25-2	Bromoform	21.7		0.21	1.00	ug/L
98-82-8	Isopropylbenzene	19.9		0.13	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.8		0.27	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.4		0.24	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.7		0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.6		0.19	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.1		0.46	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.8		0.42	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.3		0.51	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	45.4		74 - 125	91%	SPK: 50
1868-53-7	Dibromofluoromethane	48.6		75 - 124	97%	SPK: 50
2037-26-5	Toluene-d8	47.8		86 - 113	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.0		77 - 121	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	145000	8.218			
540-36-3	1,4-Difluorobenzene	235000	9.1			
3114-55-4	Chlorobenzene-d5	212000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	111000	13.788			

Report of Analysis

Client:	Walsh Construction Company II, LLC		Date Collected:	
Project:	NYCDEP C547A - Shafts 17B-1 & 18B-1 Stage 2		Date Received:	
Client Sample ID:	VN1113WBSD01		SDG No.:	P4722
Lab Sample ID:	VN1113WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084823.D	1		11/13/24 14:00	VN111324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>WALS01</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>P4722</u>
Instrument ID:	<u>MSVOA_N</u>	SAS No.:	<u>P4722</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>P4722</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>10/30/2024</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>11:46</u>
			<u>13:45</u>

LAB FILE ID:	RRF100 = VN084570.D		RRF050 = VN084571.D		RRF020 = VN084572.D			
	RRF010 = VN084573.D		RRF005 = VN084574.D		RRF001 = VN084575.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD
Dichlorodifluoromethane	0.571	0.552	0.552	0.598	0.594	0.581	0.575	3.4
Chloromethane	0.658	0.672	0.725	0.871	0.995	1.789	0.952	45.2
Vinyl Chloride	0.613	0.605	0.623	0.636	0.651	0.581	0.618	4
Bromomethane	0.292	0.296	0.310	0.336	0.405		0.328	14.2
Chloroethane	0.378	0.376	0.413	0.426	0.475	0.863	0.488	38.3
Trichlorofluoromethane	0.971	0.959	1.017	1.022	1.070	1.071	1.018	4.7
1,1,2-Trichlorotrifluoroethane	0.566	0.557	0.571	0.585	0.588	0.586	0.575	2.2
1,1-Dichloroethene	0.548	0.538	0.560	0.575	0.552	0.644	0.569	6.8
Acetone	0.209	0.204	0.213	0.223	0.241	0.338	0.238	21.3
Carbon Disulfide	1.604	1.603	1.700	1.714	1.784	2.117	1.753	10.9
Methyl tert-butyl Ether	1.773	1.758	1.802	1.779	1.786	1.572	1.745	4.9
Methyl Acetate	0.731	0.749	0.954	1.044	1.266	2.291	1.172	49.7
Methylene Chloride	0.604	0.602	0.633	0.658	0.714	0.600	0.635	7.1
trans-1,2-Dichloroethene	0.565	0.563	0.596	0.600	0.584	0.601	0.585	2.9
1,1-Dichloroethane	1.067	1.066	1.114	1.127	1.203	1.033	1.102	5.5
Cyclohexane	0.956	0.938	0.956	1.043	1.093		0.997	6.8
2-Butanone	0.316	0.315	0.348	0.338	0.370	0.334	0.337	6.1
Carbon Tetrachloride	0.530	0.514	0.532	0.548	0.537	0.488	0.525	4
cis-1,2-Dichloroethene	0.675	0.662	0.697	0.685	0.705	0.673	0.683	2.4
Bromochloromethane	0.483	0.511	0.388	0.415	0.408	0.429	0.439	10.8
Chloroform	1.099	1.086	1.142	1.154	1.222	1.025	1.121	6
1,1,1-Trichloroethane	1.000	0.991	1.046	1.073	1.032	0.980	1.021	3.5
Methylcyclohexane	0.546	0.509	0.495	0.487	0.458	0.371	0.478	12.5
Benzene	1.494	1.448	1.509	1.507	1.546	1.540	1.507	2.4
1,2-Dichloroethane	0.488	0.494	0.493	0.492	0.503	0.459	0.488	3.1
Trichloroethene	0.339	0.335	0.345	0.338	0.341	0.387	0.348	5.7
1,2-Dichloropropane	0.356	0.348	0.358	0.357	0.373	0.330	0.354	4
Bromodichloromethane	0.528	0.521	0.526	0.522	0.529	0.530	0.526	0.7
4-Methyl-2-Pentanone	0.424	0.423	0.416	0.417	0.412	0.344	0.406	7.6
Toluene	0.923	0.899	0.919	0.902	0.891	0.757	0.882	7.1

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722

Instrument ID: MSVOA_N Calibration Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Calibration Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:								
RRF100 = VN084570.D			RRF050 = VN084571.D			RRF020 = VN084572.D		
RRF010 = VN084573.D			RRF005 = VN084574.D			RRF001 = VN084575.D		
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD
t-1,3-Dichloropropene	0.552	0.543	0.538	0.532	0.547	0.555	0.544	1.6
cis-1,3-Dichloropropene	0.592	0.581	0.582	0.569	0.584	0.548	0.576	2.7
1,1,2-Trichloroethane	0.336	0.329	0.334	0.342	0.342	0.309	0.332	3.7
2-Hexanone	0.314	0.312	0.301	0.297	0.294	0.256	0.296	7.1
Dibromochloromethane	0.404	0.394	0.391	0.393	0.377	0.312	0.378	8.9
1,2-Dibromoethane	0.340	0.333	0.341	0.335	0.355	0.336	0.340	2.4
Tetrachloroethene	0.326	0.313	0.333	0.347	0.351	0.325	0.333	4.3
Chlorobenzene	1.068	1.061	1.149	1.123	1.165	1.146	1.119	3.9
Ethyl Benzene	1.957	1.891	1.928	1.880	1.849	1.697	1.867	4.9
m/p-Xylenes	0.737	0.728	0.737	0.701	0.683	0.654	0.707	4.8
o-Xylene	0.703	0.690	0.701	0.679	0.645	0.550	0.661	8.9
Styrene	1.223	1.205	1.206	1.144	1.093	1.050	1.154	6.1
Bromoform	0.287	0.295	0.298	0.289	0.302	0.286	0.293	2.3
Isopropylbenzene	3.558	3.570	3.701	3.605	3.402	3.188	3.504	5.2
1,1,2,2-Tetrachloroethane	1.052	1.073	1.163	1.190	1.317	1.222	1.170	8.4
1,3-Dichlorobenzene	1.642	1.668	1.770	1.802	1.884	2.264	1.838	12.3
1,4-Dichlorobenzene	1.646	1.674	1.782	1.867	1.879	2.773	1.937	21.7
1,2-Dichlorobenzene	1.601	1.618	1.748	1.732	1.879	2.021	1.766	9.1
1,2-Dibromo-3-Chloropropane	0.204	0.217	0.229	0.226	0.274	0.272	0.237	12.3
1,2,4-Trichlorobenzene	0.852	0.865	0.895	0.881	0.956	1.353	0.967	19.9
1,2,3-Trichlorobenzene	0.832	0.841	0.953	0.877	0.939	1.189	0.939	14.1
1,2-Dichloroethane-d4	0.689	0.721	0.708	0.722	0.771		0.722	4.2
Dibromofluoromethane	0.334	0.344	0.326	0.336	0.353		0.338	3.1
Toluene-d8	1.267	1.303	1.216	1.231	1.217		1.247	3
4-Bromofluorobenzene	0.481	0.493	0.450	0.451	0.454		0.466	4.3

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722

Instrument ID: MSVOA_N Calibration Date/Time: 11/13/2024 09:52

Lab File ID: VN084815.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.575	0.546		-5.04	20
Chloromethane	0.952	0.608	0.1	-36.13	20
Vinyl Chloride	0.618	0.665		7.61	20
Bromomethane	0.328	0.336		2.44	20
Chloroethane	0.488	0.417		-14.55	20
Trichlorofluoromethane	1.018	1.038		1.97	20
1,1,2-Trichlorotrifluoroethane	0.575	0.578		0.52	20
1,1-Dichloroethene	0.569	0.539		-5.27	20
Acetone	0.238	0.205		-13.87	20
Carbon Disulfide	1.753	1.531		-12.66	20
Methyl tert-butyl Ether	1.745	1.800		3.15	20
Methyl Acetate	1.172	0.631		-46.16	20
Methylene Chloride	0.635	0.600		-5.51	20
trans-1,2-Dichloroethene	0.585	0.566		-3.25	20
1,1-Dichloroethane	1.102	1.068	0.1	-3.09	20
Cyclohexane	0.997	0.913		-8.43	20
2-Butanone	0.337	0.325		-3.56	20
Carbon Tetrachloride	0.525	0.567		8	20
cis-1,2-Dichloroethene	0.683	0.673		-1.46	20
Bromochloromethane	0.439	0.474		7.97	20
Chloroform	1.121	1.116		-0.45	20
1,1,1-Trichloroethane	1.021	1.034		1.27	20
Methylcyclohexane	0.478	0.545		14.02	20
Benzene	1.507	1.535		1.86	20
1,2-Dichloroethane	0.488	0.515		5.53	20
Trichloroethene	0.348	0.357		2.59	20
1,2-Dichloropropane	0.354	0.368		3.95	20
Bromodichloromethane	0.526	0.545		3.61	20
4-Methyl-2-Pentanone	0.406	0.412		1.48	20
Toluene	0.882	0.942		6.8	20
t-1,3-Dichloropropene	0.544	0.546		0.37	20
cis-1,3-Dichloropropene	0.576	0.603		4.69	20
1,1,2-Trichloroethane	0.332	0.345		3.92	20
2-Hexanone	0.296	0.312		5.41	20
Dibromochloromethane	0.378	0.403		6.61	20
1,2-Dibromoethane	0.340	0.339		-0.29	20
Tetrachloroethene	0.333	0.370		11.11	20
Chlorobenzene	1.119	1.142	0.3	2.06	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722

Instrument ID: MSVOA_N Calibration Date/Time: 11/13/2024 09:52

Lab File ID: VN084815.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.867	2.028		8.62	20
m/p-Xylenes	0.707	0.777		9.9	20
o-Xylene	0.661	0.751		13.62	20
Styrene	1.154	1.259		9.1	20
Bromoform	0.293	0.315	0.1	7.51	20
Isopropylbenzene	3.504	3.640		3.88	20
1,1,2,2-Tetrachloroethane	1.170	1.056	0.3	-9.74	20
1,3-Dichlorobenzene	1.838	1.698		-7.62	20
1,4-Dichlorobenzene	1.937	1.702		-12.13	20
1,2-Dichlorobenzene	1.766	1.640		-7.14	20
1,2-Dibromo-3-Chloropropane	0.237	0.194		-18.14	20
1,2,4-Trichlorobenzene	0.967	0.823		-14.89	20
1,2,3-Trichlorobenzene	0.939	0.834		-11.18	20
1,2-Dichloroethane-d4	0.722	0.681		-5.68	20
Dibromofluoromethane	0.338	0.348		2.96	20
Toluene-d8	1.247	1.247		0	20
4-Bromofluorobenzene	0.466	0.482		3.43	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.