

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WALS01

Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG No.: P4722

Instrument ID: MSVOA_N Calibration Date/Time: 11/08/2024 09:39

Lab File ID: VN084737.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.618	0.547		-11.49	20
1,1-Dichloroethene	0.569	0.508		-10.72	20
2-Butanone	0.337	0.361		7.12	20
Carbon Tetrachloride	0.525	0.529		0.76	20
Chloroform	1.121	1.095		-2.32	20
Benzene	1.507	1.426		-5.38	20
1,2-Dichloroethane	0.488	0.484		-0.82	20
Trichloroethene	0.348	0.335		-3.74	20
Tetrachloroethene	0.333	0.341		2.4	20
Chlorobenzene	1.119	1.084	0.3	-3.13	20
1,2-Dichloroethane-d4	0.722	0.672		-6.93	20
Dibromofluoromethane	0.338	0.337		-0.3	20
Toluene-d8	1.247	1.241		-0.48	20
4-Bromofluorobenzene	0.466	0.466		0	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.