

Surrogate Summary

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4722-02	WC-1(3.5)	1,2-Dichloroethane-d4	50	49.8	99	74	125
		Dibromofluoromethane	50	49.1	98	75	124
		Toluene-d8	50	48.5	97	86	113
		4-Bromofluorobenzene	50	50.3	101	77	121
P4722-07	WC-2(4)	1,2-Dichloroethane-d4	50	47.9	96	74	125
		Dibromofluoromethane	50	48.3	97	75	124
		Toluene-d8	50	48.0	96	86	113
		4-Bromofluorobenzene	50	50.1	100	77	121
P4722-12	WC-3(3)	1,2-Dichloroethane-d4	50	47.9	96	74	125
		Dibromofluoromethane	50	49.0	98	75	124
		Toluene-d8	50	47.7	95	86	113
		4-Bromofluorobenzene	50	48.4	97	77	121
P4722-16	WC-1(5.5)	1,2-Dichloroethane-d4	50	49.4	99	74	125
		Dibromofluoromethane	50	48.6	97	75	124
		Toluene-d8	50	48.4	97	86	113
		4-Bromofluorobenzene	50	48.7	97	77	121
P4722-18	WC-2(2)	1,2-Dichloroethane-d4	50	48.7	97	74	125
		Dibromofluoromethane	50	49.4	99	75	124
		Toluene-d8	50	48.5	97	86	113
		4-Bromofluorobenzene	50	49.0	98	77	121
P4722-20	WC-3(5)	1,2-Dichloroethane-d4	50	49.3	99	74	125
		Dibromofluoromethane	50	48.2	96	75	124
		Toluene-d8	50	48.0	96	86	113
		4-Bromofluorobenzene	50	48.3	97	77	121

Surrogate Summary

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN1108WBL01	VN1108WBL01	1,2-Dichloroethane-d4	50	48.7	97	74	125
		Dibromofluoromethane	50	48.9	98	75	124
		Toluene-d8	50	45.6	91	86	113
		4-Bromofluorobenzene	50	45.9	92	77	121
VN1108WBS01	VN1108WBS01	1,2-Dichloroethane-d4	50	46.2	92	74	125
		Dibromofluoromethane	50	47.6	95	75	124
		Toluene-d8	50	47.1	94	86	113
		4-Bromofluorobenzene	50	50.6	101	77	121
VN1108WBSD0	VN1108WBSD01	1,2-Dichloroethane-d4	50	47.7	95	74	125
		Dibromofluoromethane	50	47.1	94	75	124
		Toluene-d8	50	47.3	95	86	113
		4-Bromofluorobenzene	50	46.8	94	77	121
VN1111WBL01	VN1111WBL01	1,2-Dichloroethane-d4	50	48.0	96	74	125
		Dibromofluoromethane	50	49.1	98	75	124
		Toluene-d8	50	45.9	92	86	113
		4-Bromofluorobenzene	50	48.1	96	77	121
VN1111WBS01	VN1111WBS01	1,2-Dichloroethane-d4	50	41.8	84	74	125
		Dibromofluoromethane	50	46.0	92	75	124
		Toluene-d8	50	44.8	90	86	113
		4-Bromofluorobenzene	50	46.7	93	77	121