

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Datafile : VN084741.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1108WBSD01	Vinyl chloride	20	17.6	ug/L	88	8		65	117	20
	1,1-Dichloroethene	20	17.9	ug/L	90	7		74	110	20
	2-Butanone	100	110	ug/L	110	10		65	122	20
	Carbon Tetrachloride	20	19.1	ug/L	96	1		77	113	20
	Chloroform	20	19.2	ug/L	96	2		79	113	20
	Benzene	20	18.6	ug/L	93	4		82	109	20
	1,2-Dichloroethane	20	19.4	ug/L	97	1		80	115	20
	Trichloroethene	20	18.7	ug/L	94	2		77	113	20
	Tetrachloroethene	20	19.2	ug/L	96	4		67	123	20
	Chlorobenzene	20	18.3	ug/L	92	4		82	109	20