

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Datafile : VN084766.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1111WBS01	Vinyl chloride	20	15.4	ug/L	77			65	117	
	1,1-Dichloroethene	20	16.1	ug/L	81			74	110	
	2-Butanone	100	86.1	ug/L	86			65	122	
	Carbon Tetrachloride	20	19.0	ug/L	95			77	113	
	Chloroform	20	17.7	ug/L	89			79	113	
	Benzene	20	17.5	ug/L	88			82	109	
	1,2-Dichloroethane	20	18.3	ug/L	92			80	115	
	Trichloroethene	20	17.6	ug/L	88			77	113	
	Tetrachloroethene	20	18.4	ug/L	92			67	123	
	Chlorobenzene	20	17.8	ug/L	89			82	109	