

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722
Lab File ID: VN084737.D Date Analyzed: 11/08/2024
Instrument ID: MSVOA_N Time Analyzed: 09:39
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	183021	8.22	298323	9.10	261810	11.87
UPPER LIMIT	366042	8.718	596646	9.6	523620	12.365
LOWER LIMIT	91510.5	7.718	149162	8.6	130905	11.365
EPA SAMPLE NO.						
WC-1(3.5)	173136	8.22	304734	9.10	285280	11.87
VN1108WBL01	170398	8.22	297081	9.10	255691	11.87
VN1108WBS01	174962	8.22	291201	9.10	262615	11.87
VN1108WBSD01	163813	8.22	277196	9.09	246235	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.