

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
 Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722
 Lab File ID: VN084763.D Date Analyzed: 11/11/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:54
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	166446	8.22	270648	9.10	238712	11.87
UPPER LIMIT	332892	8.718	541296	9.6	477424	12.365
LOWER LIMIT	83223	7.718	135324	8.6	119356	11.365
EPA SAMPLE NO.						
WC-2 (4)	158827	8.22	275321	9.10	255227	11.87
WC-3 (3)	166359	8.22	285540	9.10	266633	11.87
WC-1 (5.5)	166984	8.22	289915	9.10	269271	11.87
WC-2 (2)	171996	8.22	301059	9.10	280468	11.87
WC-3 (5)	169890	8.22	299832	9.10	275305	11.87
VN1111WBL01	167136	8.22	283115	9.10	253987	11.87
VN1111WBS01	167933	8.22	269161	9.10	238228	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.