

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
 Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722
 Lab File ID: VN084737.D Date Analyzed: 11/08/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:39
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	135765	13.788				
UPPER LIMIT	271530	14.288				
LOWER LIMIT	67882.5	13.288				
EPA SAMPLE NO.						
WC-1(3.5)	132176	13.79				
VN1108WBL01	113797	13.79				
VN1108WBS01	135553	13.79				
VN1108WBSD01	120765	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.